

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 034

CONTROL

(Exp.) **2014-05-05**
 (YYYY-MM-DD)

Rev 1

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	
		60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80				60 60C+ ES60	80 XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.4	7.4			± 1.0	17.6	17.3			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.31	2.25			± 0.12	4.57	4.52			± 0.15	5.08	5.05			± 0.20	
HB HGB	g/dl	6.6	6.6			± 0.4	13.6	13.5			± 0.5	15.9	15.8			± 0.6	
	g/l	66	66			± 4	136	135			± 5	159	158			± 6	
	mmol/l	4.10	4.10			± 0.25	8.45	8.38			± 0.31	9.87	9.81			± 0.37	
HT HCT	%	18.5	18.7			± 1.5	37.5	37.1			± 2.0	44.7	44.4			± 2.5	
	l/l	0.185	0.187			± 0.015	0.375	0.371			± 0.020	0.447	0.444			± 0.025	
VGM MCV	µm ³ ; fl	80	83			± 5	82	82			± 5	88	88			± 5	
TGMH MCH	pg	28.6	29.3			± 2.0	29.8	29.9			± 2.0	31.3	31.3			± 2.5	
	fmol	1.77	1.82			± 0.12	1.85	1.85			± 0.12	1.94	1.94			± 0.16	
CCMH MCHC	g/dl	35.7	35.3			± 3.0	36.3	36.4			± 3.0	35.6	35.6			± 3.0	
	g/l	357	353			± 30	363	364			± 30	356	356			± 30	
	mmol/l	22.18	21.95			± 1.86	22.54	22.62			± 1.86	22.09	22.08			± 1.86	
IDR RDW	%	12.4	13.5			± 4.0	12.3	12.9			± 4.0	12.0	12.9			± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	75			± 20	254	255			± 30	485	495			± 50	
VPM MPV	µm ³ ; fl	8.7	9.0			± 2.0	8.9	9.3			± 2.0	8.5	8.9			± 2.0	
NEUT	#	1.30	1.33			± 0.30	4.12	4.13			± 0.80	12.32	11.97			± 1.90	
	%	54.0	55.3			± 10.0	55.7	55.8			± 10.0	70.0	69.2			± 10.0	
LYMPHO	#	0.73	0.71			± 0.35	2.44	2.44			± 0.65	2.82	2.91			± 1.50	
	%	30.5	29.7			± 12.0	33.0	33.0			± 8.0	16.0	16.8			± 8.0	
MONO	#	0.13	0.13			± 0.13	0.30	0.30			± 0.30	0.79	0.71			± 0.71	
	%	5.5	5.0			± 5.0	4.0	4.0			± 4.0	4.5	4.0			± 4.0	
EOS	#	0.18	0.18			± 0.18	0.29	0.29			± 0.29	0.91	0.95			± 0.91	
	%	7.0	7.0			± 7.0	3.8	3.8			± 3.8	5.0	5.5			± 5.0	
BASO	#	0.07	0.07			± 0.07	0.26	0.25			± 0.19	0.79	0.78			± 0.36	
	%	3.0	3.0			± 2.5	3.5	3.4			± 2.5	4.5	4.5			± 2.0	

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 034

CONTROL

(Exp.) **2014-05-05**
 (YYYY-MM-DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120					120 120 RETIC	DX120 DF120				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l						± 0.4						± 1.0						± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						± 0.12						± 0.15						± 0.20
HB HGB	g/dl						± 0.4						± 0.5						± 0.6
	g/l						± 4						± 5						± 6
	mmol/l						± 0.25						± 0.31						± 0.37
HT HCT	%						± 1.5						± 2.0						± 2.5
	l/l						± 0.015						± 0.020						± 0.025
VGM MCV	µm ³ ; fl						± 5						± 5						± 5
TGMH MCH	pg						± 2.0						± 2.0						± 2.5
	fmol						± 0.12						± 0.12						± 0.16
CCMH MCHC	g/dl						± 3.0						± 3.0						± 3.0
	g/l						± 30						± 30						± 30
IDR RDW	mmol/l						± 1.86						± 1.86						± 1.86
	%						± 4.0						± 4.0						± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						± 20						± 30						± 50
VPM MPV	µm ³ ; fl						± 2.0						± 2.0						± 2.0
NEUT	#						± 0.30						± 0.80						± 1.90
	%						± 10.0						± 10.0						± 10.0
LYMPHO	#						± 0.35						± 0.65						± 1.50
	%						± 12.0						± 8.0						± 8.0
MONO	#						± 0.13						± 0.30						± 0.71
	%						± 5.0						± 4.0						± 4.0
EOS	#						± 0.18						± 0.29						± 0.91
	%						± 7.0						± 3.8						± 5.0
BASO	#						± 0.07						± 0.19						± 0.36
	%						± 2.5						± 2.5						± 2.0

Ref: TEMP-0821 Rev.33 FRONT / RECTO 9930080-B