

LOT PX 402
Rev 2

CONTROL

 (Exp.) 2017-01-05
(YYYY - MM - DD)

| | | ABX Lysebio | | | | | | | | | | | | | | | | | | | | |
|--------------------------|--|--------------------|------------|-------|-------|--------|---------|-------------------------|--------------------|------------|-------|-------|---------|-------|-------------------------|--------------------|------------|-------|---------|--------|---|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | | L | TOLERANCES TOLERANCE | CONTROL | | | | | N | TOLERANCES TOLERANCE | CONTROL | | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | | | PENTRA | | | | | | | PENTRA | | | | | | |
| | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.4 | 2.4 | 2.4 | 2.4 | 2.5 | ± 0.4 | 7.6 | 7.6 | 7.6 | 7.6 | 7.6 | ± 1.0 | 18.0 | 17.7 | 17.9 | 17.7 | 18.2 | ± 2.2 | | | |
| GR RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 2.35 | 2.32 | 2.35 | 2.32 | 2.35 | ± 0.16 | 4.54 | 4.54 | 4.54 | 4.54 | 4.52 | ± 0.20 | 5.05 | 5.12 | 5.10 | 5.12 | 5.10 | ± 0.25 | | | |
| HB HGB | g/dl | 6.8 | 6.8 | 6.8 | 6.8 | 6.8 | ± 0.4 | 13.5 | 13.6 | 13.5 | 13.6 | 13.4 | ± 0.5 | 16.4 | 16.5 | 16.3 | 16.5 | 16.3 | ± 0.6 | | | |
| | g/l | 68 | 68 | 68 | 68 | 68 | ± 4 | 135 | 136 | 135 | 136 | 134 | ± 5 | 164 | 165 | 163 | 165 | 163 | ± 6 | | | |
| | mmol/l | 4.22 | 4.22 | 4.22 | 4.22 | 4.22 | ± 0.25 | 8.38 | 8.45 | 8.38 | 8.45 | 8.32 | ± 0.31 | 10.18 | 10.25 | 10.12 | 10.25 | 10.12 | ± 0.37 | | | |
| HT HCT | % | 19.0 | 19.3 | 19.0 | 19.3 | 18.9 | ± 1.5 | 36.8 | 36.8 | 36.3 | 36.8 | 36.6 | ± 2.0 | 44.4 | 45.1 | 43.9 | 45.1 | 44.6 | ± 2.5 | | | |
| | l/l | 0.190 | 0.193 | 0.190 | 0.193 | 0.189 | ± 0.015 | 0.368 | 0.368 | 0.363 | 0.368 | 0.366 | ± 0.020 | 0.444 | 0.451 | 0.439 | 0.451 | 0.446 | ± 0.025 | | | |
| VGM MCV | µm ³ ; fl | 81 | 83 | 81 | 83 | 80.5 | ± 5 | 81 | 81 | 80 | 81 | 81.0 | ± 5 | 88 | 88 | 86 | 88 | 87.5 | ± 5 | | | |
| TGMH MCH | pg | 28.9 | 29.3 | 28.9 | 29.3 | 28.9 | ± 2.0 | 29.7 | 30.0 | 29.7 | 30.0 | 29.6 | ± 2.0 | 32.5 | 32.2 | 32.0 | 32.2 | 32.0 | ± 2.5 | | | |
| | fmol | 1.80 | 1.82 | 1.80 | 1.82 | 1.80 | ± 0.12 | 1.85 | 1.86 | 1.85 | 1.86 | 1.84 | ± 0.12 | 2.02 | 2.00 | 1.98 | 2.00 | 1.98 | ± 0.16 | | | |
| CCMH MCHC | g/dl | 35.7 | 35.3 | 35.7 | 35.3 | 35.9 | ± 3.0 | 36.7 | 37.0 | 37.2 | 37.0 | 36.6 | ± 3.0 | 36.9 | 36.6 | 37.2 | 36.6 | 36.5 | ± 3.0 | | | |
| | g/l | 357 | 353 | 357 | 353 | 359 | ± 30 | 367 | 370 | 372 | 370 | 366 | ± 30 | 369 | 366 | 372 | 366 | 365 | ± 30 | | | |
| | mmol/l | 22.18 | 21.93 | 22.18 | 21.93 | 22.32 | ± 1.86 | 22.80 | 22.97 | 23.08 | 22.97 | 22.73 | ± 1.86 | 22.92 | 22.74 | 23.08 | 22.74 | 22.68 | ± 1.86 | | | |
| IDR RDW | % | 13.4 | 13.2 | 12.7 | 13.2 | 12.0 | ± 4.0 | 13.4 | 13.8 | 12.6 | 13.8 | 13.0 | ± 4.0 | 12.7 | 13.3 | 12.0 | 13.3 | 12.4 | ± 4.0 | | | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 70 | 68 | 68 | 68 | 70 | ± 20 | 248 | 243 | 248 | 243 | 241 | ± 30 | 490 | 495 | 510 | 495 | 485 | ± 50 | | | |
| VPM MPV | µm ³ ; fl | 8.9 | 9.2 | 9.2 | 9.2 | 8.7 | ± 2.0 | 9.2 | 9.4 | 9.2 | 9.4 | 8.8 | ± 2.0 | 8.8 | 9.0 | 8.9 | 9.0 | 8.4 | ± 2.0 | | | |
| NEUT | # | 1.38 | 1.41 | 1.38 | 1.41 | 1.47 | ± 0.35 | 4.40 | 4.45 | 4.37 | 4.45 | 4.45 | ± 0.90 | 12.96 | 12.96 | 13.07 | 12.96 | 13.41 | ± 1.90 | | | |
| | % | 57.5 | 58.7 | 57.5 | 58.7 | 58.9 | ± 10.0 | 57.9 | 58.5 | 57.5 | 58.5 | 58.6 | ± 10.0 | 72.0 | 73.2 | 73.0 | 73.2 | 73.7 | ± 10.0 | | | |
| LYMPHO | # | 0.69 | 0.67 | 0.70 | 0.67 | 0.72 | ± 0.33 | 2.28 | 2.26 | 2.39 | 2.26 | 2.33 | ± 0.70 | 2.52 | 2.37 | 2.51 | 2.37 | 2.60 | ± 1.50 | | | |
| | % | 28.7 | 27.8 | 29.0 | 27.8 | 28.6 | ± 12.0 | 30.0 | 29.7 | 31.5 | 29.7 | 30.7 | ± 8.0 | 14.0 | 13.4 | 14.0 | 13.4 | 14.3 | ± 8.0 | | | |
| MONO | # | 0.11 | 0.09 | 0.10 | 0.09 | 0.11 | ± 0.09 | 0.34 | 0.30 | 0.27 | 0.30 | 0.29 | ± 0.27 | 0.81 | 0.80 | 0.64 | 0.80 | 0.66 | ± 0.64 | | | |
| | % | 4.5 | 3.7 | 4.2 | 3.7 | 4.2 | ± 3.7 | 4.5 | 4.0 | 3.5 | 4.0 | 3.8 | ± 3.5 | 4.5 | 4.5 | 3.6 | 4.5 | 3.6 | ± 3.6 | | | |
| EOS | # | 0.14 | 0.16 | 0.14 | 0.16 | 0.12 | ± 0.12 | 0.29 | 0.32 | 0.30 | 0.32 | 0.27 | ± 0.27 | 0.90 | 0.80 | 0.90 | 0.80 | 0.76 | ± 0.76 | | | |
| | % | 6.0 | 6.5 | 6.0 | 6.5 | 4.9 | ± 4.9 | 3.8 | 4.2 | 4.0 | 4.2 | 3.5 | ± 3.5 | 5.0 | 4.5 | 5.0 | 4.5 | 4.2 | ± 4.2 | | | |
| BASO | # | 0.08 | 0.08 | 0.08 | 0.08 | 0.09 | ± 0.08 | 0.29 | 0.27 | 0.27 | 0.27 | 0.26 | ± 0.26 | 0.81 | 0.78 | 0.79 | 0.78 | 0.76 | ± 0.76 | | | |
| | % | 3.3 | 3.3 | 3.3 | 3.3 | 3.4 | ± 3.3 | 3.8 | 3.6 | 3.5 | 3.6 | 3.4 | ± 3.4 | 4.5 | 4.4 | 4.4 | 4.4 | 4.2 | ± 4.2 | | | |

LOT PX 402
Rev 2

CONTROL

 (Exp.) 2017-01-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | |
|--------------------------|-------|--|-------------|-----------|----------|----------|---------|-----------|-------|----------|--|---------|---------|-------|----------|--|---------|-------------------------|
| | | | CONTROL | | | | L | CONTROL | | | | N | CONTROL | | | | H | TOLERANCES TOLERANCE |
| | | | PENTRA | | | | | PENTRA | | | | | PENTRA | | | | | |
| | | | 120 | DX120 | DX NEXUS | | | 120 | DX120 | DX NEXUS | | | 120 | DX120 | DX NEXUS | | | |
| 120 RETIC | DF120 | DF NEXUS | | 120 RETIC | DF120 | DF NEXUS | | 120 RETIC | DF120 | DF NEXUS | | | | | | | | |
| GB | WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.4 | 2.4 | 2.4 | | ± 0.4 | 7.8 | 7.8 | 7.8 | | ± 1.0 | 18.6 | 18.6 | 18.6 | | ± 2.2 | |
| GR | RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.35 | 2.35 | 2.35 | | ± 0.16 | 4.54 | 4.54 | 4.54 | | ± 0.20 | 5.16 | 5.16 | 5.16 | | ± 0.25 | |
| HB | HGB | g/dl | 6.9 | 6.9 | 6.9 | | ± 0.4 | 13.5 | 13.5 | 13.5 | | ± 0.5 | 16.3 | 16.3 | 16.3 | | ± 0.6 | |
| | | g/l | 69 | 69 | 69 | | ± 4 | 135 | 135 | 135 | | ± 5 | 163 | 163 | 163 | | ± 6 | |
| | | mmol/l | 4.28 | 4.28 | 4.28 | | ± 0.25 | 8.38 | 8.38 | 8.38 | | ± 0.31 | 10.12 | 10.12 | 10.12 | | ± 0.37 | |
| HT | HCT | % | 19.5 | 19.5 | 19.5 | | ± 1.5 | 37.2 | 37.2 | 37.2 | | ± 2.0 | 45.4 | 45.4 | 45.4 | | ± 2.5 | |
| | | l/l | 0.195 | 0.195 | 0.195 | | ± 0.015 | 0.372 | 0.372 | 0.372 | | ± 0.020 | 0.454 | 0.454 | 0.454 | | ± 0.025 | |
| VGM | MCV | µm ³ ; fl | 83 | 83 | 83 | | ± 5 | 82 | 82 | 82 | | ± 5 | 88 | 88 | 88 | | ± 5 | |
| TGMH | MCH | pg | 29.4 | 29.4 | 29.4 | | ± 2.0 | 29.7 | 29.7 | 29.7 | | ± 2.0 | 31.6 | 31.6 | 31.6 | | ± 2.5 | |
| | | fmol | 1.82 | 1.82 | 1.82 | | ± 0.12 | 1.85 | 1.85 | 1.85 | | ± 0.12 | 1.96 | 1.96 | 1.96 | | ± 0.16 | |
| CCMH | MCHC | g/dl | 35.4 | 35.4 | 35.4 | | ± 3.0 | 36.3 | 36.3 | 36.3 | | ± 3.0 | 35.9 | 35.9 | 35.9 | | ± 3.0 | |
| | | g/l | 354 | 354 | 354 | | ± 30 | 363 | 363 | 363 | | ± 30 | 359 | 359 | 359 | | ± 30 | |
| | | mmol/l | 21.97 | 21.97 | 21.97 | | ± 1.86 | 22.52 | 22.52 | 22.52 | | ± 1.86 | 22.29 | 22.29 | 22.29 | | ± 1.86 | |
| IDR | RDW | % | 15.0 | 15.0 | 15.0 | | ± 4.0 | 16.8 | 16.8 | 16.8 | | ± 4.0 | 14.7 | 14.7 | 14.7 | | ± 4.0 | |
| PLAQ. | PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 72 | 72 | 72 | | ± 20 | 250 | 250 | 250 | | ± 30 | 490 | 490 | 490 | | ± 50 | |
| VPM | MPV | µm ³ ; fl | 9.5 | 9.5 | 9.5 | | ± 2.0 | 9.6 | 9.6 | 9.6 | | ± 2.0 | 9.2 | 9.2 | 9.2 | | ± 2.0 | |
| NEUT | # | # | 1.41 | 1.46 | 1.46 | | ± 0.35 | 4.56 | 4.68 | 4.68 | | ± 0.90 | 13.90 | 14.00 | 14.00 | | ± 1.90 | |
| | | % | 58.8 | 60.9 | 60.9 | | ± 10.0 | 58.5 | 60.0 | 60.0 | | ± 10.0 | 74.5 | 75.4 | 75.4 | | ± 10.0 | |
| LYMPHO | # | # | 0.68 | 0.62 | 0.62 | | ± 0.33 | 2.33 | 2.18 | 2.18 | | ± 0.70 | 2.38 | 2.21 | 2.21 | | ± 1.50 | |
| | | % | 28.3 | 25.7 | 25.7 | | ± 12.0 | 29.9 | 27.9 | 27.9 | | ± 8.0 | 12.8 | 11.9 | 11.9 | | ± 8.0 | |
| MONO | # | # | 0.10 | 0.10 | 0.10 | | ± 0.10 | 0.35 | 0.37 | 0.37 | | ± 0.35 | 0.93 | 0.93 | 0.93 | | ± 0.93 | |
| | | % | 4.1 | 4.2 | 4.2 | | ± 4.1 | 4.5 | 4.7 | 4.7 | | ± 4.5 | 5.0 | 5.0 | 5.0 | | ± 5.0 | |
| EOS | # | # | 0.15 | 0.16 | 0.16 | | ± 0.15 | 0.32 | 0.34 | 0.34 | | ± 0.32 | 0.97 | 0.97 | 0.97 | | ± 0.97 | |
| | | % | 6.3 | 6.7 | 6.7 | | ± 6.3 | 4.1 | 4.4 | 4.4 | | ± 4.1 | 5.2 | 5.2 | 5.2 | | ± 5.2 | |
| BASO | # | # | 0.06 | 0.06 | 0.06 | | ± 0.06 | 0.23 | 0.23 | 0.23 | | ± 0.23 | 0.47 | 0.47 | 0.47 | | ± 0.47 | |
| | | % | 2.5 | 2.5 | 2.5 | | ± 2.5 | 3.0 | 3.0 | 3.0 | | ± 3.0 | 2.5 | 2.5 | 2.5 | | ± 2.5 | |

1300027264-A
FRONT / RECTO
Ref: TEMP-0821 Rev.39