


LOT PX 412
Rev 1


CONTROL

 (Exp.) 2018-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio															
		CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					
		80 XL80	XLR				80 XL80	XLR				80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5				± 0.4	7.3	7.3				± 1.0	15.9	15.9		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.27	2.27			± 0.16	4.60	4.60			± 0.20	5.04	5.04			± 0.25	
HB HGB	g/dl	6.6	6.6			± 0.4	13.9	13.9			± 0.5	15.8	15.8			± 0.6	
	g/l	66	66			± 4	139	139			± 5	158	158			± 6	
	mmol/l	4.10	4.10			± 0.25	8.63	8.63			± 0.31	9.81	9.81			± 0.37	
HT HCT	%	18.8	18.8			± 1.5	38.2	38.2			± 2.0	43.8	43.3			± 2.5	
	l/l	0.188	0.188			± 0.015	0.382	0.382			± 0.020	0.438	0.433			± 0.025	
VGM MCV	µm ³ ; fl	83	83			± 5	83	83			± 5	87	86			± 5	
TGMH MCH	pg	29.1	29.1			± 2.0	30.2	30.2			± 2.0	31.3	31.3			± 2.5	
	fmol	1.81	1.81			± 0.12	1.88	1.88			± 0.12	1.95	1.95			± 0.16	
CCMH MCHC	g/dl	35.0	35.0			± 3.0	36.4	36.4			± 3.0	36.0	36.5			± 3.0	
	g/l	350	350			± 30	364	364			± 30	360	365			± 30	
	mmol/l	21.75	21.75			± 1.86	22.61	22.61			± 1.86	22.38	22.64			± 1.86	
IDR RDW	%	12.8	11.7			± 4.0	13.8	12.5			± 4.0	13.8	12.7			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66			± 20	243	243			± 30	480	480			± 50	
VPM MPV	µm ³ ; fl	9.7	9.7			± 2.0	9.7	10.0			± 2.0	9.4	9.7			± 2.0	
NEUT	#	1.42	1.42			± 0.35	4.15	4.15			± 0.90	10.80	10.80			± 1.90	
	%	56.8	56.8			± 10.0	56.9	56.9			± 10.0	67.9	67.9			± 10.0	
LYMPHO	#	0.75	0.75			± 0.33	2.39	2.39			± 0.70	2.91	2.91			± 1.50	
	%	30.1	30.1			± 12.0	32.7	32.7			± 8.0	18.3	18.3			± 8.0	
MONO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.59	0.59			± 0.59	
	%	3.4	3.4			± 3.4	3.5	3.5			± 3.5	3.7	3.7			± 3.7	
EOS	#	0.16	0.16			± 0.16	0.24	0.24			± 0.24	0.92	0.92			± 0.92	
	%	6.4	6.4			± 6.4	3.3	3.3			± 3.3	5.8	5.8			± 5.8	
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.68	0.68			± 0.68	
	%	3.3	3.3			± 3.3	3.6	3.6			± 3.6	4.3	4.3			± 4.3	

LOT PX 412
Rev 1

CONTROL

 (Exp.) 2018-09-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5		± 0.4	7.3	7.3	7.4		± 1.0	16.0	15.9	16.4		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.30	2.30		± 0.16	4.60	4.60	4.60		± 0.20	5.04	5.02	5.06		± 0.25			
	g/dl	6.7	6.6	6.7		± 0.4	14.0	13.8	13.9		± 0.5	15.8	15.7	15.8		± 0.6			
HB HGB	g/l	67	66	67		± 4	140	138	139		± 5	158	157	158		± 6			
	mmol/l	4.16	4.10	4.16		± 0.25	8.69	8.57	8.63		± 0.31	9.81	9.75	9.81		± 0.37			
HT HCT	%	18.9	18.4	18.6		± 1.5	38.2	37.7	37.7		± 2.0	43.8	43.2	43.0		± 2.5			
	l/l	0.189	0.184	0.186		± 0.015	0.382	0.377	0.377		± 0.020	0.438	0.432	0.430		± 0.025			
VGM MCV	µm ³ ; fl	81	80	81.0		± 5	83	82	82.0		± 5	87	86	85.0		± 5			
TGMH MCH	pg	28.8	28.7	29.1		± 2.0	30.4	30.0	30.2		± 2.0	31.3	31.3	31.2		± 2.5			
	fmol	1.79	1.78	1.81		± 0.12	1.89	1.86	1.88		± 0.12	1.95	1.94	1.94		± 0.16			
	g/dl	35.5	35.9	36.0		± 3.0	36.7	36.6	36.9		± 3.0	36.0	36.4	36.7		± 3.0			
CCMH MCHC	g/l	355	359	360		± 30	367	366	369		± 30	360	364	367		± 30			
	mmol/l	22.05	22.28	22.33		± 1.86	22.77	22.72	22.88		± 1.86	22.38	22.58	22.81		± 1.86			
IDR RDW	%	12.5	12.2	12.4		± 4.0	13.0	13.0	12.0		± 4.0	13.0	12.5	12.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	68		± 20	250	250	240		± 30	490	490	475		± 50			
VPM MPV	µm ³ ; fl	9.5	9.4	9.2		± 2.0	9.4	9.4	9.1		± 2.0	9.2	9.2	8.8		± 2.0			
NEUT	#	1.39	1.37	1.47		± 0.35	4.20	4.13	4.29		± 0.90	11.09	10.84	11.48		± 1.90			
	%	58.0	57.2	58.6		± 10.0	57.5	56.6	58.0		± 10.0	69.3	68.2	70.0		± 10.0			
LYMPHO	#	0.70	0.71	0.73		± 0.33	2.35	2.37	2.41		± 0.70	2.74	2.81	2.90		± 1.50			
	%	29.1	29.6	29.2		± 12.0	32.2	32.5	32.5		± 8.0	17.1	17.7	17.7		± 8.0			
MONO	#	0.08	0.09	0.08		± 0.08	0.23	0.28	0.22		± 0.22	0.56	0.65	0.49		± 0.49			
	%	3.2	3.7	3.0		± 3.0	3.2	3.9	3.0		± 3.0	3.5	4.1	3.0		± 3.0			
EOS	#	0.15	0.15	0.15		± 0.15	0.26	0.26	0.22		± 0.22	0.93	0.92	0.87		± 0.87			
	%	6.2	6.2	5.8		± 5.8	3.5	3.5	3.0		± 3.0	5.8	5.8	5.3		± 5.3			
BASO	#	0.08	0.08	0.09		± 0.08	0.26	0.26	0.26		± 0.26	0.69	0.67	0.66		± 0.66			
	%	3.5	3.3	3.4		± 3.3	3.6	3.5	3.5		± 3.5	4.3	4.2	4.0		± 4.0			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42