


**LOT** PX 422  
Rev 2


**CONTROL**

 (Exp.) 2020-05-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.5</b>	<b>2.5</b>			± 0.4	<b>7.5</b>	<b>7.5</b>			± 1.0	<b>17.8</b>	<b>17.8</b>			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.35</b>	<b>2.35</b>			± 0.16	<b>4.67</b>	<b>4.67</b>			± 0.20	<b>5.25</b>	<b>5.25</b>			± 0.25			
HB HGB	g/dl	<b>6.6</b>	<b>6.6</b>			± 0.4	<b>14.0</b>	<b>14.0</b>			± 0.5	<b>16.7</b>	<b>16.7</b>			± 0.6			
	g/l	<b>66</b>	<b>66</b>			± 4	<b>140</b>	<b>140</b>			± 5	<b>167</b>	<b>167</b>			± 6			
	mmol/l	<b>4.10</b>	<b>4.10</b>			± 0.25	<b>8.69</b>	<b>8.69</b>			± 0.31	<b>10.37</b>	<b>10.37</b>			± 0.37			
HT HCT	%	<b>19.3</b>	<b>19.3</b>			± 1.5	<b>39.2</b>	<b>39.2</b>			± 2.0	<b>46.2</b>	<b>46.2</b>			± 2.5			
	l/l	<b>0.193</b>	<b>0.193</b>			± 0.015	<b>0.392</b>	<b>0.392</b>			± 0.020	<b>0.462</b>	<b>0.462</b>			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	<b>82</b>	<b>82</b>			± 5	<b>84</b>	<b>84</b>			± 5	<b>88</b>	<b>88</b>			± 5			
TGMH MCH	pg	<b>28.1</b>	<b>28.1</b>			± 2.0	<b>30.0</b>	<b>30.0</b>			± 2.0	<b>31.8</b>	<b>31.8</b>			± 2.5			
	fmol	<b>1.74</b>	<b>1.74</b>			± 0.12	<b>1.86</b>	<b>1.86</b>			± 0.12	<b>1.98</b>	<b>1.98</b>			± 0.16			
CCMH MCHC	g/dl	<b>34.3</b>	<b>34.3</b>			± 3.0	<b>35.7</b>	<b>35.7</b>			± 3.0	<b>36.1</b>	<b>36.1</b>			± 3.0			
	g/l	<b>343</b>	<b>343</b>			± 30	<b>357</b>	<b>357</b>			± 30	<b>361</b>	<b>361</b>			± 30			
	mmol/l	<b>21.27</b>	<b>21.27</b>			± 1.86	<b>22.16</b>	<b>22.16</b>			± 1.86	<b>22.45</b>	<b>22.45</b>			± 1.86			
IDR RDW	%	<b>14.0</b>	<b>13.0</b>			± 4.0	<b>14.0</b>	<b>12.5</b>			± 4.0	<b>13.5</b>	<b>12.0</b>			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>63</b>	<b>63</b>			± 20	<b>241</b>	<b>241</b>			± 30	<b>487</b>	<b>487</b>			± 50			
VPM MPV	µm <sup>3</sup> ; fl	<b>10.4</b>	<b>10.4</b>			± 2.0	<b>9.6</b>	<b>9.6</b>			± 2.0	<b>9.8</b>	<b>9.8</b>			± 2.0			
NEUT	#	<b>1.39</b>	<b>1.39</b>			± 0.35	<b>4.22</b>	<b>4.24</b>			± 0.90	<b>12.44</b>	<b>12.34</b>			± 1.90			
	%	<b>55.7</b>	<b>55.5</b>			± 10.0	<b>56.2</b>	<b>56.5</b>			± 10.0	<b>69.9</b>	<b>69.3</b>			± 10.0			
LYMPHO	#	<b>0.77</b>	<b>0.78</b>			± 0.33	<b>2.46</b>	<b>2.41</b>			± 0.70	<b>2.83</b>	<b>2.83</b>			± 1.50			
	%	<b>30.7</b>	<b>31.0</b>			± 12.0	<b>32.8</b>	<b>32.1</b>			± 8.0	<b>15.9</b>	<b>15.9</b>			± 8.0			
MONO	#	<b>0.09</b>	<b>0.10</b>			± 0.09	<b>0.27</b>	<b>0.32</b>			± 0.27	<b>0.77</b>	<b>0.87</b>			± 0.77			
	%	<b>3.4</b>	<b>3.9</b>			± 3.4	<b>3.6</b>	<b>4.2</b>			± 3.6	<b>4.3</b>	<b>4.9</b>			± 4.3			
EOS	#	<b>0.17</b>	<b>0.16</b>			± 0.16	<b>0.29</b>	<b>0.28</b>			± 0.28	<b>0.98</b>	<b>0.96</b>			± 0.96			
	%	<b>6.9</b>	<b>6.3</b>			± 6.3	<b>3.9</b>	<b>3.7</b>			± 3.7	<b>5.5</b>	<b>5.4</b>			± 5.4			
BASO	#	<b>0.08</b>	<b>0.08</b>			± 0.08	<b>0.26</b>	<b>0.26</b>			± 0.26	<b>0.78</b>	<b>0.80</b>			± 0.78			
	%	<b>3.3</b>	<b>3.3</b>			± 3.3	<b>3.5</b>	<b>3.5</b>			± 3.5	<b>4.4</b>	<b>4.5</b>			± 4.4			

**LOT** PX 422  
Rev 2

**CONTROL**

 (Exp.) 2020-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5			± 0.4	7.4	7.4	7.6			± 1.0	17.9	17.8	18.1			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.39	2.36	2.34			± 0.16	4.65	4.65	4.63			± 0.20	5.21	5.20	5.23			± 0.25
		g/dl	6.6	6.5	6.6			± 0.4	13.9	13.8	13.9			± 0.5	16.7	16.5	16.6			± 0.6
HB	HGB	g/l	66	65	66			± 4	139	138	139			± 5	167	165	166			± 6
		mmol/l	4.10	4.04	4.10			± 0.25	8.63	8.57	8.63			± 0.31	10.37	10.25	10.31			± 0.37
HT	HCT	%	19.1	18.9	18.4			± 1.5	38.6	38.1	37.7			± 2.0	45.8	45.2	44.5			± 2.5
		l/l	0.191	0.189	0.184			± 0.015	0.386	0.381	0.377			± 0.020	0.458	0.452	0.445			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	80	80	78.5			± 5	83	82	81.5			± 5	88	87	85.0			± 5
TGMH	MCH	pg	27.6	27.5	28.2			± 2.0	29.9	29.7	30.0			± 2.0	32.1	31.7	31.7			± 2.5
		fmol	1.71	1.71	1.75			± 0.12	1.86	1.84	1.86			± 0.12	1.99	1.97	1.97			± 0.16
CCMH	MCHC	g/dl	34.5	34.4	35.9			± 3.0	36.0	36.2	36.8			± 3.0	36.4	36.5	37.3			± 3.0
		g/l	345	344	359			± 30	360	362	368			± 30	364	365	373			± 30
		mmol/l	21.44	21.38	22.31			± 1.86	22.37	22.48	22.88			± 1.86	22.62	22.65	23.19			± 1.86
IDR	RDW	%	13.5	13.5	12.5			± 4.0	13.0	13.0	11.5			± 4.0	13.0	12.5	11.0			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	63	65			± 20	242	240	234			± 30	480	485	475			± 50
VPM	MPV	µm <sup>3</sup> ; fl	10.2	10.2	9.9			± 2.0	9.5	9.4	9.5			± 2.0	9.6	9.7	9.5			± 2.0
NEUT		#	1.36	1.35	1.41			± 0.35	4.16	4.09	4.33			± 0.90	12.57	12.32	12.91			± 1.90
		%	56.5	56.3	56.3			± 10.0	56.2	55.3	57.0			± 10.0	70.2	69.2	71.3			± 10.0
LYMPHO		#	0.73	0.72	0.75			± 0.33	2.46	2.43	2.44			± 0.70	2.86	2.78	2.77			± 1.50
		%	30.4	30.2	30.1			± 12.0	33.2	32.9	32.1			± 8.0	16.0	15.6	15.3			± 8.0
MONO		#	0.08	0.09	0.10			± 0.08	0.24	0.34	0.30			± 0.24	0.75	1.00	0.78			± 0.75
		%	3.3	3.7	3.8			± 3.3	3.3	4.6	3.9			± 3.3	4.2	5.6	4.3			± 4.2
EOS		#	0.16	0.16	0.16			± 0.16	0.28	0.29	0.27			± 0.27	0.93	0.94	0.87			± 0.87
		%	6.5	6.6	6.3			± 6.3	3.8	3.9	3.5			± 3.5	5.2	5.3	4.8			± 4.8
BASO		#	0.08	0.08	0.09			± 0.08	0.26	0.24	0.27			± 0.24	0.79	0.77	0.78			± 0.77
		%	3.3	3.2	3.5			± 3.2	3.5	3.3	3.5			± 3.3	4.4	4.3	4.3			± 4.3

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43