


**LOT** PX 421  
Rev 1


**CONTROL**

 (Exp.) 2020-03-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6			± 0.4	7.8	7.8			± 1.0	18.3	18.3			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.37			± 0.16	4.48	4.48			± 0.20	5.14	5.14			± 0.25			
HB HGB	g/dl	7.0	7.0			± 0.4	13.2	13.2			± 0.5	16.5	16.5			± 0.6			
	g/l	70	70			± 4	132	132			± 5	165	165			± 6			
	mmol/l	4.35	4.35			± 0.25	8.20	8.20			± 0.31	10.25	10.25			± 0.37			
HT HCT	%	19.7	19.7			± 1.5	36.3	36.3			± 2.0	44.7	44.7			± 2.5			
	l/l	0.197	0.197			± 0.015	0.363	0.363			± 0.020	0.447	0.447			± 0.025			
VGM MCV	µm <sup>3</sup> fl	83	83			± 5	81	81			± 5	87	87			± 5			
TGMH MCH	pg	29.5	29.5			± 2.0	29.5	29.5			± 2.0	32.1	32.1			± 2.5			
	fmol	1.83	1.83			± 0.12	1.83	1.83			± 0.12	1.99	1.99			± 0.16			
CCMH MCHC	g/dl	35.6	35.6			± 3.0	36.4	36.4			± 3.0	36.9	36.9			± 3.0			
	g/l	356	356			± 30	364	364			± 30	369	369			± 30			
	mmol/l	22.10	22.10			± 1.86	22.59	22.59			± 1.86	22.91	22.91			± 1.86			
IDR RDW	%	13.0	12.0			± 4.0	13.2	12.3			± 4.0	13.0	11.5			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	65	65			± 20	236	236			± 30	496	496			± 50			
VPM MPV	µm <sup>3</sup> fl	10.4	10.4			± 2.0	10.0	10.0			± 2.0	9.5	9.5			± 2.0			
NEUT	#	1.43	1.43			± 0.35	4.60	4.60			± 0.90	12.81	12.81			± 1.90			
	%	54.9	54.9			± 10.0	59.0	59.0			± 10.0	70.0	70.0			± 10.0			
LYMPHO	#	0.76	0.76			± 0.33	2.30	2.30			± 0.70	2.84	2.84			± 1.50			
	%	29.2	29.2			± 12.0	29.5	29.5			± 8.0	15.5	15.5			± 8.0			
MONO	#	0.12	0.12			± 0.12	0.31	0.31			± 0.31	0.82	0.82			± 0.82			
	%	4.6	4.6			± 4.6	4.0	4.0			± 4.0	4.5	4.5			± 4.5			
EOS	#	0.21	0.21			± 0.21	0.32	0.32			± 0.32	1.04	1.04			± 1.04			
	%	8.0	8.0			± 8.0	4.1	4.1			± 4.1	5.7	5.7			± 5.7			
BASO	#	0.09	0.09			± 0.09	0.27	0.27			± 0.27	0.79	0.79			± 0.79			
	%	3.3	3.3			± 3.3	3.4	3.4			± 3.4	4.3	4.3			± 4.3			

**LOT** PX 421  
Rev 1

**CONTROL**

 (Exp.) 2020-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6	2.6		± 0.4	7.8	7.7	7.8		± 1.0	18.3	18.3	18.0		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43	2.40	2.40		± 0.16	4.48	4.45	4.46		± 0.20	5.12	5.09	5.10		± 0.25	
		g/dl	6.9	6.8	6.8		± 0.4	13.2	13.1	13.1		± 0.5	16.5	16.3	16.3		± 0.6	
HB	HGB	g/l	69	68	68		± 4	132	131	131		± 5	165	163	163		± 6	
		mmol/l	4.28	4.22	4.22		± 0.25	8.20	8.14	8.14		± 0.31	10.25	10.12	10.12		± 0.37	
HT	HCT	%	19.4	19.2	19.2		± 1.5	35.8	35.6	35.7		± 2.0	44.5	44.3	43.9		± 2.5	
		l/l	0.194	0.192	0.192		± 0.015	0.358	0.356	0.357		± 0.020	0.445	0.443	0.439		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	80	80	80.0		± 5	80	80	80.0		± 5	87	87	86.0		± 5	
TGMH	MCH	pg	28.4	28.3	28.3		± 2.0	29.5	29.4	29.4		± 2.0	32.2	32.0	32.0		± 2.5	
		fmol	1.76	1.76	1.76		± 0.12	1.83	1.83	1.82		± 0.12	2.00	1.99	1.98		± 0.16	
		g/dl	35.5	35.4	35.4		± 3.0	36.8	36.8	36.7		± 3.0	37.0	36.8	37.2		± 3.0	
CCMH	MCHC	g/l	355	354	354		± 30	368	368	367		± 30	370	368	372		± 30	
		mmol/l	22.04	21.99	21.99		± 1.86	22.87	22.85	22.80		± 1.86	23.00	22.86	23.08		± 1.86	
IDR	RDW	%	13.0	12.5	12.0		± 4.0	12.5	12.5	12.0		± 4.0	12.0	12.0	12.0		± 4.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	66	69		± 20	240	240	237		± 30	495	495	495		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.8	10.0	9.3		± 2.0	9.7	9.7	9.0		± 2.0	8.9	9.1	8.6		± 2.0	
		#	1.42	1.40	1.46		± 0.35	4.63	4.53	4.72		± 0.90	12.83	12.74	12.96		± 1.90	
		%	54.6	54.0	56.0		± 10.0	59.3	58.8	60.5		± 10.0	70.1	69.6	72.0		± 10.0	
		#	0.79	0.79	0.77		± 0.33	2.32	2.25	2.26		± 0.70	2.91	2.84	2.70		± 1.50	
		%	30.3	30.5	29.5		± 12.0	29.7	29.2	29.0		± 8.0	15.9	15.5	15.0		± 8.0	
		#	0.11	0.12	0.09		± 0.09	0.28	0.32	0.23		± 0.23	0.75	0.84	0.63		± 0.63	
		%	4.4	4.8	3.5		± 3.5	3.6	4.2	3.0		± 3.0	4.1	4.6	3.5		± 3.5	
		#	0.19	0.20	0.18		± 0.18	0.30	0.34	0.31		± 0.30	1.01	1.10	0.99		± 0.99	
		%	7.4	7.5	7.0		± 7.0	3.9	4.4	4.0		± 3.9	5.5	6.0	5.5		± 5.5	
		#	0.09	0.08	0.10		± 0.08	0.27	0.26	0.27		± 0.26	0.81	0.79	0.72		± 0.72	
		%	3.3	3.2	4.0		± 3.2	3.5	3.4	3.5		± 3.4	4.4	4.3	4.0		± 4.0	

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43