

LOT PX 425
Rev 1


CONTROL

 (Exp.) 2020-11-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.5	7.5			± 1.0	17.8	17.8			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.26	2.26			± 0.16	4.56	4.56			± 0.20	5.08	5.08			± 0.25			
HB HGB	g/dl	6.5	6.5			± 0.4	13.9	13.9			± 0.5	16.2	16.2			± 0.6			
	g/l	65	65			± 4	139	139			± 5	162	162			± 6			
	mmol/l	4.04	4.04			± 0.25	8.63	8.63			± 0.31	10.06	10.06			± 0.37			
HT HCT	%	18.8	18.8			± 1.5	38.3	38.3			± 2.0	44.7	44.7			± 2.5			
	l/l	0.188	0.188			± 0.015	0.383	0.383			± 0.020	0.447	0.447			± 0.025			
VGM MCV	µm ³ fl	83	83			± 5	84	84			± 5	88	88			± 5			
TGMH MCH	pg	28.8	28.8			± 2.0	30.5	30.5			± 2.0	31.9	31.9			± 2.5			
	fmol	1.79	1.79			± 0.12	1.89	1.89			± 0.12	1.98	1.98			± 0.16			
CCMH MCHC	g/dl	34.7	34.7			± 3.0	36.3	36.3			± 3.0	36.2	36.2			± 3.0			
	g/l	347	347			± 30	363	363			± 30	362	362			± 30			
	mmol/l	21.52	21.52			± 1.86	22.54	22.54			± 1.86	22.50	22.50			± 1.86			
IDR RDW	%	13.8	12.5			± 4.0	13.2	12.0			± 4.0	13.5	12.1			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65			± 20	256	256			± 30	490	490			± 50			
VPM MPV	µm ³ fl	9.2	9.2			± 2.0	9.6	9.6			± 2.0	9.4	9.4			± 2.0			
NEUT	#	1.38	1.38			± 0.35	4.16	4.16			± 0.90	12.10	12.10			± 1.90			
	%	55.3	55.3			± 10.0	55.4	55.4			± 10.0	68.0	68.0			± 10.0			
LYMPHO	#	0.74	0.74			± 0.33	2.28	2.28			± 0.70	2.81	2.81			± 1.50			
	%	29.4	29.4			± 12.0	30.4	30.4			± 8.0	15.8	15.8			± 8.0			
MONO	#	0.12	0.12			± 0.12	0.50	0.50			± 0.50	1.14	1.14			± 1.14			
	%	4.9	4.9			± 4.9	6.7	6.7			± 6.7	6.4	6.4			± 6.4			
EOS	#	0.18	0.18			± 0.18	0.30	0.30			± 0.30	0.98	0.98			± 0.98			
	%	7.1	7.1			± 7.1	4.0	4.0			± 4.0	5.5	5.5			± 5.5			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.77	0.77			± 0.77			
	%	3.3	3.3			± 3.3	3.5	3.5			± 3.5	4.3	4.3			± 4.3			

LOT PX 425
Rev 1

CONTROL

 (Exp.) 2020-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP					
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.4	7.5	7.4		± 1.0	17.9	18.0	17.9		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.27	2.27		± 0.16	4.55	4.52	4.56		± 0.20	5.07	5.00	5.12		± 0.25	
		g/dl	6.5	6.4	6.5		± 0.4	13.9	13.8	13.9		± 0.5	16.1	16.0	16.2		± 0.6	
HB	HGB	g/l	65	64	65		± 4	139	138	139		± 5	161	160	162		± 6	
		mmol/l	4.04	3.97	4.04		± 0.25	8.63	8.57	8.63		± 0.31	10.00	9.94	10.06		± 0.37	
HT	HCT	%	18.3	18.2	18.3		± 1.5	37.8	37.5	37.8		± 2.0	44.6	43.5	44.3		± 2.5	
		l/l	0.183	0.182	0.183		± 0.015	0.378	0.375	0.378		± 0.020	0.446	0.435	0.443		± 0.025	
VGM	MCV	µm ³ ; fl	80	80	80.5		± 5	83	83	83.0		± 5	88	87	86.5		± 5	
TGMH	MCH	pg	28.4	28.2	28.6		± 2.0	30.5	30.5	30.5		± 2.0	31.8	32.0	31.6		± 2.5	
		fmol	1.76	1.75	1.78		± 0.12	1.90	1.90	1.89		± 0.12	1.97	1.99	1.96		± 0.16	
		g/dl	35.5	35.2	35.6		± 3.0	36.8	36.8	36.7		± 3.0	36.1	36.8	36.6		± 3.0	
CCMH	MCHC	g/l	355	352	356		± 30	368	368	367		± 30	361	368	366		± 30	
		mmol/l	22.03	21.89	22.09		± 1.86	22.86	22.84	22.81		± 1.86	22.41	22.84	22.72		± 1.86	
IDR	RDW	%	13.5	13.0	13.0		± 4.0	12.5	12.5	12.0		± 4.0	12.7	12.7	12.1		± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	67	69		± 20	257	256	258		± 30	492	493	490		± 50	
VPM	MPV	µm ³ ; fl	9.2	9.0	8.6		± 2.0	9.5	9.3	8.9		± 2.0	9.4	9.2	8.7		± 2.0	
		#	1.37	1.32	1.38		± 0.35	4.18	4.07	4.26		± 0.90	12.35	12.19	12.58		± 1.90	
		%	57.0	55.1	57.5		± 10.0	56.5	54.3	57.5		± 10.0	69.0	67.7	70.3		± 10.0	
		#	0.71	0.74	0.70		± 0.33	2.33	2.45	2.28		± 0.70	2.95	3.01	2.83		± 1.50	
		%	29.5	30.7	29.0		± 12.0	31.5	32.7	30.8		± 8.0	16.5	16.7	15.8		± 8.0	
		#	0.10	0.11	0.10		± 0.10	0.37	0.44	0.35		± 0.35	0.90	1.06	0.79		± 0.79	
		%	4.0	4.7	4.0		± 4.0	5.0	5.9	4.7		± 4.7	5.0	5.9	4.4		± 4.4	
		#	0.16	0.15	0.14		± 0.14	0.26	0.26	0.25		± 0.25	0.90	0.95	0.95		± 0.90	
		%	6.5	6.2	6.0		± 6.0	3.5	3.5	3.4		± 3.4	5.0	5.3	5.3		± 5.0	
		#	0.07	0.08	0.08		± 0.07	0.26	0.27	0.27		± 0.26	0.81	0.79	0.75		± 0.75	
		%	3.0	3.3	3.5		± 3.0	3.5	3.6	3.6		± 3.5	4.5	4.4	4.2		± 4.2	

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44