


LOT PX 435
Rev 1


CONTROL

 (Exp.) **2022-07-05**
(YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XL9					XL80	XL9					XL80	XL9				
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	3.1	3.1			± 0.4	8.6	8.6			± 1.0	17.9	17.9			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.39			± 0.16	4.53	4.53			± 0.20	4.97	4.97			± 0.25			
HB HGB	g/dl	6.4	6.4			± 0.4	13.6	13.6			± 0.5	16.0	16.0			± 0.6			
	g/l	64	64			± 4	136	136			± 5	160	160			± 6			
HT HCT	mmol/l	3.97	3.97			± 0.25	8.45	8.45			± 0.31	9.94	9.94			± 0.37			
	%	19.6	19.6			± 1.5	39.9	39.9			± 2.0	46.7	46.7			± 2.5			
VGM MCV	l/l	0.196	0.196			± 0.015	0.399	0.399			± 0.020	0.467	0.467			± 0.025			
	µm ³ /fl	82	82			± 5	88	88			± 5	94	94			± 5			
TGMH MCH	pg	26.8	26.8			± 2.0	30.0	30.0			± 2.0	32.2	32.2			± 2.5			
	fmol	1.66	1.66			± 0.12	1.86	1.86			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	32.7	32.7			± 3.0	34.1	34.1			± 3.0	34.2	34.2			± 3.0			
	g/l	327	327			± 30	341	341			± 30	342	342			± 30			
IDR RDW	mmol/l	20.28	20.28			± 1.86	21.19	21.19			± 1.86	21.27	21.27			± 1.86			
	%	15.0	15.0			± 5.0	13.0	13.0			± 5.0	12.0	12.0			± 5.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74			± 20	221	221			± 30	457	457			± 50			
VPM MPV	µm ³ /fl	10.1	10.1			± 2.0	10.9	10.9			± 2.0	10.4	10.4			± 2.0			
NEUT	#	1.56	1.56			± 0.35	4.45	4.45			± 0.90	12.55	12.55			± 1.90			
	%	50.4	50.4			± 10.0	51.8	51.8			± 10.0	70.1	70.1			± 10.0			
LYMPHO	#	1.10	1.10			± 0.33	3.27	3.27			± 0.70	3.40	3.40			± 1.50			
	%	35.4	35.4			± 12.0	38.0	38.0			± 8.0	19.0	19.0			± 8.0			
MONO	#	0.20	0.20			± 0.20	0.32	0.32			± 0.32	0.38	0.38			± 0.38			
	%	6.5	6.5			± 6.5	3.7	3.7			± 3.7	2.1	2.1			± 2.1			
EOS	#	0.16	0.16			± 0.16	0.31	0.31			± 0.31	0.86	0.86			± 0.86			
	%	5.0	5.0			± 5.0	3.6	3.6			± 3.6	4.8	4.8			± 4.8			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	2.7	2.7			± 2.7	2.9	2.9			± 2.9	4.0	4.0			± 4.0			

LOT PX 435
Rev 1

CONTROL

 (Exp.) **2022-07-05**
(YYYY-MM-DD)

		ABX Lysebio													
CONTROL				L	CONTROL				N	CONTROL				H	
PENTRA				PENTRA				PENTRA							

PARAMETRES PARAMETERS	UNITES UNITS	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1	3.1	± 0.4	8.8	8.7	8.4	± 1.0	18.5	18.0	18.0	± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.47	2.45	2.41	± 0.16	4.55	4.55	4.50	± 0.20	5.01	4.96	4.97	± 0.25
HB HGB	g/dl	6.4	6.3	6.4	± 0.4	13.6	13.5	13.6	± 0.5	16.0	15.7	15.9	± 0.6
	g/l	64	63	64	± 4	136	135	136	± 5	160	157	159	± 6
HT HCT	mmol/l	3.97	3.91	3.97	± 0.25	8.45	8.38	8.45	± 0.31	9.94	9.75	9.87	± 0.37
	%	19.5	19.1	18.8	± 1.5	39.6	39.1	38.7	± 2.0	46.6	46.1	46.2	± 2.5
VGM MCV	l/l	0.195	0.191	0.188	± 0.015	0.396	0.391	0.387	± 0.020	0.466	0.461	0.462	± 0.025
	µm ³ ; fl	79	78	78	± 5	87	86	86	± 5	93	93	93	± 5
TGMH MCH	pg	25.9	25.7	26.6	± 2.0	29.9	29.7	30.2	± 2.0	31.9	31.7	32.0	± 2.5
	fmol	1.61	1.60	1.65	± 0.12	1.86	1.84	1.88	± 0.12	1.98	1.97	1.99	± 0.16
CCMH MCHC	g/dl	32.8	33.0	34.0	± 3.0	34.4	34.5	35.1	± 3.0	34.3	34.0	34.4	± 3.0
	g/l	328	330	340	± 30	344	345	351	± 30	343	340	344	± 30
IDR RDW	mmol/l	20.37	20.47	21.14	± 1.86	21.34	21.42	21.82	± 1.86	21.33	21.14	21.36	± 1.86
	%	15.5	15.5	14.0	± 4.0	12.5	12.5	11.5	± 4.0	12.0	12.0	11.0	± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	76	75	± 20	223	225	219	± 30	455	465	445	± 50
VPM MPV	µm ³ ; fl	9.8	10.0	9.0	± 2.0	10.2	10.5	9.5	± 2.0	9.7	10.0	9.0	± 2.0
NEUT	#	1.53	1.55	1.52	± 0.35	4.44	4.49	4.33	± 0.90	12.84	12.62	12.60	± 1.90
	%	49.2	50.1	49.0	± 10.0	50.4	51.6	51.5	± 10.0	69.4	70.1	70.0	± 10.0
LYMPHO	#	1.13	1.09	1.13	± 0.33	3.48	3.31	3.28	± 0.70	3.64	3.37	3.60	± 1.50
	%	36.3	35.1	36.5	± 12.0	39.5	38.0	39.0	± 8.0	19.7	18.7	20.0	± 8.0
MONO	#	0.21	0.21	0.20	± 0.20	0.34	0.34	0.29	± 0.29	0.41	0.38	0.27	± 0.27
	%	6.7	6.8	6.5	± 6.5	3.9	3.9	3.5	± 3.5	2.2	2.1	1.5	± 1.5
EOS	#	0.16	0.16	0.16	± 0.16	0.29	0.31	0.25	± 0.25	0.87	0.90	0.81	± 0.81
	%	5.1	5.3	5.0	± 5.0	3.3	3.6	3.0	± 3.0	4.7	5.0	4.5	± 4.5
BASO	#	0.08	0.08	0.09	± 0.08	0.26	0.25	0.25	± 0.25	0.74	0.74	0.72	± 0.72
	%	2.7	2.7	3.0	± 2.7	2.9	2.9	3.0	± 2.9	4.0	4.1	4.0	± 4.0

Ref: TEMP-0821 Rev.45 FRONT / RECTO 1300090306