

ABX Difftrol



LOT PX 406
Rev 1

CONTROL

(Exp.) 2017-09-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.6	2.6	2.6	2.6	± 0.4	7.8	7.8	7.9	7.8	7.9	± 1.0	19.0	19.0	19.1	19.0	19.5	± 2.2		
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.24	2.30	2.24	2.28	± 0.16	4.51	4.53	4.51	4.53	4.52	± 0.20	5.08	5.11	5.06	5.11	5.10	± 0.25		
HB	HGB	g/dl	6.6	6.5	6.6	6.5	6.5	± 0.4	13.1	13.1	13.1	13.1	13.1	± 0.5	16.1	16.1	16.2	16.1	16.0	± 0.6		
		g/l	66	65	66	65	65	± 4	131	131	131	131	131	± 5	161	161	162	161	160	± 6		
		mmol/l	4.10	4.04	4.10	4.04	4.04	± 0.25	8.14	8.14	8.14	8.14	8.14	± 0.31	10.00	10.00	10.06	10.00	9.94	± 0.37		
HT	HCT	%	18.6	18.8	18.4	18.8	18.5	± 1.5	36.1	36.7	36.1	36.7	36.2	± 2.0	44.7	45.0	44.0	44.5	43.9	± 2.5		
		l/l	0.186	0.188	0.184	0.188	0.185	± 0.015	0.361	0.367	0.361	0.367	0.362	± 0.020	0.447	0.450	0.440	0.445	0.439	± 0.025		
VGM	MCV	µm ³ ; fl	81	84	80	84	81.0	± 5	80	81	80	81	80.0	± 5	88	88	87	87	86.0	± 5		
TGMH	MCH	pg	28.7	29.0	28.7	29.0	28.5	± 2.0	29.0	28.9	29.0	28.9	29.0	± 2.0	31.7	31.5	32.0	31.5	31.4	± 2.5		
		fmol	1.78	1.80	1.78	1.80	1.77	± 0.12	1.80	1.80	1.80	1.80	1.80	± 0.12	1.97	1.96	1.99	1.96	1.95	± 0.16		
CCMH	MCHC	g/dl	35.4	34.5	35.9	34.5	35.2	± 3.0	36.3	35.7	36.3	35.7	36.2	± 3.0	36.0	35.8	36.8	36.2	36.5	± 3.0		
		g/l	354	345	359	345	352	± 30	363	357	363	357	362	± 30	360	358	368	362	365	± 30		
		mmol/l	22.00	21.45	22.28	21.45	21.86	± 1.86	22.55	22.17	22.55	22.17	22.50	± 1.86	22.37	22.23	22.85	22.49	22.65	± 1.86		
IDR	RDW	%	13.4	13.5	12.8	13.5	12.3	± 4.0	13.0	13.5	12.8	13.5	12.5	± 4.0	12.6	13.0	12.0	13.0	11.8	± 4.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	67	68	67	70	± 20	240	240	247	240	240	± 30	490	490	507	490	485	± 50		
VPM	MPV	µm ³ ; fl	9.5	9.7	9.6	9.7	9.0	± 2.0	9.2	9.5	9.2	9.5	8.9	± 2.0	8.5	8.7	8.6	8.7	8.2	± 2.0		
NEUT		#	1.49	1.55	1.55	1.55	1.53	± 0.35	4.42	4.50	4.45	4.50	4.38	± 0.90	13.22	12.98	13.26	12.98	13.65	± 1.90		
		%	59.4	59.8	59.6	59.8	58.9	± 10.0	56.7	57.7	56.3	57.7	55.4	± 10.0	69.6	68.3	69.4	68.3	70.0	± 10.0		
LYMPHO		#	0.72	0.71	0.75	0.71	0.78	± 0.33	2.48	2.32	2.54	2.32	2.63	± 0.70	3.00	2.98	2.87	2.98	3.08	± 1.50		
		%	28.6	27.4	28.7	27.4	30.0	± 12.0	31.8	29.8	32.1	29.8	33.3	± 8.0	15.8	15.7	15.0	15.7	15.8	± 8.0		
MONO		#	0.09	0.11	0.09	0.11	0.09	± 0.09	0.36	0.44	0.36	0.44	0.40	± 0.36	0.95	1.20	1.09	1.20	1.05	± 0.95		
		%	3.5	4.2	3.3	4.2	3.6	± 3.3	4.6	5.6	4.6	5.6	5.1	± 4.6	5.0	6.3	5.7	6.3	5.4	± 5.0		
EOS		#	0.13	0.14	0.13	0.14	0.10	± 0.10	0.27	0.27	0.28	0.27	0.22	± 0.22	0.97	1.01	1.05	1.01	0.88	± 0.88		
		%	5.1	5.3	5.0	5.3	4.0	± 4.0	3.4	3.5	3.5	3.5	2.8	± 2.8	5.1	5.3	5.5	5.3	4.5	± 4.5		
BASO		#	0.09	0.09	0.09	0.09	0.09	± 0.09	0.27	0.27	0.28	0.27	0.27	± 0.27	0.86	0.84	0.84	0.84	0.84	± 0.84		
		%	3.4	3.3	3.4	3.3	3.5	± 3.3	3.5	3.4	3.5	3.4	3.4	± 3.4	4.5	4.4	4.4	4.4	4.3	± 4.3		

ABX Difftrol



LOT PX 406
Rev 1

CONTROL

(Exp.) 2017-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		ABX Lysebio																					
		CONTROL			L		CONTROL			N		CONTROL			H		TOLERANCES TOLERANCE						
		PENTRA						PENTRA						PENTRA									
		120	DX120	DX NEXUS				120	DX120	DX NEXUS				120	DX120	DX NEXUS							
120 RETIC	DF120	DF NEXUS				120 RETIC	DF120	DF NEXUS				120 RETIC	DF120	DF NEXUS									
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6				± 0.4	8.0	8.0	8.0				± 1.0	20.5	20.5	20.5				± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.30	2.30	2.30				± 0.16	4.54	4.54	4.54				± 0.20	5.14	5.14	5.14				± 0.25
HB	HGB	g/dl	6.6	6.6	6.6				± 0.4	13.1	13.1	13.1				± 0.5	16.0	16.0	16.0				± 0.6
		g/l	66	66	66				± 4	131	131	131				± 5	160	160	160				± 6
		mmol/l	4.10	4.10	4.10				± 0.25	8.14	8.14	8.14				± 0.31	9.94	9.94	9.94				± 0.37
HT	HCT	%	19.3	19.3	19.6				± 1.5	36.8	36.8	37.2				± 2.0	44.7	44.7	45.2				± 2.5
		l/l	0.193	0.193	0.196				± 0.015	0.368	0.368	0.372				± 0.020	0.447	0.447	0.452				± 0.025
VGM	MCV	µm ³ ; fl	84	84	85				± 5	81	81	82				± 5	87	87	88				± 5
TGMH	MCH	pg	28.7	28.7	28.7				± 2.0	28.9	28.9	28.9				± 2.0	31.1	31.1	31.1				± 2.5
		fmol	1.78	1.78	1.78				± 0.12	1.79	1.79	1.79				± 0.12	1.93	1.93	1.93				± 0.16
CCMH	MCHC	g/dl	34.2	34.2	33.8				± 3.0	35.6	35.6	35.2				± 3.0	35.8	35.8	35.4				± 3.0
		g/l	342	342	338				± 30	356	356	352				± 30	358	358	354				± 30
		mmol/l	21.21	21.21	20.96				± 1.86	22.12	22.12	21.85				± 1.86	22.22	22.22	21.97				± 1.86
IDR	RDW	%	15.0	15.0	15.0				± 4.0	16.0	16.0	16.0				± 4.0	15.0	15.0	15.0				± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70				± 20	245	245	245				± 30	488	488	488				± 50
VPM	MPV	µm ³ ; fl	9.7	9.7	9.7				± 2.0	9.4	9.4	9.4				± 2.0	8.7	8.7	8.7				± 2.0
NEUT		#	1.62	1.67	1.67				± 0.35	4.61	4.72	4.72				± 0.90	14.50	14.70	14.70				± 1.90
		%	62.2	64.1	64.1				± 10.0	57.6	59.0	59.0				± 10.0	70.6	71.6	71.6				± 10.0
LYMPHO		#	0.67	0.61	0.61				± 0.33	2.34	2.19	2.19				± 0.70	2.95	2.75	2.75				± 1.50
		%	25.8	23.5	23.5				± 12.0	29.3	27.4	27.4				± 8.0	14.4	13.4	13.4				± 8.0
MONO		#	0.12	0.12	0.12				± 0.12	0.53	0.54	0.54				± 0.53	1.64	1.62	1.62				± 1.62
		%	4.7	4.8	4.8				± 4.7	6.6	6.8	6.8				± 6.6	8.0	7.9	7.9				± 7.9
EOS		#	0.12	0.13	0.13				± 0.12	0.28	0.30	0.30				± 0.28	0.92	0.94	0.94				± 0.92
		%	4.8	5.1	5.1				± 4.8	3.5	3.8	3.8				± 3.5	4.5	4.6	4.6				± 4.5
BASO		#	0.07	0.07	0.07				± 0.07	0.24	0.24	0.24				± 0.24	0.51	0.51	0.51				± 0.51
		%	2.5	2.5	2.5				± 2.5	3.0	3.0	3.0				± 3.0	2.5	2.5	2.5				± 2.5

Ref: TEMP-0821 Rev.40 FRONT / RECTO 1300032435-A