

LOT PX 426
Rev 2


CONTROL

 (Exp.) 2021-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					
			80 XL80	XLR				80 XL80	XLR				80 XL80	XLR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8				± 0.4	7.8	7.8				± 1.0	18.0	18.0		
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.23	2.23			± 0.16	4.52	4.52			± 0.20	5.00	5.00			± 0.25	
HB	HGB	g/dl	6.3	6.3			± 0.4	13.6	13.6			± 0.5	16.2	16.2			± 0.6	
		g/l	63	63			± 4	136	136			± 5	162	162			± 6	
		mmol/l	3.91	3.91			± 0.25	8.45	8.45			± 0.31	10.06	10.06			± 0.37	
HT	HCT	%	18.3	18.3			± 1.5	37.5	37.5			± 2.0	44.5	44.5			± 2.5	
		l/l	0.183	0.183			± 0.015	0.375	0.375			± 0.020	0.445	0.445			± 0.025	
VGM	MCV	µm ³ ; fl	82	82			± 5	83	83			± 5	89	89			± 5	
TGMH	MCH	pg	28.3	28.3			± 2.0	30.1	30.1			± 2.0	32.4	32.4			± 2.5	
		fmol	1.75	1.75			± 0.12	1.87	1.87			± 0.12	2.01	2.01			± 0.16	
CCMH	MCHC	g/dl	34.5	34.5			± 3.0	36.3	36.3			± 3.0	36.4	36.4			± 3.0	
		g/l	345	345			± 30	363	363			± 30	364	364			± 30	
		mmol/l	21.40	21.40			± 1.86	22.51	22.51			± 1.86	22.61	22.61			± 1.86	
IDR	RDW	%	14.0	13.0			± 4.0	13.3	12.5			± 4.0	13.0	12.0			± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66			± 20	245	245			± 30	490	490			± 50	
VPM	MPV	µm ³ ; fl	9.9	9.9			± 2.0	8.9	8.9			± 2.0	8.9	8.9			± 2.0	
NEUT		#	1.47	1.47			± 0.35	4.05	4.05			± 0.90	12.02	12.02			± 1.90	
		%	52.4	52.4			± 10.0	51.9	51.9			± 10.0	66.8	66.8			± 10.0	
LYMPHO		#	0.80	0.80			± 0.33	2.35	2.35			± 0.70	2.84	2.84			± 1.50	
		%	28.6	28.6			± 12.0	30.1	30.1			± 8.0	15.8	15.8			± 8.0	
MONO		#	0.23	0.23			± 0.23	0.73	0.73			± 0.73	1.46	1.46			± 1.46	
		%	8.1	8.1			± 8.1	9.3	9.3			± 9.3	8.1	8.1			± 8.1	
EOS		#	0.21	0.21			± 0.21	0.38	0.38			± 0.38	0.86	0.86			± 0.86	
		%	7.5	7.5			± 7.5	4.9	4.9			± 4.9	4.8	4.8			± 4.8	
BASO		#	0.10	0.10			± 0.10	0.30	0.30			± 0.30	0.81	0.81			± 0.81	
		%	3.4	3.4			± 3.4	3.8	3.8			± 3.8	4.5	4.5			± 4.5	

LOT PX 426
Rev 2

CONTROL

 (Exp.) 2021-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60	MS60	MSCRP			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.7		± 0.4	7.8	7.9	7.8		± 1.0	18.5	18.6	18.5		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.28	2.26	2.26		± 0.16	4.50	4.49	4.50		± 0.20	5.00	4.96	5.00		± 0.25			
HB HGB	g/dl	6.4	6.3	6.4		± 0.4	13.6	13.4	13.5		± 0.5	16.3	16.0	16.2		± 0.6			
	g/l	64	63	64		± 4	136	134	135		± 5	163	160	162		± 6			
HT HCT	mmol/l	3.97	3.91	3.97		± 0.25	8.45	8.32	8.38		± 0.31	10.12	9.94	10.06		± 0.37			
	%	18.2	17.9	17.7		± 1.5	37.4	37.3	37.4		± 2.0	45.0	44.1	44.0		± 2.5			
VGM MCV	l/l	0.182	0.179	0.177		± 0.015	0.374	0.373	0.374		± 0.020	0.450	0.441	0.440		± 0.025			
	µm ³ /fl	80	79	78.5		± 5	83	83	83.0		± 5	90	89	88.0		± 5			
TGMH MCH	pg	28.1	27.9	28.3		± 2.0	30.2	29.8	30.0		± 2.0	32.6	32.3	32.4		± 2.5			
	fmol	1.74	1.73	1.76		± 0.12	1.88	1.85	1.86		± 0.12	2.02	2.00	2.01		± 0.16			
CCMH MCHC	g/dl	35.1	35.3	36.1		± 3.0	36.4	36.0	36.1		± 3.0	36.2	36.2	36.8		± 3.0			
	g/l	351	353	361		± 30	364	360	361		± 30	362	362	368		± 30			
IDR RDW	mmol/l	21.79	21.91	22.40		± 1.86	22.61	22.33	22.45		± 1.86	22.49	22.51	22.86		± 1.86			
	%	13.5	13.0	13.2		± 4.0	12.5	12.0	12.0		± 4.0	12.0	12.0	11.5		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	67	65		± 20	247	247	244		± 30	487	490	486		± 50			
VPM MPV	µm ³ /fl	9.7	9.9	9.2		± 2.0	8.6	8.6	8.3		± 2.0	8.6	8.6	8.3		± 2.0			
NEUT	#	1.48	1.44	1.50		± 0.35	4.17	4.01	4.35		± 0.90	12.56	12.03	12.67		± 1.90			
	%	52.7	51.4	55.5		± 10.0	53.5	50.7	55.8		± 10.0	67.9	64.7	68.5		± 10.0			
LYMPHO	#	0.79	0.80	0.75		± 0.33	2.31	2.31	2.18		± 0.70	2.87	2.85	2.68		± 1.50			
	%	28.3	28.7	27.7		± 12.0	29.6	29.3	28.0		± 8.0	15.5	15.3	14.5		± 8.0			
MONO	#	0.22	0.27	0.18		± 0.18	0.65	0.92	0.62		± 0.62	1.33	1.97	1.41		± 1.33			
	%	7.7	9.5	6.5		± 6.5	8.3	11.6	8.0		± 8.0	7.2	10.6	7.6		± 7.2			
EOS	#	0.22	0.20	0.18		± 0.18	0.37	0.36	0.35		± 0.35	0.89	0.89	0.93		± 0.89			
	%	7.9	7.0	6.8		± 6.8	4.8	4.6	4.5		± 4.5	4.8	4.8	5.0		± 4.8			
BASO	#	0.10	0.10	0.09		± 0.09	0.30	0.30	0.29		± 0.29	0.85	0.86	0.81		± 0.81			
	%	3.4	3.4	3.5		± 3.4	3.8	3.8	3.7		± 3.7	4.6	4.6	4.4		± 4.4			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44