

LOT PX 417
Rev 3


CONTROL

 (Exp.) 2019-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.3	7.3			± 1.0	18.0	18.0			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.29	2.29			± 0.16	4.59	4.59			± 0.20	5.19	5.19			± 0.25			
HB HGB	g/dl	6.6	6.6			± 0.4	13.9	13.9			± 0.5	16.6	16.6			± 0.6			
	g/l	66	66			± 4	139	139			± 5	166	166			± 6			
	mmol/l	4.10	4.10			± 0.25	8.63	8.63			± 0.31	10.31	10.31			± 0.37			
HT HCT	%	19.0	19.0			± 1.5	39.0	39.0			± 2.0	45.7	45.7			± 2.5			
	l/l	0.190	0.190			± 0.015	0.390	0.390			± 0.020	0.457	0.457			± 0.025			
VGM MCV	µm ³ fl	83	83			± 5	85	85			± 5	88	88			± 5			
TGMH MCH	pg	28.8	28.8			± 2.0	30.3	30.3			± 2.0	32.0	32.0			± 2.5			
	fmol	1.79	1.79			± 0.12	1.88	1.88			± 0.12	1.99	1.99			± 0.16			
CCMH MCHC	g/dl	34.7	34.7			± 3.0	35.6	35.6			± 3.0	36.3	36.3			± 3.0			
	g/l	347	347			± 30	356	356			± 30	363	363			± 30			
	mmol/l	21.56	21.56			± 1.86	22.12	22.12			± 1.86	22.57	22.57			± 1.86			
IDR RDW	%	13.0	11.5			± 4.0	13.5	12.5			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	245	245			± 30	495	495			± 50			
VPM MPV	µm ³ fl	10.0	10.0			± 2.0	10.0	10.3			± 2.0	9.6	9.9			± 2.0			
NEUT	#	1.43	1.43			± 0.35	4.02	4.02			± 0.90	12.44	12.44			± 1.90			
	%	57.1	57.1			± 10.0	55.1	55.1			± 10.0	69.1	69.1			± 10.0			
LYMPHO	#	0.75	0.75			± 0.33	2.37	2.37			± 0.70	2.79	2.79			± 1.50			
	%	30.1	30.1			± 12.0	32.4	32.4			± 8.0	15.5	15.5			± 8.0			
MONO	#	0.11	0.11			± 0.11	0.30	0.30			± 0.30	0.88	0.88			± 0.88			
	%	4.3	4.3			± 4.3	4.1	4.1			± 4.1	4.9	4.9			± 4.9			
EOS	#	0.13	0.13			± 0.13	0.35	0.35			± 0.35	1.08	1.08			± 1.08			
	%	5.2	5.2			± 5.2	4.8	4.8			± 4.8	6.0	6.0			± 6.0			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.81	0.81			± 0.81			
	%	3.3	3.3			± 3.3	3.6	3.6			± 3.6	4.5	4.5			± 4.5			

LOT PX 417
Rev 3

CONTROL

 (Exp.) 2019-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5		± 0.4	7.3	7.2	7.4		± 1.0	18.3	18.0	18.4		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.31	2.32		± 0.16	4.59	4.57	4.60		± 0.20	5.15	5.16	5.19		± 0.25			
	g/dl	6.6	6.6	6.6		± 0.4	13.9	13.8	13.8		± 0.5	16.5	16.4	16.4		± 0.6			
HB HGB	g/l	66	66	66		± 4	139	138	138		± 5	165	164	164		± 6			
	mmol/l	4.10	4.10	4.10		± 0.25	8.63	8.57	8.57		± 0.31	10.25	10.18	10.18		± 0.37			
HT HCT	%	18.8	18.5	18.6		± 1.5	38.6	37.9	38.0		± 2.0	45.3	45.4	44.9		± 2.5			
	l/l	0.188	0.185	0.186		± 0.015	0.386	0.379	0.380		± 0.020	0.453	0.454	0.449		± 0.025			
VGM MCV	µm ³ ; fl	80	80	80.0		± 5	84	83	82.6		± 5	88	88	86.5		± 5			
TGMH MCH	pg	28.1	28.6	28.4		± 2.0	30.3	30.2	30.0		± 2.0	32.0	31.8	31.6		± 2.5			
	fmol	1.74	1.77	1.77		± 0.12	1.88	1.88	1.86		± 0.12	1.99	1.97	1.96		± 0.16			
	g/dl	35.1	35.7	35.6		± 3.0	36.1	36.4	36.3		± 3.0	36.4	36.1	36.5		± 3.0			
CCMH MCHC	g/l	351	357	356		± 30	361	364	363		± 30	364	361	365		± 30			
	mmol/l	21.80	22.18	22.08		± 1.86	22.39	22.59	22.55		± 1.86	22.61	22.43	22.69		± 1.86			
IDR RDW	%	12.5	13.0	12.5		± 4.0	13.0	12.5	12.0		± 4.0	12.8	12.5	12.5		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	66	70		± 20	245	245	240		± 30	485	495	475		± 50			
VPM MPV	µm ³ ; fl	9.7	9.8	9.2		± 2.0	9.8	9.7	9.3		± 2.0	9.5	9.6	8.7		± 2.0			
NEUT	#	1.35	1.37	1.44		± 0.35	4.09	4.02	4.13		± 0.90	12.83	12.60	12.90		± 1.90			
	%	56.4	57.0	57.6		± 10.0	56.0	55.9	55.8		± 10.0	70.1	70.0	70.1		± 10.0			
LYMPHO	#	0.73	0.71	0.76		± 0.33	2.40	2.34	2.48		± 0.70	2.76	2.65	2.94		± 1.50			
	%	30.4	29.5	30.5		± 12.0	32.9	32.5	33.5		± 8.0	15.1	14.7	16.0		± 8.0			
MONO	#	0.11	0.12	0.10		± 0.10	0.30	0.35	0.30		± 0.30	0.93	0.99	0.83		± 0.83			
	%	4.5	4.8	4.0		± 4.0	4.1	4.8	4.0		± 4.0	5.1	5.5	4.5		± 4.5			
EOS	#	0.13	0.13	0.12		± 0.12	0.25	0.24	0.24		± 0.24	0.95	0.97	0.92		± 0.92			
	%	5.4	5.4	4.6		± 4.6	3.4	3.4	3.2		± 3.2	5.2	5.4	5.0		± 5.0			
BASO	#	0.08	0.08	0.08		± 0.08	0.26	0.24	0.26		± 0.24	0.82	0.79	0.81		± 0.79			
	%	3.3	3.3	3.3		± 3.3	3.6	3.4	3.5		± 3.4	4.5	4.4	4.4		± 4.4			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43