


**LOT** PX 418  
Rev 2


**CONTROL**

 (Exp.) 2019-09-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5			± 0.4	7.5	7.5			± 1.0	18.0	18.0			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.27	2.27			± 0.16	4.58	4.58			± 0.20	5.21	5.21			± 0.25			
HB HGB	g/dl	6.8	6.8			± 0.4	14.0	14.0			± 0.5	16.8	16.8			± 0.6			
	g/l	68	68			± 4	140	140			± 5	168	168			± 6			
	mmol/l	4.22	4.22			± 0.25	8.69	8.69			± 0.31	10.43	10.43			± 0.37			
HT HCT	%	19.5	19.5			± 1.5	38.9	38.9			± 2.0	45.8	45.8			± 2.5			
	l/l	0.195	0.195			± 0.015	0.389	0.389			± 0.020	0.458	0.458			± 0.025			
VGM MCV	µm <sup>3</sup> fl	86	86			± 5	85	85			± 5	88	88			± 5			
TGMH MCH	pg	30.0	30.0			± 2.0	30.6	30.6			± 2.0	32.2	32.2			± 2.5			
	fmol	1.86	1.86			± 0.12	1.90	1.90			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	34.8	34.8			± 3.0	36.0	36.0			± 3.0	36.6	36.6			± 3.0			
	g/l	348	348			± 30	360	360			± 30	366	366			± 30			
	mmol/l	21.63	21.63			± 1.86	22.33	22.33			± 1.86	22.76	22.76			± 1.86			
IDR RDW	%	13.5	11.5			± 4.0	13.3	11.5			± 4.0	13.0	11.8			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	67			± 20	246	246			± 30	503	503			± 50			
VPM MPV	µm <sup>3</sup> fl	9.5	9.9			± 2.0	9.2	9.7			± 2.0	9.5	9.9			± 2.0			
NEUT	#	1.45	1.45			± 0.35	4.18	4.18			± 0.90	12.53	12.53			± 1.90			
	%	57.8	57.8			± 10.0	55.7	55.7			± 10.0	69.6	69.6			± 10.0			
LYMPHO	#	0.72	0.72			± 0.33	2.45	2.45			± 0.70	2.93	2.93			± 1.50			
	%	28.7	28.7			± 12.0	32.6	32.6			± 8.0	16.3	16.3			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.35	0.35			± 0.35	0.79	0.79			± 0.79			
	%	3.8	3.8			± 3.8	4.6	4.6			± 4.6	4.4	4.4			± 4.4			
EOS	#	0.16	0.16			± 0.16	0.27	0.27			± 0.27	0.95	0.95			± 0.95			
	%	6.3	6.3			± 6.3	3.6	3.6			± 3.6	5.3	5.3			± 5.3			
BASO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.79	0.79			± 0.79			
	%	3.4	3.4			± 3.4	3.5	3.5			± 3.5	4.4	4.4			± 4.4			

**LOT** PX 418  
Rev 2

**CONTROL**

 (Exp.) 2019-09-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5		± 0.4	7.5	7.4	7.5		± 1.0	18.2	17.8	18.0		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.33	2.31	2.31		± 0.16	4.58	4.55	4.56		± 0.20	5.16	5.18	5.19		± 0.25			
	g/dl	6.8	6.7	6.7		± 0.4	13.9	13.8	13.9		± 0.5	16.7	16.5	16.6		± 0.6			
HB HGB	g/l	68	67	67		± 4	139	138	139		± 5	167	165	166		± 6			
	mmol/l	4.22	4.16	4.16		± 0.25	8.63	8.57	8.63		± 0.31	10.37	10.25	10.31		± 0.37			
HT HCT	%	19.3	18.9	18.9		± 1.5	38.0	37.3	38.1		± 2.0	45.4	45.1	45.2		± 2.5			
	l/l	0.193	0.189	0.189		± 0.015	0.380	0.373	0.381		± 0.020	0.454	0.451	0.452		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	82	82.0		± 5	83	82	83.5		± 5	88	87	87.0		± 5			
TGMH MCH	pg	29.2	29.0	29.0		± 2.0	30.3	30.3	30.5		± 2.0	32.4	31.9	32.0		± 2.5			
	fmol	1.81	1.80	1.80		± 0.12	1.88	1.88	1.89		± 0.12	2.01	1.98	1.99		± 0.16			
	g/dl	35.2	35.4	35.4		± 3.0	36.6	37.0	36.5		± 3.0	36.8	36.6	36.8		± 3.0			
CCMH MCHC	g/l	352	354	354		± 30	366	370	365		± 30	368	366	368		± 30			
	mmol/l	21.84	21.97	21.97		± 1.86	22.71	22.97	22.67		± 1.86	22.84	22.74	22.83		± 1.86			
IDR RDW	%	12.7	12.7	12.5		± 4.0	12.5	12.0	12.0		± 4.0	12.3	12.0	12.5		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	71	69	69		± 20	250	250	248		± 30	503	505	495		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.3	9.5	9.0		± 2.0	9.2	9.2	8.6		± 2.0	9.4	9.3	8.7		± 2.0			
NEUT	#	1.41	1.39	1.45		± 0.35	4.20	4.12	4.24		± 0.90	12.72	12.34	12.69		± 1.90			
	%	58.7	57.9	58.0		± 10.0	56.0	55.7	56.5		± 10.0	69.9	69.3	70.5		± 10.0			
LYMPHO	#	0.66	0.66	0.73		± 0.33	2.42	2.31	2.48		± 0.70	2.93	2.74	2.97		± 1.50			
	%	27.5	27.4	29.0		± 12.0	32.3	31.2	33.0		± 8.0	16.1	15.4	16.5		± 8.0			
MONO	#	0.09	0.11	0.09		± 0.09	0.32	0.42	0.34		± 0.32	0.80	1.01	0.72		± 0.72			
	%	3.6	4.6	3.5		± 3.5	4.3	5.7	4.5		± 4.3	4.4	5.7	4.0		± 4.0			
EOS	#	0.16	0.16	0.15		± 0.15	0.29	0.30	0.23		± 0.23	0.95	0.94	0.81		± 0.81			
	%	6.7	6.6	6.0		± 6.0	3.8	4.0	3.0		± 3.0	5.2	5.3	4.5		± 4.5			
BASO	#	0.08	0.08	0.09		± 0.08	0.27	0.25	0.23		± 0.23	0.80	0.77	0.81		± 0.77			
	%	3.5	3.5	3.5		± 3.5	3.6	3.4	3.0		± 3.0	4.4	4.3	4.5		± 4.3			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43