


**LOT** PX 429  
Rev 2


**CONTROL**

 (Exp.) 2021-07-05  
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		XL80	XLR					XL80	XLR					XL80	XLR					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.7	2.7				± 0.4	7.5	7.5				± 1.0	17.7	17.7				± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.11	2.11				± 0.16	4.59	4.59				± 0.20	5.14	5.14				± 0.25	
HB HGB	g/dl	6.0	6.0				± 0.4	13.9	13.9				± 0.5	16.8	16.8				± 0.6	
	g/l	60	60				± 4	139	139				± 5	168	168				± 6	
	mmol/l	3.73	3.73				± 0.25	8.63	8.63				± 0.31	10.43	10.43				± 0.37	
HT HCT	%	17.5	17.5				± 1.5	39.0	39.0				± 2.0	45.7	45.7				± 2.5	
	l/l	0.175	0.175				± 0.015	0.390	0.390				± 0.020	0.457	0.457				± 0.025	
VGM MCV	µm <sup>3</sup> fl	83	83				± 5	85	85				± 5	89	89				± 5	
TGMH MCH	pg	28.4	28.4				± 2.0	30.3	30.3				± 2.0	32.7	32.7				± 2.5	
	fmol	1.77	1.77				± 0.12	1.88	1.88				± 0.12	2.03	2.03				± 0.16	
CCMH MCHC	g/dl	34.3	34.3				± 3.0	35.6	35.6				± 3.0	36.7	36.7				± 3.0	
	g/l	343	343				± 30	356	356				± 30	367	367				± 30	
	mmol/l	21.28	21.28				± 1.86	22.12	22.12				± 1.86	22.81	22.81				± 1.86	
IDR RDW	%	14.0	14.0				± 5.0	13.5	13.5				± 5.0	12.7	12.7				± 5.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	67				± 20	238	238				± 30	485	485				± 50	
VPM MPV	µm <sup>3</sup> fl	10.0	10.0				± 2.0	10.3	10.3				± 2.0	10.0	10.0				± 2.0	
NEUT	#	1.36	1.36				± 0.35	4.12	4.12				± 0.90	12.09	12.09				± 1.90	
	%	50.5	50.5				± 10.0	54.9	54.9				± 10.0	68.3	68.3				± 10.0	
LYMPHO	#	0.80	0.80				± 0.33	2.33	2.33				± 0.70	2.64	2.64				± 1.50	
	%	29.7	29.7				± 12.0	31.1	31.1				± 8.0	14.9	14.9				± 8.0	
MONO	#	0.15	0.15				± 0.15	0.52	0.52				± 0.52	1.29	1.29				± 1.29	
	%	5.6	5.6				± 5.6	6.9	6.9				± 6.9	7.3	7.3				± 7.3	
EOS	#	0.30	0.30				± 0.30	0.29	0.29				± 0.29	0.92	0.92				± 0.92	
	%	11.1	11.1				± 11.1	3.8	3.8				± 3.8	5.2	5.2				± 5.2	
BASO	#	0.08	0.08				± 0.08	0.25	0.25				± 0.25	0.76	0.76				± 0.76	
	%	3.1	3.1				± 3.1	3.3	3.3				± 3.3	4.3	4.3				± 4.3	

**LOT** PX 429  
Rev 2

**CONTROL**

 (Exp.) 2021-07-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60		MS60	MSCRP		60 60C+ ES60	MS60		MSCRP		60 60C+ ES60	MS60	MSCRP	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.7</b>	<b>2.6</b>	<b>2.6</b>		± 0.4	<b>7.5</b>	<b>7.4</b>	<b>7.3</b>		± 1.0	<b>18.3</b>	<b>17.8</b>	<b>17.6</b>		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.17</b>	<b>2.13</b>	<b>2.14</b>		± 0.16	<b>4.58</b>	<b>4.55</b>	<b>4.58</b>		± 0.20	<b>5.15</b>	<b>5.08</b>	<b>5.16</b>		± 0.25			
	g/dl	<b>6.0</b>	<b>5.9</b>	<b>6.0</b>		± 0.4	<b>13.9</b>	<b>13.7</b>	<b>13.8</b>		± 0.5	<b>16.8</b>	<b>16.5</b>	<b>16.7</b>		± 0.6			
HB HGB	g/l	<b>60</b>	<b>59</b>	<b>60</b>		± 4	<b>139</b>	<b>137</b>	<b>138</b>		± 5	<b>168</b>	<b>165</b>	<b>167</b>		± 6			
	mmol/l	<b>3.73</b>	<b>3.66</b>	<b>3.73</b>		± 0.25	<b>8.63</b>	<b>8.51</b>	<b>8.57</b>		± 0.31	<b>10.43</b>	<b>10.25</b>	<b>10.37</b>		± 0.37			
HT HCT	%	<b>17.1</b>	<b>16.6</b>	<b>16.9</b>		± 1.5	<b>38.0</b>	<b>37.3</b>	<b>38.0</b>		± 2.0	<b>45.8</b>	<b>44.7</b>	<b>44.9</b>		± 2.5			
	l/l	<b>0.171</b>	<b>0.166</b>	<b>0.169</b>		± 0.015	<b>0.380</b>	<b>0.373</b>	<b>0.380</b>		± 0.020	<b>0.458</b>	<b>0.447</b>	<b>0.449</b>		± 0.025			
VGM MCV	µm <sup>3</sup> /fl	<b>79</b>	<b>78</b>	<b>79</b>		± 5	<b>83</b>	<b>82</b>	<b>83</b>		± 5	<b>89</b>	<b>88</b>	<b>87</b>		± 5			
TGMH MCH	pg	<b>27.6</b>	<b>27.7</b>	<b>28.0</b>		± 2.0	<b>30.3</b>	<b>30.1</b>	<b>30.1</b>		± 2.0	<b>32.6</b>	<b>32.5</b>	<b>32.4</b>		± 2.5			
	fmol	<b>1.72</b>	<b>1.72</b>	<b>1.74</b>		± 0.12	<b>1.88</b>	<b>1.87</b>	<b>1.87</b>		± 0.12	<b>2.03</b>	<b>2.02</b>	<b>2.01</b>		± 0.16			
	g/dl	<b>35.0</b>	<b>35.5</b>	<b>35.5</b>		± 3.0	<b>36.6</b>	<b>36.7</b>	<b>36.3</b>		± 3.0	<b>36.7</b>	<b>36.9</b>	<b>37.2</b>		± 3.0			
CCMH MCHC	g/l	<b>350</b>	<b>355</b>	<b>355</b>		± 30	<b>366</b>	<b>367</b>	<b>363</b>		± 30	<b>367</b>	<b>369</b>	<b>372</b>		± 30			
	mmol/l	<b>21.73</b>	<b>22.05</b>	<b>22.04</b>		± 1.86	<b>22.71</b>	<b>22.80</b>	<b>22.54</b>		± 1.86	<b>22.76</b>	<b>22.92</b>	<b>23.10</b>		± 1.86			
IDR RDW	%	<b>13.5</b>	<b>12.5</b>	<b>13.5</b>		± 4.0	<b>12.5</b>	<b>12.5</b>	<b>12.5</b>		± 4.0	<b>12.0</b>	<b>11.5</b>	<b>11.5</b>		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>72</b>	<b>69</b>	<b>70</b>		± 20	<b>247</b>	<b>239</b>	<b>240</b>		± 30	<b>494</b>	<b>489</b>	<b>482</b>		± 50			
VPM MPV	µm <sup>3</sup> /fl	<b>9.4</b>	<b>9.6</b>	<b>9.0</b>		± 2.0	<b>9.5</b>	<b>9.8</b>	<b>9.1</b>		± 2.0	<b>9.1</b>	<b>9.3</b>	<b>8.7</b>		± 2.0			
NEUT	#	<b>1.35</b>	<b>1.29</b>	<b>1.43</b>		± 0.35	<b>4.04</b>	<b>4.00</b>	<b>4.12</b>		± 0.90	<b>12.70</b>	<b>12.28</b>	<b>12.58</b>		± 1.90			
	%	<b>50.1</b>	<b>49.8</b>	<b>55.0</b>		± 10.0	<b>53.8</b>	<b>54.0</b>	<b>56.5</b>		± 10.0	<b>69.4</b>	<b>69.0</b>	<b>71.5</b>		± 10.0			
LYMPHO	#	<b>0.82</b>	<b>0.82</b>	<b>0.74</b>		± 0.33	<b>2.41</b>	<b>2.52</b>	<b>2.34</b>		± 0.70	<b>2.75</b>	<b>2.83</b>	<b>2.55</b>		± 1.50			
	%	<b>30.4</b>	<b>31.4</b>	<b>28.5</b>		± 12.0	<b>32.1</b>	<b>34.0</b>	<b>32.0</b>		± 8.0	<b>15.0</b>	<b>15.9</b>	<b>14.5</b>		± 8.0			
MONO	#	<b>0.18</b>	<b>0.14</b>	<b>0.13</b>		± 0.13	<b>0.56</b>	<b>0.38</b>	<b>0.37</b>		± 0.37	<b>1.15</b>	<b>1.07</b>	<b>0.88</b>		± 0.88			
	%	<b>6.6</b>	<b>5.2</b>	<b>5.0</b>		± 5.0	<b>7.5</b>	<b>5.1</b>	<b>5.0</b>		± 5.0	<b>6.3</b>	<b>6.0</b>	<b>5.0</b>		± 5.0			
EOS	#	<b>0.27</b>	<b>0.27</b>	<b>0.21</b>		± 0.21	<b>0.26</b>	<b>0.27</b>	<b>0.22</b>		± 0.22	<b>0.93</b>	<b>0.89</b>	<b>0.79</b>		± 0.79			
	%	<b>9.9</b>	<b>10.5</b>	<b>8.0</b>		± 8.0	<b>3.4</b>	<b>3.7</b>	<b>3.0</b>		± 3.0	<b>5.1</b>	<b>5.0</b>	<b>4.5</b>		± 4.5			
BASO	#	<b>0.08</b>	<b>0.08</b>	<b>0.09</b>		± 0.08	<b>0.24</b>	<b>0.24</b>	<b>0.26</b>		± 0.24	<b>0.77</b>	<b>0.73</b>	<b>0.79</b>		± 0.73			
	%	<b>3.0</b>	<b>3.1</b>	<b>3.5</b>		± 3.0	<b>3.2</b>	<b>3.2</b>	<b>3.5</b>		± 3.2	<b>4.2</b>	<b>4.1</b>	<b>4.5</b>		± 4.1			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44