


LOT PX 431
Rev 2


CONTROL

 (Exp.) 2021-11-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6			± 0.4	7.9	7.9			± 1.0	17.8	17.8			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.25	2.25			± 0.16	4.45	4.45			± 0.20	5.09	5.09			± 0.25			
HB HGB	g/dl	6.5	6.5			± 0.4	13.5	13.5			± 0.5	16.4	16.4			± 0.6			
	g/l	65	65			± 4	135	135			± 5	164	164			± 6			
	mmol/l	4.04	4.04			± 0.25	8.38	8.38			± 0.31	10.18	10.18			± 0.37			
HT HCT	%	18.9	18.9			± 1.5	37.8	37.8			± 2.0	45.3	45.3			± 2.5			
	l/l	0.189	0.189			± 0.015	0.378	0.378			± 0.020	0.453	0.453			± 0.025			
VGM MCV	µm ³ fl	84	84			± 5	85	85			± 5	89	89			± 5			
TGMH MCH	pg	28.9	28.9			± 2.0	30.3	30.3			± 2.0	32.2	32.2			± 2.5			
	fmol	1.79	1.79			± 0.12	1.88	1.88			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	34.4	34.4			± 3.0	35.7	35.7			± 3.0	36.2	36.2			± 3.0			
	g/l	344	344			± 30	357	357			± 30	362	362			± 30			
	mmol/l	21.36	21.36			± 1.86	22.16	22.16			± 1.86	22.48	22.48			± 1.86			
IDR RDW	%	13.0	13.0			± 5.0	13.0	13.0			± 5.0	13.0	13.0			± 5.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	246	246			± 30	495	495			± 50			
VPM MPV	µm ³ fl	10.0	10.0			± 2.0	10.0	10.0			± 2.0	9.7	9.7			± 2.0			
NEUT	#	1.42	1.42			± 0.35	4.65	4.65			± 0.90	12.23	12.23			± 1.90			
	%	54.6	54.6			± 10.0	58.8	58.8			± 10.0	68.7	68.7			± 10.0			
LYMPHO	#	0.75	0.75			± 0.33	2.16	2.16			± 0.70	2.56	2.56			± 1.50			
	%	28.8	28.8			± 12.0	27.4	27.4			± 8.0	14.4	14.4			± 8.0			
MONO	#	0.16	0.16			± 0.16	0.47	0.47			± 0.47	1.12	1.12			± 1.12			
	%	6.2	6.2			± 6.2	6.0	6.0			± 6.0	6.3	6.3			± 6.3			
EOS	#	0.19	0.19			± 0.19	0.36	0.36			± 0.36	1.14	1.14			± 1.14			
	%	7.3	7.3			± 7.3	4.5	4.5			± 4.5	6.4	6.4			± 6.4			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.75	0.75			± 0.75			
	%	3.1	3.1			± 3.1	3.3	3.3			± 3.3	4.2	4.2			± 4.2			

LOT PX 431
Rev 2

CONTROL

 (Exp.) 2021-11-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.4	2.4		± 0.4	7.8	7.8	7.6		± 1.0	18.1	17.8	17.8		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.31	2.29	2.26		± 0.16	4.48	4.46	4.44		± 0.20	5.08	5.06	5.10		± 0.25			
HB HGB	g/dl	6.5	6.5	6.4		± 0.4	13.5	13.4	13.3		± 0.5	16.3	16.1	16.1		± 0.6			
	g/l	65	65	64		± 4	135	134	133		± 5	163	161	161		± 6			
HT HCT	mmol/l	4.04	4.04	3.97		± 0.25	8.38	8.32	8.26		± 0.31	10.12	10.00	10.00		± 0.37			
	%	18.5	18.3	18.1		± 1.5	37.2	37.0	36.9		± 2.0	44.7	44.0	44.4		± 2.5			
VGM MCV	l/l	0.185	0.183	0.181		± 0.015	0.372	0.370	0.369		± 0.020	0.447	0.440	0.444		± 0.025			
	µm ³ /fl	80	80	80.0		± 5	83	83	83.0		± 5	88	87	87.0		± 5			
TGMH MCH	pg	28.1	28.4	28.3		± 2.0	30.1	30.0	30.0		± 2.0	32.1	31.8	31.6		± 2.5			
	fmol	1.75	1.76	1.76		± 0.12	1.87	1.87	1.86		± 0.12	1.99	1.98	1.96		± 0.16			
CCMH MCHC	g/dl	35.2	35.5	35.4		± 3.0	36.3	36.2	36.1		± 3.0	36.5	36.6	36.3		± 3.0			
	g/l	352	355	354		± 30	363	362	361		± 30	365	366	363		± 30			
IDR RDW	mmol/l	21.84	22.03	21.98		± 1.86	22.55	22.48	22.41		± 1.86	22.64	22.71	22.53		± 1.86			
	%	13.0	13.0	13.0		± 4.0	12.5	12.5	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	69	64		± 20	246	249	240		± 30	485	498	475		± 50			
VPM MPV	µm ³ /fl	9.8	9.9	9.6		± 2.0	9.3	9.5	9.0		± 2.0	9.0	8.8	8.6		± 2.0			
NEUT	#	1.36	1.35	1.37		± 0.35	4.49	4.73	4.60		± 0.90	12.27	12.50	12.82		± 1.90			
	%	54.2	56.3	57.0		± 10.0	57.5	60.6	60.5		± 10.0	67.8	70.2	72.0		± 10.0			
LYMPHO	#	0.75	0.72	0.72		± 0.33	2.25	2.18	2.17		± 0.70	2.79	2.60	2.49		± 1.50			
	%	30.1	29.9	30.0		± 12.0	28.8	27.9	28.5		± 8.0	15.4	14.6	14.0		± 8.0			
MONO	#	0.14	0.08	0.08		± 0.08	0.48	0.30	0.30		± 0.30	1.14	0.85	0.71		± 0.71			
	%	5.7	3.5	3.5		± 3.5	6.2	3.8	4.0		± 3.8	6.3	4.8	4.0		± 4.0			
EOS	#	0.17	0.18	0.14		± 0.14	0.33	0.35	0.27		± 0.27	1.14	1.14	0.98		± 0.98			
	%	6.9	7.5	6.0		± 6.0	4.2	4.5	3.5		± 3.5	6.3	6.4	5.5		± 5.5			
BASO	#	0.08	0.07	0.08		± 0.07	0.26	0.25	0.27		± 0.25	0.76	0.71	0.80		± 0.71			
	%	3.1	2.8	3.5		± 2.8	3.3	3.2	3.5		± 3.2	4.2	4.0	4.5		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.45