


LOT PX 075
Rev 1


CONTROL

 (Exp.) 2015-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.6	2.6	2.6	2.6	± 0.4	7.5	7.6	7.7	7.6	7.5	± 1.0	18.0	18.1	18.1	18.1	18.1	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38	2.34	2.35	2.34	2.34	± 0.16	4.63	4.64	4.62	4.64	4.59	± 0.20	5.15	5.16	5.15	5.19	5.10	± 0.25	
HB	HGB	g/dl	6.6	6.6	6.6	6.6	6.5	± 0.4	13.6	13.6	13.6	13.6	13.7	± 0.5	16.4	16.4	16.3	16.4	16.3	± 0.6	
		g/l	66	66	66	66	65	± 4	136	136	136	136	137	± 5	164	164	163	164	163	± 6	
		mmol/l	4.10	4.10	4.10	4.10	4.04	± 0.25	8.45	8.45	8.45	8.45	8.51	± 0.31	10.18	10.18	10.12	10.18	10.12	± 0.37	
HT	HCT	%	18.8	19.2	18.3	19.0	18.1	± 1.5	38.0	38.0	37.0	37.6	36.9	± 2.0	44.8	44.9	44.3	44.6	43.9	± 2.5	
		l/l	0.188	0.192	0.183	0.190	0.181	± 0.015	0.380	0.380	0.370	0.376	0.369	± 0.020	0.448	0.449	0.443	0.446	0.439	± 0.025	
VGM	MCV	µm ³ ; fl	79	82	78	81	77.5	± 5	82	82	80	81	80.5	± 5	87	87	86	86	86.0	± 5	
TGMH	MCH	pg	27.7	28.2	28.1	28.2	27.8	± 2.0	29.4	29.3	29.4	29.3	29.8	± 2.0	31.8	31.8	31.7	31.6	32.0	± 2.5	
		fmol	1.72	1.75	1.74	1.75	1.73	± 0.12	1.82	1.82	1.83	1.82	1.85	± 0.12	1.98	1.97	1.97	1.96	1.98	± 0.16	
CCMH	MCHC	g/dl	35.1	34.4	36.0	34.8	35.8	± 3.0	35.8	35.7	36.8	36.2	37.1	± 3.0	36.6	36.5	36.8	36.7	37.2	± 3.0	
		g/l	351	344	360	348	358	± 30	358	357	368	362	371	± 30	366	365	368	367	372	± 30	
		mmol/l	21.80	21.36	22.36	21.62	22.26	± 1.86	22.25	22.20	22.85	22.47	23.03	± 1.86	22.73	22.69	22.85	22.82	23.08	± 1.86	
IDR	RDW	%	12.4	13.0	12.0	13.0	10.3	± 4.0	12.4	13.3	12.2	13.3	10.5	± 4.0	12.0	13.3	11.6	13.3	10.5	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	72	72	70	± 20	255	250	255	250	245	± 30	485	480	495	480	465	± 50	
VPM	MPV	µm ³ ; fl	9.6	9.9	9.7	9.9	9.1	± 2.0	9.3	9.7	9.5	9.7	8.9	± 2.0	9.3	9.6	9.4	9.6	8.7	± 2.0	
NEUT	#	#	1.40	1.48	1.48	1.48	1.60	± 0.35	4.22	4.29	4.39	4.29	4.31	± 0.90	12.44	12.62	12.76	12.62	12.94	± 1.90	
		%	56.1	57.0	57.0	57.0	61.5	± 10.0	56.2	56.5	57.0	56.5	57.5	± 10.0	69.1	69.7	70.5	69.7	71.5	± 10.0	
LYMPHO	#	#	0.74	0.77	0.77	0.77	0.70	± 0.33	2.44	2.47	2.50	2.47	2.36	± 0.70	2.88	2.90	2.81	2.90	2.62	± 1.50	
		%	29.7	29.5	29.5	29.5	27.0	± 12.0	32.5	32.5	32.5	32.5	31.5	± 8.0	16.0	16.0	15.5	16.0	14.5	± 8.0	
MONO	#	#	0.13	0.10	0.10	0.10	0.09	± 0.09	0.30	0.27	0.23	0.27	0.23	± 0.23	0.81	0.69	0.63	0.69	0.63	± 0.63	
		%	5.2	4.0	4.0	4.0	3.5	± 3.5	4.0	3.5	3.0	3.5	3.0	± 3.0	4.5	3.8	3.5	3.8	3.5	± 3.5	
EOS	#	#	0.14	0.16	0.16	0.16	0.13	± 0.13	0.29	0.30	0.31	0.30	0.34	± 0.29	1.08	1.09	1.09	1.09	1.09	± 1.08	
		%	5.7	6.0	6.0	6.0	5.0	± 5.0	3.8	4.0	4.0	4.0	4.5	± 3.8	6.0	6.0	6.0	6.0	6.0	± 6.0	
BASO	#	#	0.08	0.09	0.09	0.09	0.08	± 0.08	0.26	0.27	0.27	0.27	0.26	± 0.26	0.79	0.81	0.81	0.81	0.81	± 0.79	
		%	3.3	3.5	3.5	3.5	3.0	± 3.0	3.5	3.5	3.5	3.5	3.5	± 3.5	4.4	4.5	4.5	4.5	4.5	± 4.4	

LOT PX 075
Rev 1

CONTROL

 (Exp.) 2015-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.8	7.8	7.8		± 1.0	18.8	18.8	18.8		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.39	2.39		± 0.16	4.64	4.64	4.64		± 0.20	5.22	5.22	5.22		± 0.25		
	g/dl	6.7	6.7	6.7		± 0.4	13.6	13.6	13.6		± 0.5	16.2	16.2	16.2		± 0.6		
HB HGB	g/l	67	67	67		± 4	136	136	136		± 5	162	162	162		± 6		
	mmol/l	4.16	4.16	4.16		± 0.25	8.45	8.45	8.45		± 0.31	10.06	10.06	10.06		± 0.37		
HT HCT	%	19.6	19.6	19.6		± 1.5	38.5	38.5	38.5		± 2.0	45.9	45.9	45.9		± 2.5		
	l/l	0.196	0.196	0.196		± 0.015	0.385	0.385	0.385		± 0.020	0.459	0.459	0.459		± 0.025		
VGM MCV	µm ³ ; fl	82	82	82		± 5	83	83	83		± 5	88	88	88		± 5		
TGMH MCH	pg	28.0	28.0	28.0		± 2.0	29.3	29.3	29.3		± 2.0	31.0	31.0	31.0		± 2.5		
	fmol	1.74	1.74	1.74		± 0.12	1.82	1.82	1.82		± 0.12	1.93	1.93	1.93		± 0.16		
	g/dl	34.2	34.2	34.2		± 3.0	35.3	35.3	35.3		± 3.0	35.3	35.3	35.3		± 3.0		
CCMH MCHC	g/l	342	342	342		± 30	353	353	353		± 30	353	353	353		± 30		
	mmol/l	21.23	21.23	21.23		± 1.86	21.93	21.93	21.93		± 1.86	21.90	21.90	21.90		± 1.86		
IDR RDW	%	15.0	15.0	15.0		± 4.0	15.8	15.8	15.8		± 4.0	14.0	14.0	14.0		± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78	78	78		± 20	260	260	260		± 30	495	495	495		± 50		
VPM MPV	µm ³ ; fl	10.0	10.0	10.0		± 2.0	9.4	9.4	9.4		± 2.0	9.3	9.3	9.3		± 2.0		
NEUT	#	1.48	1.54	1.54		± 0.35	4.55	4.68	4.68		± 0.90	13.20	13.50	13.50		± 1.90		
	%	56.8	59.1	59.1		± 10.0	58.3	60.0	60.0		± 10.0	69.9	71.4	71.4		± 10.0		
LYMPHO	#	0.73	0.66	0.66		± 0.33	2.29	2.13	2.13		± 0.70	2.74	2.46	2.46		± 1.50		
	%	28.0	25.2	25.2		± 12.0	29.4	27.3	27.3		± 8.0	14.6	13.1	13.1		± 8.0		
MONO	#	0.18	0.18	0.18		± 0.18	0.47	0.48	0.48		± 0.47	1.45	1.43	1.43		± 1.43		
	%	6.9	7.1	7.1		± 6.9	6.0	6.2	6.2		± 6.0	7.7	7.6	7.6		± 7.6		
EOS	#	0.15	0.16	0.16		± 0.15	0.26	0.27	0.27		± 0.26	1.00	1.02	1.02		± 1.00		
	%	5.8	6.1	6.1		± 5.8	3.3	3.5	3.5		± 3.3	5.3	5.4	5.4		± 5.3		
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.47	0.47	0.47		± 0.47		
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5		

9930096-A
FRONT / RECTO
Ref: TEMP-0821 Rev.36