

LOT PX 014

CONTROL

(Exp.) **2014-03-05**
 (YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE	
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3	2.3	2.3	± 0.4		7.2	7.2	7.2	7.2		± 1.0	17.6	17.2	17.3			17.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.30	2.34	2.30	± 0.12	4.38	4.35	4.37	4.35	± 0.15	5.20	5.20	5.17	5.20	± 0.20			
HB HGB	g/dl	6.5	6.5	6.5	6.5	± 0.4	13.1	13.1	13.1	13.1	± 0.5	16.8	16.8	16.7	16.8	± 0.6			
	g/l	65	65	65	65	± 4	131	131	131	131	± 5	168	168	167	168	± 6			
	mmol/l	4.04	4.04	4.04	4.04	± 0.25	8.14	8.14	8.14	8.14	± 0.31	10.43	10.43	10.37	10.43	± 0.37			
HT HCT	%	18.6	18.9	18.3	18.9	± 1.5	35.9	35.7	35.0	35.7	± 2.0	45.8	45.8	44.5	45.8	± 2.5			
	l/l	0.186	0.189	0.183	0.189	± 0.015	0.359	0.357	0.350	0.357	± 0.020	0.458	0.458	0.445	0.458	± 0.025			
VGM MCV	µm ³ ; fl	79	82	78	82	± 5	82	82	80	82	± 5	88	88	86	88	± 5			
TGMH MCH	pg	27.7	28.3	27.8	28.3	± 2.0	29.9	30.1	30.0	30.1	± 2.0	32.3	32.3	32.3	32.3	± 2.5			
	fmol	1.72	1.76	1.73	1.76	± 0.12	1.86	1.87	1.86	1.87	± 0.12	2.01	2.01	2.01	2.01	± 0.16			
CCMH MCHC	g/dl	35.0	34.5	35.6	34.5	± 3.0	36.5	36.7	37.5	36.7	± 3.0	36.7	36.7	37.6	36.7	± 3.0			
	g/l	350	345	356	345	± 30	365	367	375	367	± 30	367	367	376	367	± 30			
	mmol/l	21.74	21.40	22.12	21.40	± 1.86	22.65	22.81	23.27	22.81	± 1.86	22.80	22.80	23.32	22.80	± 1.86			
IDR RDW	%	12.7	13.3	12.5	13.3	± 4.0	12.0	13.0	12.0	13.0	± 4.0	11.6	13.0	11.5	13.0	± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	72	72	72	± 20	258	260	265	260	± 30	510	520	535	520	± 50			
VPM MPV	µm ³ ; fl	9.2	9.4	9.2	9.4	± 2.0	8.6	8.8	8.6	8.8	± 2.0	8.9	9.1	9.1	9.1	± 2.0			
NEUT	#	1.27	1.27	1.29	1.27	± 0.25	4.14	4.18	4.18	4.18	± 0.80	12.32	12.04	12.20	12.04	± 1.90			
	%	55.0	55.3	56.0	55.3	± 10.0	57.5	58.0	58.0	58.0	± 10.0	70.0	70.0	70.5	70.0	± 10.0			
LYMPHO	#	0.70	0.70	0.69	0.70	± 0.20	2.23	2.20	2.23	2.20	± 0.60	2.90	2.84	2.77	2.84	± 1.50			
	%	30.5	30.5	30.0	30.5	± 8.0	31.0	30.5	31.0	30.5	± 8.0	16.5	16.5	16.0	16.5	± 8.0			
MONO	#	0.12	0.10	0.10	0.10	± 0.10	0.32	0.29	0.29	0.29	± 0.29	0.79	0.69	0.69	0.69	± 0.69			
	%	5.0	4.5	4.5	4.5	± 4.5	4.5	4.0	4.0	4.0	± 4.0	4.5	4.0	4.0	4.0	± 4.0			
EOS	#	0.15	0.15	0.15	0.15	± 0.12	0.25	0.29	0.25	0.29	± 0.21	0.88	0.86	0.95	0.86	± 0.73			
	%	6.5	6.5	6.5	6.5	± 5.0	3.5	4.0	3.5	4.0	± 2.8	5.0	5.0	5.5	5.0	± 4.0			
BASO	#	0.07	0.07	0.07	0.07	± 0.07	0.25	0.25	0.25	0.25	± 0.19	0.70	0.77	0.69	0.77	± 0.46			
	%	3.0	3.2	3.0	3.2	± 3.0	3.5	3.5	3.5	3.5	± 2.5	4.0	4.5	4.0	4.5	± 2.5			

LOT PX 014

CONTROL

(Exp.) 2014-03-05
 (YYYY-MM-DD)

Rev 1

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.4	7.4	7.4		± 1.0	18.3	18.3	18.3		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35	2.35		± 0.12	4.38	4.38	4.38		± 0.15	5.26	5.26	5.26		± 0.20			
HB HGB	g/dl	6.6	6.6	6.6		± 0.4	13.1	13.1	13.1		± 0.5	16.6	16.6	16.6		± 0.6			
	g/l	66	66	66		± 4	131	131	131		± 5	166	166	166		± 6			
	mmol/l	4.10	4.10	4.10		± 0.25	8.14	8.14	8.14		± 0.31	10.31	10.31	10.31		± 0.37			
HT HCT	%	19.0	19.0	19.0		± 1.5	36.4	36.4	36.4		± 2.0	46.8	46.8	46.8		± 2.5			
	l/l	0.190	0.190	0.190		± 0.015	0.364	0.364	0.364		± 0.020	0.468	0.468	0.468		± 0.025			
VGM MCV	µm ³ ; fl	81	81	81		± 5	83	83	83		± 5	89	89	89		± 5			
TGMH MCH	pg	28.1	28.1	28.1		± 2.0	29.9	29.9	29.9		± 2.0	31.6	31.6	31.6		± 2.5			
	fmol	1.74	1.74	1.74		± 0.12	1.86	1.86	1.86		± 0.12	1.96	1.96	1.96		± 0.16			
CCMH MCHC	g/dl	34.7	34.7	34.7		± 3.0	36.0	36.0	36.0		± 3.0	35.5	35.5	35.5		± 3.0			
	g/l	347	347	347		± 30	360	360	360		± 30	355	355	355		± 30			
	mmol/l	21.53	21.53	21.53		± 1.86	22.38	22.38	22.38		± 1.86	22.02	22.02	22.02		± 1.86			
IDR RDW	%	16.1	16.1	16.1		± 4.0	16.8	16.8	16.8		± 4.0	16.1	16.1	16.1		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	268	268	268		± 30	525	525	525		± 50			
VPM MPV	µm ³ ; fl	9.1	9.1	9.1		± 2.0	8.5	8.5	8.5		± 2.0	8.8	8.8	8.8		± 2.0			
NEUT	#	1.25	1.38	1.38		± 0.25	4.26	4.43	4.43		± 0.80	12.90	13.20	13.20		± 1.90			
	%	52.0	57.6	57.6		± 10.0	57.5	59.8	59.8		± 10.0	70.5	72.2	72.2		± 10.0			
LYMPHO	#	0.78	0.65	0.65		± 0.20	2.31	2.09	2.09		± 0.60	3.02	2.69	2.69		± 1.50			
	%	32.6	26.9	26.9		± 8.0	31.2	28.3	28.3		± 8.0	16.5	14.7	14.7		± 8.0			
MONO	#	0.16	0.13	0.13		± 0.10	0.44	0.39	0.39		± 0.29	1.19	1.08	1.08		± 0.69			
	%	6.8	5.4	5.4		± 4.5	6.0	5.3	5.3		± 4.0	6.5	5.9	5.9		± 4.0			
EOS	#	0.13	0.17	0.17		± 0.12	0.21	0.27	0.27		± 0.21	0.73	0.86	0.86		± 0.73			
	%	5.6	7.1	7.1		± 5.0	2.8	3.6	3.6		± 2.8	4.0	4.7	4.7		± 4.0			
BASO	#	0.07	0.07	0.07		± 0.07	0.19	0.22	0.22		± 0.19	0.46	0.46	0.46		± 0.46			
	%	3.0	3.0	3.0		± 3.0	2.5	3.0	3.0		± 2.5	2.5	2.5	2.5		± 2.5			

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B