

LOT PX 015
 Rev 1

CONTROL

(Exp.) 2015-03-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | | | | |
|--------------------------|------|--|--------------------|------------|-------|-------|--------|---------|--------------------|------------|-------|-------|--------|---------|--------------------|------------|-------|-------|--------|---------|-------------------------|
| | | | CONTROL | | | | | L | CONTROL | | | | | N | CONTROL | | | | | H | TOLERANCES TOLERANCE |
| | | | PENTRA | | | | | MSCRIP | PENTRA | | | | | MSCRIP | PENTRA | | | | | MSCRIP | |
| | | | 60 60C+ ES60 | 80 XL80 | MS60 | XL80 | MSCRIP | | 60 60C+ ES60 | 80 XL80 | MS60 | XL80 | MSCRIP | | 60 60C+ ES60 | 80 XL80 | MS60 | XL80 | MSCRIP | | |
| GB | WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.4 | 2.4 | 2.4 | 2.4 | 2.4 | ± 0.4 | 7.5 | 7.5 | 7.5 | 7.5 | 7.4 | ± 1.0 | 18.0 | 17.6 | 17.7 | 17.6 | 17.9 | ± 2.2 | |
| GR | RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.44 | 2.39 | 2.45 | 2.39 | 2.39 | ± 0.16 | 4.56 | 4.57 | 4.58 | 4.57 | 4.51 | ± 0.20 | 5.10 | 5.12 | 5.12 | 5.12 | 5.03 | ± 0.25 | |
| HB | HGB | g/dl | 7.0 | 7.0 | 7.0 | 7.0 | 6.8 | ± 0.4 | 13.5 | 13.5 | 13.5 | 13.5 | 13.3 | ± 0.5 | 16.1 | 16.2 | 16.1 | 16.2 | 15.9 | ± 0.6 | |
| | | g/l | 70 | 70 | 70 | 70 | 68 | ± 4 | 135 | 135 | 135 | 135 | 133 | ± 5 | 161 | 162 | 161 | 162 | 159 | ± 6 | |
| | | mmol/l | 4.35 | 4.35 | 4.35 | 4.35 | 4.22 | ± 0.25 | 8.38 | 8.38 | 8.38 | 8.38 | 8.26 | ± 0.31 | 10.00 | 10.06 | 10.00 | 10.06 | 9.87 | ± 0.37 | |
| HT | HCT | % | 19.8 | 19.8 | 19.6 | 19.8 | 19.2 | ± 1.5 | 36.9 | 37.0 | 36.2 | 37.0 | 36.7 | ± 2.0 | 44.4 | 44.5 | 43.5 | 44.5 | 43.9 | ± 2.5 | |
| | | l/l | 0.198 | 0.198 | 0.196 | 0.198 | 0.192 | ± 0.015 | 0.369 | 0.370 | 0.362 | 0.370 | 0.367 | ± 0.020 | 0.444 | 0.445 | 0.435 | 0.445 | 0.439 | ± 0.025 | |
| VGM | MCV | µm ³ ; fl | 81 | 83 | 80 | 83 | 80.3 | ± 5 | 81 | 81 | 79 | 81 | 81.3 | ± 5 | 87 | 87 | 85 | 87 | 87.3 | ± 5 | |
| TGMH | MCH | pg | 28.7 | 29.3 | 28.6 | 29.3 | 28.5 | ± 2.0 | 29.6 | 29.5 | 29.5 | 29.5 | 29.5 | ± 2.0 | 31.6 | 31.6 | 31.4 | 31.6 | 31.6 | ± 2.5 | |
| | | fmol | 1.78 | 1.82 | 1.77 | 1.82 | 1.77 | ± 0.12 | 1.84 | 1.83 | 1.83 | 1.83 | 1.83 | ± 0.12 | 1.96 | 1.96 | 1.95 | 1.96 | 1.96 | ± 0.16 | |
| CCMH | MCHC | g/dl | 35.4 | 35.3 | 35.7 | 35.3 | 35.4 | ± 3.0 | 36.5 | 36.5 | 37.3 | 36.5 | 36.3 | ± 3.0 | 36.3 | 36.4 | 37.0 | 36.4 | 36.2 | ± 3.0 | |
| | | g/l | 354 | 353 | 357 | 353 | 354 | ± 30 | 365 | 365 | 373 | 365 | 363 | ± 30 | 363 | 364 | 370 | 364 | 362 | ± 30 | |
| | | mmol/l | 21.99 | 21.91 | 22.18 | 21.91 | 22.00 | ± 1.86 | 22.70 | 22.65 | 23.17 | 22.65 | 22.53 | ± 1.86 | 22.53 | 22.58 | 22.97 | 22.58 | 22.49 | ± 1.86 | |
| IDR | RDW | % | 12.3 | 13.3 | 11.8 | 13.3 | 10.5 | ± 4.0 | 12.0 | 13.5 | 12.0 | 13.5 | 11.2 | ± 4.0 | 11.5 | 13.5 | 11.3 | 13.5 | 11.1 | ± 4.0 | |
| PLAQ. | PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 76 | 74 | 74 | 74 | 73 | ± 20 | 248 | 245 | 248 | 245 | 252 | ± 30 | 470 | 470 | 485 | 470 | 483 | ± 50 | |
| VPM | MPV | µm ³ ; fl | 9.8 | 10.2 | 10.0 | 10.2 | 9.1 | ± 2.0 | 9.7 | 10.0 | 9.8 | 10.0 | 8.9 | ± 2.0 | 8.4 | 8.7 | 8.5 | 8.7 | 7.7 | ± 2.0 | |
| NEUT | # | % | 1.39 | 1.45 | 1.43 | 1.45 | 1.46 | ± 0.35 | 4.22 | 4.43 | 4.35 | 4.43 | 4.40 | ± 0.90 | 12.69 | 12.58 | 12.66 | 12.58 | 13.19 | ± 1.90 | |
| | | % | 58.0 | 60.5 | 59.5 | 60.5 | 61.3 | ± 10.0 | 56.3 | 59.0 | 58.0 | 59.0 | 59.7 | ± 10.0 | 70.5 | 71.5 | 71.5 | 71.5 | 73.8 | ± 10.0 | |
| LYMPHO | # | % | 0.65 | 0.60 | 0.65 | 0.60 | 0.58 | ± 0.31 | 2.36 | 2.25 | 2.29 | 2.25 | 2.12 | ± 0.70 | 2.70 | 2.64 | 2.57 | 2.64 | 2.23 | ± 1.50 | |
| | | % | 27.0 | 25.0 | 27.0 | 25.0 | 24.1 | ± 12.0 | 31.5 | 30.0 | 30.5 | 30.0 | 28.8 | ± 8.0 | 15.0 | 15.0 | 14.5 | 15.0 | 12.5 | ± 8.0 | |
| MONO | # | % | 0.11 | 0.08 | 0.07 | 0.08 | 0.10 | ± 0.07 | 0.32 | 0.23 | 0.26 | 0.23 | 0.26 | ± 0.23 | 0.81 | 0.53 | 0.62 | 0.53 | 0.63 | ± 0.53 | |
| | | % | 4.5 | 3.5 | 3.0 | 3.5 | 4.0 | ± 3.0 | 4.2 | 3.0 | 3.5 | 3.0 | 3.5 | ± 3.0 | 4.5 | 3.0 | 3.5 | 3.0 | 3.5 | ± 3.0 | |
| EOS | # | % | 0.17 | 0.18 | 0.17 | 0.18 | 0.17 | ± 0.17 | 0.34 | 0.34 | 0.34 | 0.34 | 0.34 | ± 0.34 | 0.99 | 1.06 | 1.06 | 1.06 | 1.02 | ± 0.99 | |
| | | % | 7.0 | 7.5 | 7.0 | 7.5 | 7.0 | ± 7.0 | 4.5 | 4.5 | 4.5 | 4.5 | 4.5 | ± 4.5 | 5.5 | 6.0 | 6.0 | 6.0 | 5.7 | ± 5.5 | |
| BASO | # | % | 0.08 | 0.08 | 0.08 | 0.08 | 0.09 | ± 0.08 | 0.26 | 0.26 | 0.26 | 0.26 | 0.26 | ± 0.26 | 0.81 | 0.79 | 0.80 | 0.79 | 0.80 | ± 0.79 | |
| | | % | 3.5 | 3.5 | 3.5 | 3.5 | 3.6 | ± 3.5 | 3.5 | 3.5 | 3.5 | 3.5 | 3.5 | ± 3.5 | 4.5 | 4.5 | 4.5 | 4.5 | 4.5 | ± 4.5 | |

LOT PX 015
 Rev 1

CONTROL

(Exp.) 2015-03-05
 (YYYY-MM-DD)

| | | ABX Lysebio | | | | | | | | | | | | | | | | | | |
|--------------------------|--|------------------|----------------|----------------------|--|--|---------|-------------------------|------------------|----------------|----------------------|--|---------|---------|-------------------------|------------------|----------------|----------------------|---------|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | | L | CONTROL | | | | | N | CONTROL | | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | | PENTRA | | | | | | PENTRA | | | | | | |
| | | 120 120 RETIC | DX120 DF120 | DX NEXUS DF NEXUS | | | | TOLERANCES TOLERANCE | 120 120 RETIC | DX120 DF120 | DX NEXUS DF NEXUS | | | | TOLERANCES TOLERANCE | 120 120 RETIC | DX120 DF120 | DX NEXUS DF NEXUS | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.5 | 2.5 | 2.5 | | | ± 0.4 | 8.0 | 8.0 | 8.0 | | | ± 1.0 | 19.8 | 19.8 | 19.8 | | | ± 2.2 | |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.45 | 2.45 | 2.45 | | | ± 0.16 | 4.58 | 4.58 | 4.58 | | | ± 0.20 | 5.15 | 5.15 | 5.15 | | | ± 0.25 | |
| HB HGB | g/dl | 7.0 | 7.0 | 7.0 | | | ± 0.4 | 13.4 | 13.4 | 13.4 | | | ± 0.5 | 16.1 | 16.1 | 16.1 | | | ± 0.6 | |
| | g/l | 70 | 70 | 70 | | | ± 4 | 134 | 134 | 134 | | | ± 5 | 161 | 161 | 161 | | | ± 6 | |
| | mmol/l | 4.35 | 4.35 | 4.35 | | | ± 0.25 | 8.32 | 8.32 | 8.32 | | | ± 0.31 | 10.00 | 10.00 | 10.00 | | | ± 0.37 | |
| HT HCT | % | 20.3 | 20.3 | 20.3 | | | ± 1.5 | 37.6 | 37.6 | 37.6 | | | ± 2.0 | 45.3 | 45.3 | 45.3 | | | ± 2.5 | |
| | l/l | 0.203 | 0.203 | 0.203 | | | ± 0.015 | 0.376 | 0.376 | 0.376 | | | ± 0.020 | 0.453 | 0.453 | 0.453 | | | ± 0.025 | |
| VGM MCV | µm ³ ·fl | 83 | 83 | 83 | | | ± 5 | 82 | 82 | 82 | | | ± 5 | 88 | 88 | 88 | | | ± 5 | |
| TGMH MCH | pg | 28.6 | 28.6 | 28.6 | | | ± 2.0 | 29.3 | 29.3 | 29.3 | | | ± 2.0 | 31.3 | 31.3 | 31.3 | | | ± 2.5 | |
| | fmol | 1.77 | 1.77 | 1.77 | | | ± 0.12 | 1.82 | 1.82 | 1.82 | | | ± 0.12 | 1.94 | 1.94 | 1.94 | | | ± 0.16 | |
| CCMH MCHC | g/dl | 34.4 | 34.4 | 34.4 | | | ± 3.0 | 35.7 | 35.7 | 35.7 | | | ± 3.0 | 35.5 | 35.5 | 35.5 | | | ± 3.0 | |
| | g/l | 344 | 344 | 344 | | | ± 30 | 357 | 357 | 357 | | | ± 30 | 355 | 355 | 355 | | | ± 30 | |
| | mmol/l | 21.38 | 21.38 | 21.38 | | | ± 1.86 | 22.16 | 22.16 | 22.16 | | | ± 1.86 | 22.06 | 22.06 | 22.06 | | | ± 1.86 | |
| IDR RDW | % | 15.0 | 15.0 | 15.0 | | | ± 4.0 | 16.0 | 16.0 | 16.0 | | | ± 4.0 | 14.8 | 14.8 | 14.8 | | | ± 4.0 | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 78 | 78 | 78 | | | ± 20 | 255 | 255 | 255 | | | ± 30 | 465 | 465 | 465 | | | ± 50 | |
| VPM MPV | µm ³ ·fl | 9.9 | 9.9 | 9.9 | | | ± 2.0 | 9.7 | 9.7 | 9.7 | | | ± 2.0 | 8.4 | 8.4 | 8.4 | | | ± 2.0 | |
| NEUT | # | 1.53 | 1.58 | 1.58 | | | ± 0.35 | 4.74 | 4.88 | 4.88 | | | ± 0.90 | 14.50 | 14.70 | 14.70 | | | ± 1.90 | |
| | % | 61.2 | 63.2 | 63.2 | | | ± 10.0 | 59.2 | 61.0 | 61.0 | | | ± 10.0 | 72.9 | 74.2 | 74.2 | | | ± 10.0 | |
| LYMPHO | # | 0.63 | 0.57 | 0.57 | | | ± 0.31 | 2.27 | 2.10 | 2.10 | | | ± 0.70 | 2.59 | 2.34 | 2.34 | | | ± 1.50 | |
| | % | 25.1 | 22.6 | 22.6 | | | ± 12.0 | 28.4 | 26.2 | 26.2 | | | ± 8.0 | 13.1 | 11.8 | 11.8 | | | ± 8.0 | |
| MONO | # | 0.11 | 0.11 | 0.11 | | | ± 0.11 | 0.43 | 0.44 | 0.44 | | | ± 0.43 | 1.19 | 1.17 | 1.17 | | | ± 1.17 | |
| | % | 4.4 | 4.5 | 4.5 | | | ± 4.4 | 5.4 | 5.5 | 5.5 | | | ± 5.4 | 6.0 | 5.9 | 5.9 | | | ± 5.9 | |
| EOS | # | 0.17 | 0.18 | 0.18 | | | ± 0.17 | 0.32 | 0.34 | 0.34 | | | ± 0.32 | 1.09 | 1.11 | 1.11 | | | ± 1.09 | |
| | % | 6.8 | 7.2 | 7.2 | | | ± 6.8 | 4.0 | 4.3 | 4.3 | | | ± 4.0 | 5.5 | 5.6 | 5.6 | | | ± 5.5 | |
| BASO | # | 0.06 | 0.06 | 0.06 | | | ± 0.06 | 0.24 | 0.24 | 0.24 | | | ± 0.24 | 0.50 | 0.50 | 0.50 | | | ± 0.50 | |
| | % | 2.5 | 2.5 | 2.5 | | | ± 2.5 | 3.0 | 3.0 | 3.0 | | | ± 3.0 | 2.5 | 2.5 | 2.5 | | | ± 2.5 | |

9930091-A
 Ref. TEMP-0821 Rev.35