


LOT PX 035
Rev 1


CONTROL

 (Exp.) 2015-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.5	± 0.4	7.6	7.6	7.6	7.6	7.6	± 1.0	18.1	17.7	18.0	17.7	17.8	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.39	2.33	2.37	2.33	2.33	± 0.16	4.66	4.66	4.69	4.66	4.60	± 0.20	5.10	5.10	5.08	5.10	5.04	± 0.25	
HB	HGB	g/dl	6.8	6.8	6.8	6.8	6.7	± 0.4	13.7	13.8	13.7	13.8	13.6	± 0.5	16.2	16.3	16.2	16.3	16.1	± 0.6	
		g/l	68	68	68	68	67	± 4	137	138	137	138	136	± 5	162	163	162	163	161	± 6	
		mmol/l	4.22	4.22	4.22	4.22	4.16	± 0.25	8.51	8.57	8.51	8.57	8.45	± 0.31	10.06	10.12	10.06	10.12	10.00	± 0.37	
HT	HCT	%	19.1	19.3	19.0	19.3	18.8	± 1.5	37.7	38.2	37.5	38.2	37.2	± 2.0	44.9	44.9	44.2	44.9	44.4	± 2.5	
		l/l	0.191	0.193	0.190	0.193	0.188	± 0.015	0.377	0.382	0.375	0.382	0.372	± 0.020	0.449	0.449	0.442	0.449	0.444	± 0.025	
VGM	MCV	µm ³ ; fl	80	83	80	83	80.7	± 5	81	82	80	82	80.8	± 5	88	88	87	88	88.0	± 5	
TGMH	MCH	pg	28.5	29.2	28.7	29.2	28.8	± 2.0	29.4	29.6	29.2	29.6	29.6	± 2.0	31.8	32.0	31.9	32.0	31.9	± 2.5	
		fmol	1.77	1.81	1.78	1.81	1.79	± 0.12	1.83	1.84	1.81	1.84	1.84	± 0.12	1.97	1.98	1.98	1.98	1.98	± 0.16	
CCMH	MCHC	g/dl	35.6	35.2	35.9	35.2	35.6	± 3.0	36.3	36.1	36.5	36.1	36.6	± 3.0	36.1	36.3	36.7	36.3	36.3	± 3.0	
		g/l	356	352	359	352	356	± 30	363	361	365	361	366	± 30	361	363	367	363	363	± 30	
		mmol/l	22.09	21.84	22.27	21.84	22.13	± 1.86	22.54	22.43	22.68	22.43	22.72	± 1.86	22.42	22.55	22.76	22.55	22.54	± 1.86	
IDR	RDW	%	12.6	14.3	12.3	14.3	11.5	± 4.0	12.3	13.5	12.2	13.5	10.8	± 4.0	11.8	13.3	11.7	13.3	10.6	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	70	68	70	74	± 20	255	255	255	255	255	± 30	495	505	505	505	500	± 50	
VPM	MPV	µm ³ ; fl	8.3	8.4	8.4	8.4	7.9	± 2.0	9.1	9.4	9.2	9.4	8.6	± 2.0	8.2	8.4	8.2	8.4	7.7	± 2.0	
NEUT	#	#	1.46	1.49	1.50	1.49	1.56	± 0.35	4.37	4.45	4.48	4.45	4.60	± 0.90	12.94	12.76	13.14	12.76	13.23	± 1.90	
		%	58.5	59.5	60.0	59.5	62.2	± 10.0	57.5	58.5	59.0	58.5	60.5	± 10.0	71.5	72.1	73.0	72.1	74.3	± 10.0	
LYMPHO	#	#	0.71	0.70	0.69	0.70	0.66	± 0.31	2.39	2.37	2.33	2.37	2.24	± 0.70	2.62	2.66	2.61	2.66	2.49	± 1.50	
		%	28.5	28.0	27.5	28.0	26.3	± 12.0	31.5	31.2	30.7	31.2	29.5	± 8.0	14.5	15.0	14.5	15.0	14.0	± 8.0	
MONO	#	#	0.09	0.08	0.08	0.08	0.08	± 0.08	0.30	0.23	0.23	0.23	0.23	± 0.23	0.81	0.62	0.53	0.62	0.53	± 0.53	
		%	3.5	3.0	3.0	3.0	3.0	± 3.0	4.0	3.0	3.0	3.0	3.0	± 3.0	4.5	3.5	3.0	3.5	3.0	± 3.0	
EOS	#	#	0.15	0.15	0.15	0.15	0.13	± 0.13	0.27	0.29	0.29	0.29	0.27	± 0.27	0.91	0.89	0.90	0.89	0.80	± 0.80	
		%	6.0	6.0	6.0	6.0	5.0	± 5.0	3.5	3.8	3.8	3.8	3.5	± 3.5	5.0	5.0	5.0	5.0	4.5	± 4.5	
BASO	#	#	0.09	0.09	0.09	0.09	0.09	± 0.09	0.27	0.27	0.27	0.27	0.27	± 0.27	0.81	0.78	0.81	0.78	0.75	± 0.75	
		%	3.5	3.5	3.5	3.5	3.5	± 3.5	3.5	3.5	3.5	3.5	3.5	± 3.5	4.5	4.4	4.5	4.4	4.2	± 4.2	

LOT PX 035
Rev 1

CONTROL

 (Exp.) 2015-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8		± 1.0	18.6	18.6	18.6		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	2.40		± 0.16	4.67	4.67	4.67		± 0.20	5.13	5.13	5.13		± 0.25		
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.7	13.7	13.7		± 0.5	16.1	16.1	16.1		± 0.6		
	g/l	69	69	69		± 4	137	137	137		± 5	161	161	161		± 6		
	mmol/l	4.28	4.28	4.28		± 0.25	8.51	8.51	8.51		± 0.31	10.00	10.00	10.00		± 0.37		
HT HCT	%	19.9	19.9	19.9		± 1.5	38.3	38.3	38.3		± 2.0	45.7	45.7	45.7		± 2.5		
	l/l	0.199	0.199	0.199		± 0.015	0.383	0.383	0.383		± 0.020	0.457	0.457	0.457		± 0.025		
VGM MCV	µm ³ ; fl	83	83	83		± 5	82	82	82		± 5	89	89	89		± 5		
TGMH MCH	pg	28.8	28.8	28.8		± 2.0	29.3	29.3	29.3		± 2.0	31.4	31.4	31.4		± 2.5		
	fmol	1.79	1.79	1.79		± 0.12	1.82	1.82	1.82		± 0.12	1.95	1.95	1.95		± 0.16		
CCMH MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.8	35.8	35.8		± 3.0	35.3	35.3	35.3		± 3.0		
	g/l	346	346	346		± 30	358	358	358		± 30	353	353	353		± 30		
	mmol/l	21.51	21.51	21.51		± 1.86	22.22	22.22	22.22		± 1.86	21.90	21.90	21.90		± 1.86		
IDR RDW	%	15.0	15.0	15.0		± 4.0	15.5	15.5	15.5		± 4.0	13.8	13.8	13.8		± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74	74		± 20	260	260	260		± 30	500	500	500		± 50		
VPM MPV	µm ³ ; fl	8.3	8.3	8.3		± 2.0	9.0	9.0	9.0		± 2.0	8.1	8.1	8.1		± 2.0		
NEUT	#	1.47	1.53	1.53		± 0.35	4.52	4.66	4.66		± 0.90	13.40	13.60	13.60		± 1.90		
	%	58.7	61.2	61.2		± 10.0	57.9	59.8	59.8		± 10.0	72.0	73.3	73.3		± 10.0		
LYMPHO	#	0.72	0.65	0.65		± 0.31	2.39	2.21	2.21		± 0.70	2.72	2.46	2.46		± 1.50		
	%	28.7	25.8	25.8		± 12.0	30.6	28.3	28.3		± 8.0	14.6	13.2	13.2		± 8.0		
MONO	#	0.12	0.12	0.12		± 0.12	0.41	0.42	0.42		± 0.41	1.15	1.15	1.15		± 1.15		
	%	4.7	4.8	4.8		± 4.7	5.2	5.4	5.4		± 5.2	6.2	6.2	6.2		± 6.2		
EOS	#	0.14	0.14	0.14		± 0.14	0.26	0.27	0.27		± 0.26	0.87	0.89	0.89		± 0.87		
	%	5.4	5.7	5.7		± 5.4	3.3	3.5	3.5		± 3.3	4.7	4.8	4.8		± 4.7		
BASO	#	0.06	0.06	0.06		± 0.06	0.23	0.23	0.23		± 0.23	0.47	0.47	0.47		± 0.47		
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5		

9930091-A
Ref: TEMP-0821 Rev.35