


**LOT** PX 095  
Rev 1


**CONTROL**

 (Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.5	2.5	2.5	2.6	± 0.4	7.4	7.4	7.6	7.4	7.7	± 1.0	18.3	18.1	18.3	18.1	18.6	± 2.2		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.38	2.40	2.38	2.41	± 0.16	4.54	4.54	4.52	4.54	4.54	± 0.20	5.28	5.30	5.30	5.30	5.26	± 0.25		
HB	HGB	g/dl	6.9	7.0	6.9	7.0	7.0	± 0.4	13.4	13.4	13.4	13.4	13.5	± 0.5	17.1	17.1	17.1	17.1	17.1	± 0.6		
		g/l	69	70	69	70	70	± 4	134	134	134	134	135	± 5	171	171	171	171	171	± 6		
		mmol/l	4.28	4.35	4.28	4.35	4.35	± 0.25	8.32	8.32	8.32	8.32	8.38	± 0.31	10.62	10.62	10.62	10.62	10.62	± 0.37		
HT	HCT	%	19.8	20.0	19.2	20.0	19.3	± 1.5	37.2	37.2	36.6	37.2	36.8	± 2.0	47.0	47.2	46.1	46.6	46.0	± 2.5		
		l/l	0.198	0.200	0.192	0.200	0.193	± 0.015	0.372	0.372	0.366	0.372	0.368	± 0.020	0.470	0.472	0.461	0.466	0.460	± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	81	84	80	84	80.0	± 5	82	82	81	82	81.0	± 5	89	89	87	88	87.5	± 5		
TGMH	MCH	pg	28.3	29.4	28.8	29.4	29.0	± 2.0	29.5	29.5	29.6	29.5	29.7	± 2.0	32.4	32.3	32.3	32.3	32.5	± 2.5		
		fmol	1.76	1.83	1.79	1.83	1.80	± 0.12	1.83	1.83	1.84	1.83	1.85	± 0.12	2.01	2.00	2.00	2.00	2.02	± 0.16		
CCMH	MCHC	g/dl	34.9	35.0	35.9	35.0	36.3	± 3.0	36.0	36.0	36.6	36.0	36.7	± 3.0	36.4	36.3	37.1	36.7	37.2	± 3.0		
		g/l	349	350	359	350	363	± 30	360	360	366	360	367	± 30	364	363	371	367	372	± 30		
		mmol/l	21.68	21.74	22.32	21.74	22.55	± 1.86	22.35	22.35	22.73	22.35	22.80	± 1.86	22.60	22.51	23.03	22.77	23.07	± 1.86		
IDR	RDW	%	13.0	13.0	12.8	13.0	10.8	± 4.0	12.8	13.2	12.5	13.2	11.0	± 4.0	12.8	13.2	11.8	13.2	11.2	± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	70	73	70	74	± 20	246	245	247	245	240	± 30	485	485	505	485	475	± 50		
VPM	MPV	µm <sup>3</sup> ; fl	9.5	9.7	9.5	9.7	9.0	± 2.0	9.5	9.7	9.7	9.7	9.1	± 2.0	9.2	9.5	9.4	9.5	8.8	± 2.0		
NEUT	#		1.32	1.38	1.38	1.38	1.45	± 0.35	4.14	4.19	4.33	4.19	4.45	± 0.90	12.99	12.92	13.23	12.92	13.45	± 1.90		
		%	55.0	55.0	55.3	55.0	55.6	± 10.0	56.0	56.6	57.0	56.6	57.8	± 10.0	71.0	71.4	72.3	71.4	72.3	± 10.0		
LYMPHO	#		0.79	0.83	0.83	0.83	0.83	± 0.33	2.48	2.44	2.51	2.44	2.45	± 0.70	2.93	2.90	2.80	2.90	2.79	± 1.50		
		%	33.0	33.0	33.2	33.0	31.8	± 12.0	33.5	33.0	33.0	33.0	31.8	± 8.0	16.0	16.0	15.3	16.0	15.0	± 8.0		
MONO	#		0.08	0.08	0.08	0.08	0.08	± 0.08	0.22	0.22	0.23	0.22	0.23	± 0.22	0.64	0.54	0.55	0.54	0.56	± 0.54		
		%	3.5	3.3	3.0	3.3	3.0	± 3.0	3.0	3.0	3.0	3.0	3.0	± 3.0	3.5	3.0	3.0	3.0	3.0	± 3.0		
EOS	#		0.13	0.14	0.14	0.14	0.16	± 0.13	0.30	0.30	0.30	0.30	0.31	± 0.30	0.92	0.94	0.92	0.94	1.02	± 0.92		
		%	5.3	5.5	5.5	5.5	6.2	± 5.3	4.0	4.0	4.0	4.0	4.0	± 4.0	5.0	5.2	5.0	5.2	5.5	± 5.0		
BASO	#		0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.25	0.23	0.25	0.26	± 0.23	0.82	0.80	0.81	0.80	0.78	± 0.78		
		%	3.2	3.2	3.0	3.2	3.4	± 3.0	3.5	3.4	3.0	3.4	3.4	± 3.0	4.5	4.4	4.4	4.4	4.2	± 4.2		

**LOT** PX 095  
Rev 1

**CONTROL**

 (Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.6</b>	<b>2.6</b>	<b>2.6</b>		± 0.4	<b>7.8</b>	<b>7.8</b>	<b>7.8</b>		± 1.0	<b>19.2</b>	<b>19.2</b>	<b>19.2</b>		± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.44</b>	<b>2.44</b>	<b>2.44</b>		± 0.16	<b>4.58</b>	<b>4.58</b>	<b>4.58</b>		± 0.20	<b>5.39</b>	<b>5.39</b>	<b>5.39</b>		± 0.25		
	g/dl	<b>7.0</b>	<b>7.0</b>	<b>7.0</b>		± 0.4	<b>13.4</b>	<b>13.4</b>	<b>13.4</b>		± 0.5	<b>16.9</b>	<b>16.9</b>	<b>16.9</b>		± 0.6		
HB HGB	g/l	<b>70</b>	<b>70</b>	<b>70</b>		± 4	<b>134</b>	<b>134</b>	<b>134</b>		± 5	<b>169</b>	<b>169</b>	<b>169</b>		± 6		
	mmol/l	<b>4.35</b>	<b>4.35</b>	<b>4.35</b>		± 0.25	<b>8.32</b>	<b>8.32</b>	<b>8.32</b>		± 0.31	<b>10.49</b>	<b>10.49</b>	<b>10.49</b>		± 0.37		
HT HCT	%	<b>20.3</b>	<b>20.3</b>	<b>20.3</b>		± 1.5	<b>38.0</b>	<b>38.0</b>	<b>38.0</b>		± 2.0	<b>48.0</b>	<b>48.0</b>	<b>48.0</b>		± 2.5		
	l/l	<b>0.203</b>	<b>0.203</b>	<b>0.203</b>		± 0.015	<b>0.380</b>	<b>0.380</b>	<b>0.380</b>		± 0.020	<b>0.480</b>	<b>0.480</b>	<b>0.480</b>		± 0.025		
VGM MCV	µm <sup>3</sup> ; fl	<b>83</b>	<b>83</b>	<b>83</b>		± 5	<b>83</b>	<b>83</b>	<b>83</b>		± 5	<b>89</b>	<b>89</b>	<b>89</b>		± 5		
TGMH MCH	pg	<b>28.7</b>	<b>28.7</b>	<b>28.7</b>		± 2.0	<b>29.3</b>	<b>29.3</b>	<b>29.3</b>		± 2.0	<b>31.4</b>	<b>31.4</b>	<b>31.4</b>		± 2.5		
	fmol	<b>1.78</b>	<b>1.78</b>	<b>1.78</b>		± 0.12	<b>1.82</b>	<b>1.82</b>	<b>1.82</b>		± 0.12	<b>1.95</b>	<b>1.95</b>	<b>1.95</b>		± 0.16		
CCMH MCHC	g/dl	<b>34.6</b>	<b>34.6</b>	<b>34.6</b>		± 3.0	<b>35.3</b>	<b>35.3</b>	<b>35.3</b>		± 3.0	<b>35.2</b>	<b>35.2</b>	<b>35.2</b>		± 3.0		
	g/l	<b>346</b>	<b>346</b>	<b>346</b>		± 30	<b>353</b>	<b>353</b>	<b>353</b>		± 30	<b>352</b>	<b>352</b>	<b>352</b>		± 30		
	mmol/l	<b>21.46</b>	<b>21.46</b>	<b>21.46</b>		± 1.86	<b>21.89</b>	<b>21.89</b>	<b>21.89</b>		± 1.86	<b>21.88</b>	<b>21.88</b>	<b>21.88</b>		± 1.86		
IDR RDW	%	<b>15.0</b>	<b>15.0</b>	<b>15.0</b>		± 4.0	<b>15.6</b>	<b>15.6</b>	<b>15.6</b>		± 4.0	<b>14.8</b>	<b>14.8</b>	<b>14.8</b>		± 4.0		
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>77</b>	<b>77</b>	<b>77</b>		± 20	<b>257</b>	<b>257</b>	<b>257</b>		± 30	<b>500</b>	<b>500</b>	<b>500</b>		± 50		
VPM MPV	µm <sup>3</sup> ; fl	<b>9.7</b>	<b>9.7</b>	<b>9.7</b>		± 2.0	<b>9.5</b>	<b>9.5</b>	<b>9.5</b>		± 2.0	<b>9.3</b>	<b>9.3</b>	<b>9.3</b>		± 2.0		
NEUT	#	<b>1.45</b>	<b>1.52</b>	<b>1.52</b>		± 0.35	<b>4.52</b>	<b>4.67</b>	<b>4.67</b>		± 0.90	<b>14.20</b>	<b>14.50</b>	<b>14.50</b>		± 1.90		
	%	<b>55.6</b>	<b>58.3</b>	<b>58.3</b>		± 10.0	<b>58.0</b>	<b>59.9</b>	<b>59.9</b>		± 10.0	<b>73.8</b>	<b>75.3</b>	<b>75.3</b>		± 10.0		
LYMPHO	#	<b>0.82</b>	<b>0.74</b>	<b>0.74</b>		± 0.33	<b>2.46</b>	<b>2.29</b>	<b>2.29</b>		± 0.70	<b>2.92</b>	<b>2.63</b>	<b>2.63</b>		± 1.50		
	%	<b>31.6</b>	<b>28.4</b>	<b>28.4</b>		± 12.0	<b>31.6</b>	<b>29.3</b>	<b>29.3</b>		± 8.0	<b>15.2</b>	<b>13.7</b>	<b>13.7</b>		± 8.0		
MONO	#	<b>0.15</b>	<b>0.16</b>	<b>0.16</b>		± 0.15	<b>0.30</b>	<b>0.30</b>	<b>0.30</b>		± 0.30	<b>0.77</b>	<b>0.75</b>	<b>0.75</b>		± 0.75		
	%	<b>5.8</b>	<b>6.0</b>	<b>6.0</b>		± 5.8	<b>3.8</b>	<b>3.9</b>	<b>3.9</b>		± 3.8	<b>4.0</b>	<b>3.9</b>	<b>3.9</b>		± 3.9		
EOS	#	<b>0.12</b>	<b>0.12</b>	<b>0.12</b>		± 0.12	<b>0.28</b>	<b>0.30</b>	<b>0.30</b>		± 0.28	<b>0.86</b>	<b>0.88</b>	<b>0.88</b>		± 0.86		
	%	<b>4.5</b>	<b>4.8</b>	<b>4.8</b>		± 4.5	<b>3.6</b>	<b>3.9</b>	<b>3.9</b>		± 3.6	<b>4.5</b>	<b>4.6</b>	<b>4.6</b>		± 4.5		
BASO	#	<b>0.07</b>	<b>0.07</b>	<b>0.07</b>		± 0.07	<b>0.23</b>	<b>0.23</b>	<b>0.23</b>		± 0.23	<b>0.48</b>	<b>0.48</b>	<b>0.48</b>		± 0.48		
	%	<b>2.5</b>	<b>2.5</b>	<b>2.5</b>		± 2.5	<b>3.0</b>	<b>3.0</b>	<b>3.0</b>		± 3.0	<b>2.5</b>	<b>2.5</b>	<b>2.5</b>		± 2.5		

9930096-A  
FRONT / RECTO  
Ref: TEMP-0821 Rev.36