


LOT PX 115
Rev 1


CONTROL

 (Exp.) 2016-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.4	2.5	2.4	± 0.4	7.8	7.8	7.8	7.8	7.8	± 1.0	17.8	17.3	17.5	17.3	17.6	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.42	2.37	2.42	2.37	2.40	± 0.16	4.52	4.51	4.53	4.51	4.47	± 0.20	5.08	5.11	5.12	5.11	5.06	± 0.25	
HB	HGB	g/dl	6.9	6.9	6.9	6.9	6.9	± 0.4	13.6	13.6	13.6	13.6	13.5	± 0.5	16.2	16.3	16.3	16.3	16.2	± 0.6	
		g/l	69	69	69	69	69	± 4	136	136	136	136	135	± 5	162	163	163	163	162	± 6	
		mmol/l	4.28	4.28	4.28	4.28	4.28	± 0.25	8.45	8.45	8.45	8.45	8.38	± 0.31	10.06	10.12	10.12	10.12	10.06	± 0.37	
HT	HCT	%	19.4	19.4	19.4	19.4	19.2	± 1.5	37.1	37.0	36.7	37.0	36.7	± 2.0	44.7	44.5	44.0	44.5	44.0	± 2.5	
		l/l	0.194	0.194	0.194	0.194	0.192	± 0.015	0.371	0.370	0.367	0.370	0.367	± 0.020	0.447	0.445	0.440	0.445	0.440	± 0.025	
VGM	MCV	µm ³ ; fl	80	82	80	82	80.0	± 5	82	82	81	82	82.0	± 5	88	87	86	87	87.0	± 5	
TGMH	MCH	pg	28.5	29.1	28.5	29.1	28.8	± 2.0	30.1	30.2	30.0	30.2	30.2	± 2.0	31.9	31.9	31.8	31.9	32.0	± 2.5	
		fmol	1.77	1.81	1.77	1.81	1.79	± 0.12	1.87	1.87	1.86	1.87	1.88	± 0.12	1.98	1.98	1.98	1.98	1.99	± 0.16	
CCMH	MCHC	g/dl	35.6	35.5	35.6	35.5	35.9	± 3.0	36.7	36.8	37.1	36.8	36.8	± 3.0	36.2	36.7	37.0	36.7	36.8	± 3.0	
		g/l	356	355	356	355	359	± 30	367	368	371	368	368	± 30	362	367	370	367	368	± 30	
		mmol/l	22.13	22.05	22.13	22.05	22.32	± 1.86	22.79	22.84	23.02	22.84	22.87	± 1.86	22.50	22.77	22.99	22.77	22.85	± 1.86	
IDR	RDW	%	13.4	13.0	12.0	13.0	11.8	± 4.0	12.9	12.5	12.2	12.5	12.0	± 4.0	12.4	12.0	11.5	12.0	11.5	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	78	75	75	± 20	257	260	265	260	260	± 30	520	520	540	520	525	± 50	
VPM	MPV	µm ³ ; fl	9.4	9.5	9.3	9.5	9.0	± 2.0	9.2	9.4	9.0	9.4	8.7	± 2.0	8.6	8.9	8.8	8.9	8.2	± 2.0	
NEUT	#	#	1.30	1.36	1.32	1.36	1.36	± 0.35	4.33	4.31	4.33	4.31	4.37	± 0.90	12.46	12.02	12.34	12.02	12.72	± 1.90	
		%	54.1	54.3	54.8	54.3	56.6	± 10.0	55.5	55.2	55.5	55.2	56.0	± 10.0	70.0	69.5	70.5	69.5	72.3	± 10.0	
LYMPHO	#	#	0.75	0.80	0.76	0.80	0.74	± 0.33	2.54	2.51	2.56	2.51	2.51	± 0.70	2.78	2.77	2.73	2.77	2.64	± 1.50	
		%	31.3	32.0	31.8	32.0	31.0	± 12.0	32.5	32.2	32.8	32.2	32.2	± 8.0	15.6	16.0	15.6	16.0	15.0	± 8.0	
MONO	#	#	0.12	0.09	0.08	0.09	0.08	± 0.08	0.35	0.30	0.31	0.30	0.27	± 0.27	0.89	0.78	0.79	0.78	0.62	± 0.62	
		%	4.8	3.5	3.2	3.5	3.2	± 3.2	4.5	3.8	4.0	3.8	3.5	± 3.5	5.0	4.5	4.5	4.5	3.5	± 3.5	
EOS	#	#	0.16	0.17	0.16	0.17	0.14	± 0.14	0.31	0.39	0.33	0.39	0.37	± 0.31	0.89	0.95	0.88	0.95	0.88	± 0.88	
		%	6.5	6.8	6.8	6.8	5.8	± 5.8	4.0	5.0	4.2	5.0	4.8	± 4.0	5.0	5.5	5.0	5.5	5.0	± 5.0	
BASO	#	#	0.08	0.09	0.08	0.09	0.08	± 0.08	0.27	0.30	0.27	0.30	0.27	± 0.27	0.78	0.78	0.77	0.78	0.74	± 0.74	
		%	3.3	3.4	3.4	3.4	3.4	± 3.3	3.5	3.8	3.5	3.8	3.5	± 3.5	4.4	4.5	4.4	4.5	4.2	± 4.2	

LOT PX 115
Rev 1

CONTROL

 (Exp.) 2016-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	± 0.4	8.0	8.0	8.0	± 1.0	18.5	18.5	18.5	± 2.2				
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.43	2.43	± 0.16	4.53	4.53	4.53	± 0.20	5.16	5.16	5.16	± 0.25				
		g/dl	7.0	7.0	7.0	± 0.4	13.6	13.6	13.6	± 0.5	16.1	16.1	16.1	± 0.6				
HB	HGB	g/l	70	70	70	± 4	136	136	136	± 5	161	161	161	± 6				
		mmol/l	4.35	4.35	4.35	± 0.25	8.45	8.45	8.45	± 0.31	10.00	10.00	10.00	± 0.37				
HT	HCT	%	20.2	20.2	20.2	± 1.5	37.6	37.6	37.6	± 2.0	45.4	45.4	45.4	± 2.5				
		l/l	0.202	0.202	0.202	± 0.015	0.376	0.376	0.376	± 0.020	0.454	0.454	0.454	± 0.025				
VGM	MCV	µm ³ ; fl	83	83	83	± 5	83	83	83	± 5	88	88	88	± 5				
TGMH	MCH	pg	28.8	28.8	28.8	± 2.0	30.0	30.0	30.0	± 2.0	31.2	31.2	31.2	± 2.5				
		fmol	1.79	1.79	1.79	± 0.12	1.86	1.86	1.86	± 0.12	1.94	1.94	1.94	± 0.16				
		g/dl	34.7	34.7	34.7	± 3.0	36.2	36.2	36.2	± 3.0	35.5	35.5	35.5	± 3.0				
CCMH	MCHC	g/l	347	347	347	± 30	362	362	362	± 30	355	355	355	± 30				
		mmol/l	21.55	21.55	21.55	± 1.86	22.46	22.46	22.46	± 1.86	22.02	22.02	22.02	± 1.86				
IDR	RDW	%	14.7	14.7	14.7	± 4.0	15.7	15.7	15.7	± 4.0	14.4	14.4	14.4	± 4.0				
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	80	80	± 20	268	268	268	± 30	520	520	520	± 50				
VPM	MPV	µm ³ ; fl	9.3	9.3	9.3	± 2.0	9.0	9.0	9.0	± 2.0	8.5	8.5	8.5	± 2.0				
		#	1.35	1.42	1.42	± 0.35	4.43	4.58	4.58	± 0.90	13.00	13.20	13.20	± 1.90				
		%	54.1	56.7	56.7	± 10.0	55.4	57.2	57.2	± 10.0	70.0	71.4	71.4	± 10.0				
		#	0.78	0.70	0.70	± 0.33	2.50	2.31	2.31	± 0.70	2.72	2.46	2.46	± 1.50				
		%	31.0	27.9	27.9	± 12.0	31.2	28.9	28.9	± 8.0	14.7	13.3	13.3	± 8.0				
		#	0.15	0.15	0.15	± 0.15	0.59	0.62	0.62	± 0.59	1.57	1.55	1.55	± 1.55				
		%	6.0	6.1	6.1	± 6.0	7.4	7.7	7.7	± 7.4	8.5	8.4	8.4	± 8.4				
		#	0.16	0.17	0.17	± 0.16	0.24	0.26	0.26	± 0.24	0.80	0.81	0.81	± 0.80				
		%	6.4	6.8	6.8	± 6.4	3.0	3.2	3.2	± 3.0	4.3	4.4	4.4	± 4.3				
		#	0.06	0.06	0.06	± 0.06	0.24	0.24	0.24	± 0.24	0.46	0.46	0.46	± 0.46				
		%	2.5	2.5	2.5	± 2.5	3.0	3.0	3.0	± 3.0	2.5	2.5	2.5	± 2.5				

9930096-A FRONT / RECTO Ref: TEMP-0821 Rev.36