


**LOT** PX 056  
Rev 4


**CONTROL**

 (Exp.) 2016-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.5	2.4	2.5	2.4	± 0.4	7.6	7.5	7.4	7.5	7.4	± 1.0	17.9	17.5	17.3	17.5	17.6	± 2.2		
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.35	2.38	2.35	2.36	± 0.16	4.60	4.60	4.58	4.60	4.57	± 0.20	5.08	5.10	5.09	5.10	5.08	± 0.25		
HB	HGB	g/dl	6.8	6.8	6.8	6.8	6.7	± 0.4	13.5	13.6	13.5	13.6	13.5	± 0.5	15.7	15.8	15.8	15.8	15.7	± 0.6		
		g/l	68	68	68	68	67	± 4	135	136	135	136	135	± 5	157	158	158	158	157	± 6		
		mmol/l	4.22	4.22	4.22	4.22	4.16	± 0.25	8.38	8.45	8.38	8.45	8.38	± 0.31	9.75	9.81	9.81	9.81	9.75	± 0.37		
HT	HCT	%	19.3	19.5	19.0	19.3	18.6	± 1.5	37.7	37.7	37.1	37.3	37.5	± 2.0	43.2	43.9	42.8	43.4	43.7	± 2.5		
		l/l	0.193	0.195	0.190	0.193	0.186	± 0.015	0.377	0.377	0.371	0.373	0.375	± 0.020	0.432	0.439	0.428	0.434	0.437	± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	80	83	80	82	79.0	± 5	82	82	81	81	82.0	± 5	85	86	84	85	86.0	± 5		
TGMH	MCH	pg	28.2	28.9	28.6	28.9	28.4	± 2.0	29.3	29.6	29.5	29.6	29.5	± 2.0	30.9	31.0	31.0	31.0	30.9	± 2.5		
		fmol	1.75	1.80	1.77	1.80	1.76	± 0.12	1.82	1.84	1.83	1.84	1.83	± 0.12	1.92	1.92	1.93	1.92	1.92	± 0.16		
CCMH	MCHC	g/dl	35.3	34.9	35.7	35.3	35.9	± 3.0	35.8	36.1	36.4	36.5	36.0	± 3.0	36.4	36.0	37.0	36.4	35.9	± 3.0		
		g/l	353	349	357	353	359	± 30	358	361	364	365	360	± 30	364	360	370	364	359	± 30		
		mmol/l	21.90	21.65	22.18	21.91	22.32	± 1.86	22.23	22.39	22.60	22.67	22.37	± 1.86	22.58	22.37	22.95	22.63	22.32	± 1.86		
IDR	RDW	%	14.1	14.1	12.8	14.1	12.5	± 4.0	13.5	14.1	12.7	14.1	12.8	± 4.0	13.4	13.3	11.9	13.3	12.0	± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	70	70	70	68	± 20	250	250	255	250	245	± 30	500	495	515	495	495	± 50		
VPM	MPV	µm <sup>3</sup> ; fl	9.5	9.7	9.7	9.7	9.0	± 2.0	9.4	9.8	9.5	9.8	8.9	± 2.0	9.0	9.3	9.1	9.3	8.4	± 2.0		
NEUT	#		1.55	1.64	1.60	1.64	1.60	± 0.35	4.18	4.07	4.14	4.07	4.20	± 0.90	12.71	12.21	12.56	12.21	12.92	± 1.90		
		%	64.5	65.4	66.5	65.4	66.7	± 10.0	55.0	54.2	55.9	54.2	56.7	± 10.0	71.0	69.8	72.6	69.8	73.4	± 10.0		
LYMPHO	#		0.53	0.52	0.51	0.52	0.49	± 0.33	2.36	2.33	2.36	2.33	2.28	± 0.70	2.36	2.21	2.23	2.21	2.11	± 1.50		
		%	22.0	20.9	21.2	20.9	20.5	± 12.0	31.0	31.0	31.9	31.0	30.8	± 8.0	13.2	12.6	12.9	12.6	12.0	± 8.0		
MONO	#		0.11	0.11	0.08	0.11	0.11	± 0.08	0.44	0.46	0.31	0.46	0.37	± 0.31	0.90	0.93	0.64	0.93	0.70	± 0.64		
		%	4.5	4.6	3.3	4.6	4.5	± 3.3	5.8	6.3	4.2	6.3	5.0	± 4.2	5.0	5.3	3.7	5.3	4.0	± 3.7		
EOS	#		0.13	0.14	0.13	0.14	0.11	± 0.11	0.34	0.38	0.33	0.38	0.30	± 0.30	1.16	1.40	1.12	1.40	1.13	± 1.12		
		%	5.5	5.4	5.3	5.4	4.5	± 4.5	4.5	5.0	4.5	5.0	4.0	± 4.0	6.5	8.0	6.5	8.0	6.4	± 6.4		
BASO	#		0.08	0.09	0.09	0.09	0.09	± 0.08	0.28	0.26	0.26	0.26	0.26	± 0.26	0.77	0.75	0.74	0.75	0.74	± 0.74		
		%	3.5	3.7	3.7	3.7	3.8	± 3.5	3.7	3.5	3.5	3.5	3.5	± 3.5	4.3	4.3	4.3	4.3	4.2	± 4.2		

**LOT** PX 056  
Rev 4

**CONTROL**

 (Exp.) 2016-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA					PENTRA					PENTRA							
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS				120	DX120
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5			± 0.4	7.8	7.8	7.8			± 1.0	18.4	18.4	18.4			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.42	2.42	2.42			± 0.16	4.62	4.62	4.62			± 0.20	5.17	5.17	5.17			± 0.25
		g/dl	6.9	6.9	6.9			± 0.4	13.6	13.6	13.6			± 0.5	15.7	15.7	15.7			± 0.6
HB	HGB	g/l	69	69	69			± 4	136	136	136			± 5	157	157	157			± 6
		mmol/l	4.28	4.28	4.28			± 0.25	8.45	8.45	8.45			± 0.31	9.75	9.75	9.75			± 0.37
HT	HCT	%	20.1	20.1	20.1			± 1.5	38.3	38.3	38.3			± 2.0	44.5	44.5	44.5			± 2.5
		l/l	0.201	0.201	0.201			± 0.015	0.383	0.383	0.383			± 0.020	0.445	0.445	0.445			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	83	83	83			± 5	83	83	83			± 5	86	86	86			± 5
TGMH	MCH	pg	28.5	28.5	28.5			± 2.0	29.4	29.4	29.4			± 2.0	30.4	30.4	30.4			± 2.5
		fmol	1.77	1.77	1.77			± 0.12	1.83	1.83	1.83			± 0.12	1.89	1.89	1.89			± 0.16
		g/dl	34.4	34.4	34.4			± 3.0	35.5	35.5	35.5			± 3.0	35.3	35.3	35.3			± 3.0
CCMH	MCHC	g/l	344	344	344			± 30	355	355	355			± 30	353	353	353			± 30
		mmol/l	21.33	21.33	21.33			± 1.86	22.02	22.02	22.02			± 1.86	21.93	21.93	21.93			± 1.86
IDR	RDW	%	16.4	16.4	16.4			± 4.0	16.7	16.7	16.7			± 4.0	15.6	15.6	15.6			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	74	74	74			± 20	260	260	260			± 30	505	505	505			± 50
VPM	MPV	µm <sup>3</sup> ; fl	9.7	9.7	9.7			± 2.0	9.6	9.6	9.6			± 2.0	9.1	9.1	9.1			± 2.0
		#	1.66	1.70	1.70			± 0.35	4.37	4.48	4.48			± 0.90	13.60	13.80	13.80			± 1.90
		%	66.5	67.9	67.9			± 10.0	56.0	57.4	57.4			± 10.0	73.9	74.8	74.8			± 10.0
		#	0.51	0.47	0.47			± 0.33	2.33	2.18	2.18			± 0.70	2.24	2.08	2.08			± 1.50
		%	20.5	18.6	18.6			± 12.0	29.9	27.9	27.9			± 8.0	12.2	11.3	11.3			± 8.0
		#	0.13	0.14	0.14			± 0.13	0.52	0.54	0.54			± 0.52	0.99	0.99	0.99			± 0.99
		%	5.3	5.4	5.4			± 5.3	6.7	6.9	6.9			± 6.7	5.4	5.4	5.4			± 5.4
		#	0.13	0.14	0.14			± 0.13	0.34	0.37	0.37			± 0.34	1.10	1.10	1.10			± 1.10
		%	5.2	5.6	5.6			± 5.2	4.4	4.8	4.8			± 4.4	6.0	6.0	6.0			± 6.0
		#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.46	0.46	0.46			± 0.46
		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

Ref: TEMP-0821 Rev.37 FRONT / RECTO 1300019740-A