


LOT PX 419
Rev 1


CONTROL

 (Exp.) 2019-11-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.5	7.5			± 1.0	17.8	17.8			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.34	2.34			± 0.16	4.58	4.58			± 0.20	5.14	5.14			± 0.25			
HB HGB	g/dl	6.8	6.8			± 0.4	13.9	13.9			± 0.5	17.0	17.0			± 0.6			
	g/l	68	68			± 4	139	139			± 5	170	170			± 6			
	mmol/l	4.22	4.22			± 0.25	8.63	8.63			± 0.31	10.56	10.56			± 0.37			
HT HCT	%	19.7	19.7			± 1.5	38.5	38.5			± 2.0	46.8	46.8			± 2.5			
	l/l	0.197	0.197			± 0.015	0.385	0.385			± 0.020	0.468	0.468			± 0.025			
VGM MCV	µm ³ fl	84	84			± 5	84	84			± 5	91	91			± 5			
TGMH MCH	pg	29.1	29.1			± 2.0	30.3	30.3			± 2.0	33.1	33.1			± 2.5			
	fmol	1.80	1.80			± 0.12	1.88	1.88			± 0.12	2.05	2.05			± 0.16			
CCMH MCHC	g/dl	34.6	34.6			± 3.0	36.1	36.1			± 3.0	36.3	36.3			± 3.0			
	g/l	346	346			± 30	361	361			± 30	363	363			± 30			
	mmol/l	21.48	21.48			± 1.86	22.44	22.44			± 1.86	22.57	22.57			± 1.86			
IDR RDW	%	13.5	12.5			± 4.0	13.4	11.9			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68			± 20	250	250			± 30	490	490			± 50			
VPM MPV	µm ³ fl	10.0	10.3			± 2.0	9.7	10.2			± 2.0	9.7	10.2			± 2.0			
NEUT	#	1.45	1.45			± 0.35	4.34	4.34			± 0.90	12.82	12.82			± 1.90			
	%	57.8	57.8			± 10.0	57.8	57.8			± 10.0	72.0	72.0			± 10.0			
LYMPHO	#	0.70	0.70			± 0.33	2.30	2.30			± 0.70	2.67	2.67			± 1.50			
	%	28.0	28.0			± 12.0	30.6	30.6			± 8.0	15.0	15.0			± 8.0			
MONO	#	0.12	0.12			± 0.12	0.35	0.35			± 0.35	0.77	0.77			± 0.77			
	%	4.6	4.6			± 4.6	4.6	4.6			± 4.6	4.3	4.3			± 4.3			
EOS	#	0.16	0.16			± 0.16	0.26	0.26			± 0.26	0.73	0.73			± 0.73			
	%	6.3	6.3			± 6.3	3.5	3.5			± 3.5	4.1	4.1			± 4.1			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.82	0.82			± 0.82			
	%	3.3	3.3			± 3.3	3.5	3.5			± 3.5	4.6	4.6			± 4.6			

LOT PX 419
Rev 1

CONTROL

 (Exp.) 2019-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR	60 60C+ ES60	MS60	MSCR						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.6	7.5	7.6			± 1.0	18.3	17.6	18.2			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.36	2.38			± 0.16	4.57	4.55	4.59			± 0.20	5.12	5.10	5.15			± 0.25
		g/dl	6.8	6.7	6.8			± 0.4	13.9	13.7	13.8			± 0.5	16.9	16.7	16.9			± 0.6
HB	HGB	g/l	68	67	68			± 4	139	137	138			± 5	169	167	169			± 6
		mmol/l	4.22	4.16	4.22			± 0.25	8.63	8.51	8.57			± 0.31	10.49	10.37	10.49			± 0.37
HT	HCT	%	19.1	18.9	18.8			± 1.5	37.9	37.3	38.1			± 2.0	46.6	45.4	46.1			± 2.5
		l/l	0.191	0.189	0.188			± 0.015	0.379	0.373	0.381			± 0.020	0.466	0.454	0.461			± 0.025
VGM	MCV	µm ³ ; fl	80	80	79.0			± 5	83	82	83.0			± 5	91	89	89.5			± 5
TGMH	MCH	pg	28.5	28.4	28.6			± 2.0	30.4	30.1	30.1			± 2.0	33.0	32.7	32.8			± 2.5
		fmol	1.77	1.76	1.77			± 0.12	1.89	1.87	1.87			± 0.12	2.05	2.03	2.04			± 0.16
		g/dl	35.6	35.5	36.2			± 3.0	36.6	36.7	36.2			± 3.0	36.3	36.8	36.7			± 3.0
CCMH	MCHC	g/l	356	355	362			± 30	366	367	362			± 30	363	368	367			± 30
		mmol/l	22.09	22.04	22.46			± 1.86	22.76	22.80	22.49			± 1.86	22.53	22.85	22.77			± 1.86
IDR	RDW	%	13.5	13.0	12.5			± 4.0	13.0	12.5	12.5			± 4.0	12.5	12.0	12.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	68			± 20	250	251	245			± 30	485	500	475			± 50
VPM	MPV	µm ³ ; fl	9.8	9.9	9.0			± 2.0	9.5	9.5	9.0			± 2.0	10.0	10.0	9.0			± 2.0
		#	1.38	1.39	1.44			± 0.35	4.30	4.25	4.33			± 0.90	13.08	12.58	13.10			± 1.90
		%	57.4	58.1	57.7			± 10.0	56.6	56.7	57.0			± 10.0	71.5	71.5	72.0			± 10.0
		#	0.70	0.69	0.75			± 0.33	2.49	2.45	2.51			± 0.70	2.82	2.64	2.86			± 1.50
		%	29.2	28.9	29.8			± 12.0	32.8	32.6	33.0			± 8.0	15.4	15.0	15.7			± 8.0
		#	0.10	0.10	0.10			± 0.10	0.27	0.30	0.27			± 0.27	0.77	0.81	0.73			± 0.73
		%	4.3	4.0	4.0			± 4.0	3.6	4.0	3.5			± 3.5	4.2	4.6	4.0			± 4.0
		#	0.14	0.13	0.13			± 0.13	0.24	0.24	0.23			± 0.23	0.79	0.77	0.73			± 0.73
		%	5.7	5.6	5.0			± 5.0	3.1	3.2	3.0			± 3.0	4.3	4.4	4.0			± 4.0
		#	0.08	0.08	0.09			± 0.08	0.30	0.26	0.27			± 0.26	0.84	0.79	0.78			± 0.78
		%	3.4	3.4	3.5			± 3.4	3.9	3.5	3.5			± 3.5	4.6	4.5	4.3			± 4.3

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43