


LOT PX 410
Rev 2


CONTROL

 (Exp.) 2018-05-05
(YYYY - MM - DD)

		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		80 XL80	XLR				80 XL80	XLR				80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	17.5	17.5			± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.26	2.26			± 0.16	4.63	4.63			± 0.20	5.01	5.01			± 0.25	
HB HGB	g/dl	6.6	6.6			± 0.4	14.0	14.0			± 0.5	15.8	15.8			± 0.6	
	g/l	66	66			± 4	140	140			± 5	158	158			± 6	
	mmol/l	4.10	4.10			± 0.25	8.69	8.69			± 0.31	9.81	9.81			± 0.37	
HT HCT	%	19.0	19.0			± 1.5	38.9	38.9			± 2.0	43.6	43.6			± 2.5	
	l/l	0.190	0.190			± 0.015	0.389	0.389			± 0.020	0.436	0.436			± 0.025	
VGM MCV	µm ³ ; fl	84	84			± 5	84	84			± 5	87	87			± 5	
TGMH MCH	pg	29.2	29.2			± 2.0	30.2	30.2			± 2.0	31.5	31.5			± 2.5	
	fmol	1.81	1.81			± 0.12	1.88	1.88			± 0.12	1.96	1.96			± 0.16	
CCMH MCHC	g/dl	34.8	34.8			± 3.0	36.0	36.0			± 3.0	36.2	36.2			± 3.0	
	g/l	348	348			± 30	360	360			± 30	362	362			± 30	
	mmol/l	21.59	21.59			± 1.86	22.35	22.35			± 1.86	22.51	22.51			± 1.86	
IDR RDW	%	13.5	12.5			± 4.0	14.0	13.0			± 4.0	13.0	12.0			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68			± 20	245	245			± 30	486	486			± 50	
VPM MPV	µm ³ ; fl	9.6	9.6			± 2.0	9.8	9.8			± 2.0	9.3	9.3			± 2.0	
NEUT	#	1.39	1.39			± 0.35	4.15	4.15			± 0.90	12.04	12.04			± 1.90	
	%	55.7	55.7			± 10.0	56.1	56.1			± 10.0	68.8	68.8			± 10.0	
LYMPHO	#	0.75	0.75			± 0.33	2.34	2.34			± 0.70	2.92	2.92			± 1.50	
	%	30.0	30.0			± 12.0	31.6	31.6			± 8.0	16.7	16.7			± 8.0	
MONO	#	0.10	0.10			± 0.10	0.36	0.36			± 0.36	0.86	0.86			± 0.86	
	%	4.0	4.0			± 4.0	4.8	4.8			± 4.8	4.9	4.9			± 4.9	
EOS	#	0.18	0.18			± 0.18	0.28	0.28			± 0.28	0.91	0.91			± 0.91	
	%	7.1	7.1			± 7.1	3.8	3.8			± 3.8	5.2	5.2			± 5.2	
BASO	#	0.08	0.08			± 0.08	0.27	0.27			± 0.27	0.77	0.77			± 0.77	
	%	3.2	3.2			± 3.2	3.7	3.7			± 3.7	4.4	4.4			± 4.4	

LOT PX 410
Rev 2

CONTROL

 (Exp.) 2018-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.5	7.5	7.8			± 1.0	17.8	17.7	18.5			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.30	2.33			± 0.16	4.64	4.64	4.67			± 0.20	5.00	5.02	5.05			± 0.25
		g/dl	6.7	6.6	6.7			± 0.4	14.1	14.0	14.1			± 0.5	15.8	15.8	15.8			± 0.6
HB	HGB	g/l	67	66	67			± 4	141	140	141			± 5	158	158	158			± 6
		mmol/l	4.16	4.10	4.16			± 0.25	8.76	8.69	8.76			± 0.31	9.81	9.81	9.81			± 0.37
HT	HCT	%	19.0	18.6	18.6			± 1.5	38.5	38.5	38.8			± 2.0	43.5	43.2	43.4			± 2.5
		l/l	0.190	0.186	0.186			± 0.015	0.385	0.385	0.388			± 0.020	0.435	0.432	0.434			± 0.025
VGM	MCV	µm ³ ; fl	81	81	80.0			± 5	83	83	83.0			± 5	87	86	86.0			± 5
TGMH	MCH	pg	28.5	28.7	28.8			± 2.0	30.4	30.2	30.2			± 2.0	31.6	31.5	31.3			± 2.5
		fmol	1.77	1.78	1.79			± 0.12	1.89	1.87	1.87			± 0.12	1.96	1.95	1.94			± 0.16
		g/dl	35.2	35.4	35.9			± 3.0	36.6	36.4	36.4			± 3.0	36.3	36.6	36.4			± 3.0
CCMH	MCHC	g/l	352	354	359			± 30	366	364	364			± 30	363	366	364			± 30
		mmol/l	21.86	22.00	22.32			± 1.86	22.74	22.57	22.59			± 1.86	22.56	22.73	22.59			± 1.86
IDR	RDW	%	14.2	12.5	13.0			± 4.0	14.5	13.0	13.0			± 4.0	13.6	12.0	12.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	70			± 20	253	252	245			± 30	490	500	485			± 50
VPM	MPV	µm ³ ; fl	9.4	9.5	8.8			± 2.0	9.4	9.4	9.0			± 2.0	9.0	9.1	8.6			± 2.0
		#	1.34	1.33	1.41			± 0.35	4.22	4.18	4.45			± 0.90	12.48	12.32	13.14			± 1.90
		%	55.8	55.4	56.5			± 10.0	56.3	55.7	57.0			± 10.0	70.1	69.6	71.0			± 10.0
		#	0.72	0.72	0.75			± 0.33	2.39	2.35	2.46			± 0.70	2.88	2.81	2.96			± 1.50
		%	30.2	30.0	30.0			± 12.0	31.8	31.3	31.5			± 8.0	16.2	15.9	16.0			± 8.0
		#	0.10	0.11	0.10			± 0.10	0.35	0.42	0.35			± 0.35	0.82	0.97	0.83			± 0.82
		%	4.0	4.5	4.0			± 4.0	4.6	5.6	4.5			± 4.5	4.6	5.5	4.5			± 4.5
		#	0.16	0.17	0.15			± 0.15	0.28	0.29	0.27			± 0.27	0.84	0.81	0.83			± 0.81
		%	6.8	6.9	6.0			± 6.0	3.7	3.8	3.5			± 3.5	4.7	4.6	4.5			± 4.5
		#	0.08	0.08	0.09			± 0.08	0.27	0.27	0.27			± 0.27	0.78	0.78	0.74			± 0.74
		%	3.2	3.2	3.5			± 3.2	3.6	3.6	3.5			± 3.5	4.4	4.4	4.0			± 4.0

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42