


LOT PX 420
Rev 2


CONTROL

 (Exp.) 2020-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7			± 0.4	7.8	7.8			± 1.0	18.6	18.6			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.44	2.44			± 0.16	4.48	4.48			± 0.20	5.10	5.10			± 0.25			
HB HGB	g/dl	7.3	7.3			± 0.4	13.5	13.5			± 0.5	16.2	16.2			± 0.6			
	g/l	73	73			± 4	135	135			± 5	162	162			± 6			
	mmol/l	4.53	4.53			± 0.25	8.38	8.38			± 0.31	10.06	10.06			± 0.37			
HT HCT	%	21.0	21.0			± 1.5	37.6	37.6			± 2.0	44.4	44.4			± 2.5			
	l/l	0.210	0.210			± 0.015	0.376	0.376			± 0.020	0.444	0.444			± 0.025			
VGM MCV	µm ³ fl	86	86			± 5	84	84			± 5	87	87			± 5			
TGMH MCH	pg	29.9	29.9			± 2.0	30.1	30.1			± 2.0	31.8	31.8			± 2.5			
	fmol	1.86	1.86			± 0.12	1.87	1.87			± 0.12	1.97	1.97			± 0.16			
CCMH MCHC	g/dl	34.8	34.8			± 3.0	35.9	35.9			± 3.0	36.5	36.5			± 3.0			
	g/l	348	348			± 30	359	359			± 30	365	365			± 30			
	mmol/l	21.60	21.60			± 1.86	22.28	22.28			± 1.86	22.67	22.67			± 1.86			
IDR RDW	%	13.5	12.5			± 4.0	13.8	13.0			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65			± 20	241	241			± 30	500	500			± 50			
VPM MPV	µm ³ fl	9.3	9.4			± 2.0	9.1	9.4			± 2.0	10.0	10.5			± 2.0			
NEUT	#	1.49	1.49			± 0.35	4.27	4.27			± 0.90	12.72	12.72			± 1.90			
	%	55.2	55.2			± 10.0	54.7	54.7			± 10.0	68.4	68.4			± 10.0			
LYMPHO	#	0.73	0.73			± 0.33	2.29	2.29			± 0.70	2.66	2.66			± 1.50			
	%	26.9	26.9			± 12.0	29.4	29.4			± 8.0	14.3	14.3			± 8.0			
MONO	#	0.19	0.19			± 0.19	0.53	0.53			± 0.53	1.21	1.21			± 1.21			
	%	7.1	7.1			± 7.1	6.8	6.8			± 6.8	6.5	6.5			± 6.5			
EOS	#	0.20	0.20			± 0.20	0.44	0.44			± 0.44	1.19	1.19			± 1.19			
	%	7.4	7.4			± 7.4	5.7	5.7			± 5.7	6.4	6.4			± 6.4			
BASO	#	0.09	0.09			± 0.09	0.27	0.27			± 0.27	0.82	0.82			± 0.82			
	%	3.4	3.4			± 3.4	3.4	3.4			± 3.4	4.4	4.4			± 4.4			

LOT PX 420
Rev 2

CONTROL

 (Exp.) 2020-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRCP		60 60C+ ES60	MS60	MSCRCP		60 60C+ ES60	MS60	MSCRCP							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.7			± 0.4	7.9	7.8	7.9			± 1.0	18.9	18.7	19.0			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.50	2.45	2.45			± 0.16	4.48	4.45	4.45			± 0.20	5.08	5.05	5.08			± 0.25
		g/dl	7.3	7.2	7.1			± 0.4	13.5	13.3	13.3			± 0.5	16.0	16.0	15.9			± 0.6
HB	HGB	g/l	73	72	71			± 4	135	133	133			± 5	160	160	159			± 6
		mmol/l	4.53	4.47	4.41			± 0.25	8.38	8.26	8.26			± 0.31	9.94	9.94	9.87			± 0.37
HT	HCT	%	20.5	20.1	19.8			± 1.5	37.2	36.5	36.3			± 2.0	44.2	43.4	42.7			± 2.5
		l/l	0.205	0.201	0.198			± 0.015	0.372	0.365	0.363			± 0.020	0.442	0.434	0.427			± 0.025
VGM	MCV	µm ³ ·fl	82	82	81.0			± 5	83	82	81.5			± 5	87	86	84.0			± 5
TGMH	MCH	pg	29.2	29.4	29.0			± 2.0	30.1	29.9	29.9			± 2.0	31.5	31.7	31.3			± 2.5
		fmol	1.81	1.82	1.80			± 0.12	1.87	1.86	1.86			± 0.12	1.96	1.97	1.94			± 0.16
		g/dl	35.6	35.8	35.8			± 3.0	36.3	36.4	36.7			± 3.0	36.2	36.8	37.3			± 3.0
CCMH	MCHC	g/l	356	358	358			± 30	363	364	367			± 30	362	368	373			± 30
		mmol/l	22.11	22.26	22.22			± 1.86	22.55	22.63	22.77			± 1.86	22.48	22.88	23.14			± 1.86
IDR	RDW	%	13.0	13.0	12.5			± 4.0	13.0	13.0	12.5			± 4.0	12.5	12.0	12.0			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	66	65			± 20	243	247	235			± 30	503	513	480			± 50
VPM	MPV	µm ³ ·fl	9.1	9.3	8.7			± 2.0	9.1	9.2	8.5			± 2.0	10.2	10.4	9.2			± 2.0
		#	1.42	1.43	1.49			± 0.35	4.34	4.27	4.31			± 0.90	12.87	12.72	13.11			± 1.90
		%	54.8	55.1	55.0			± 10.0	54.9	54.7	54.5			± 10.0	68.1	68.0	69.0			± 10.0
		#	0.75	0.74	0.77			± 0.33	2.54	2.45	2.57			± 0.70	3.12	2.97	3.14			± 1.50
		%	28.8	28.5	28.5			± 12.0	32.1	31.4	32.5			± 8.0	16.5	15.9	16.5			± 8.0
		#	0.14	0.16	0.16			± 0.14	0.40	0.45	0.40			± 0.40	0.93	1.03	0.86			± 0.86
		%	5.5	6.0	6.0			± 5.5	5.0	5.8	5.0			± 5.0	4.9	5.5	4.5			± 4.5
		#	0.20	0.18	0.18			± 0.18	0.34	0.37	0.36			± 0.34	1.13	1.16	1.05			± 1.05
		%	7.5	7.0	6.5			± 6.5	4.3	4.7	4.5			± 4.3	6.0	6.2	5.5			± 5.5
		#	0.09	0.09	0.11			± 0.09	0.29	0.27	0.28			± 0.27	0.85	0.82	0.86			± 0.82
		%	3.4	3.4	4.0			± 3.4	3.7	3.4	3.5			± 3.4	4.5	4.4	4.5			± 4.4

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43