


**LOT** PX 414  
Rev 2

**CONTROL**

 (Exp.) 2019-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.1		± 0.4	7.1		7.1	± 1.0		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.49	2.49	± 0.16	4.73	4.73	± 0.20	5.19	5.19	± 0.25	
HB HGB	g/dl	7.1	7.1	± 0.4	14.0	14.0	± 0.5	16.3	16.3	± 0.6	
	g/l	71	71	± 4	140	140	± 5	163	163	± 6	
	mmol/l	4.41	4.41	± 0.25	8.69	8.69	± 0.31	10.12	10.12	± 0.37	
HT HCT	%	20.7	20.7	± 1.5	39.3	39.3	± 2.0	45.7	45.7	± 2.5	
	l/l	0.207	0.207	± 0.015	0.393	0.393	± 0.020	0.457	0.457	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	83	83	± 5	83	83	± 5	88	88	± 5	
TGMH MCH	pg	28.5	28.5	± 2.0	29.6	29.6	± 2.0	31.4	31.4	± 2.5	
	fmol	1.77	1.77	± 0.12	1.84	1.84	± 0.12	1.95	1.95	± 0.16	
CCMH MCHC	g/dl	34.4	34.4	± 3.0	35.7	35.7	± 3.0	35.7	35.7	± 3.0	
	g/l	344	344	± 30	357	357	± 30	357	357	± 30	
	mmol/l	21.33	21.33	± 1.86	22.15	22.15	± 1.86	22.16	22.16	± 1.86	
IDR-CV RDW-CV	%	12.5	12.5	± 4.0	12.5	12.5	± 4.0	12.0	12.0	± 4.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	75	± 20	270	270	± 30	525	525	± 50	
VPM MPV	µm <sup>3</sup> ; fl	11.0	11.0	± 2.0	11.0	11.0	± 2.0	10.5	10.5	± 2.0	
NEUT	#	1.17	1.17	± 0.35	3.83	3.83	± 0.90	11.26	11.26	± 1.90	
	%	55.7	55.7	± 10.0	54.0	54.0	± 10.0	65.1	65.1	± 10.0	
LYMPHO	#	0.43	0.43	± 0.33	1.59	1.59	± 0.70	1.78	1.78	± 1.50	
	%	20.7	20.7	± 12.0	22.4	22.4	± 8.0	10.3	10.3	± 8.0	
MONO	#	0.21	0.21	± 0.14	0.97	0.97	± 0.65	2.25	2.25	± 1.50	
	%	10.0	10.0	± 6.7	13.7	13.7	± 9.1	13.0	13.0	± 8.7	
EOS	#	0.19	0.19	± 0.19	0.37	0.37	± 0.37	1.18	1.18	± 1.18	
	%	8.9	8.9	± 8.9	5.2	5.2	± 5.2	6.8	6.8	± 6.8	
BASO	#	0.10	0.10	± 0.10	0.33	0.33	± 0.33	0.83	0.83	± 0.83	
	%	4.7	4.7	± 4.7	4.7	4.7	± 4.7	4.8	4.8	± 4.8	
IDR-SD RDW-SD	fl	38.0	38.0	± 8.0	38.0	38.0	± 8.0	40.5	40.5	± 8.0	
ERB NRBC	#	0.23	0.23	± 0.13	0.86	0.86	± 0.30	0.80	0.80	± 0.25	
	%	11.0	11.0	± 4.0	12.1	12.1	± 5.0	4.6	4.6	± 1.5	
PLTS OPT	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	82	± 30	N/A	255	± 40	N/A	500	± 50	

1300045049  
BACK / VERSO  
Ref. TEMP-0821 Rev.42