


**LOT** PX 428  
Rev 2

**CONTROL**

 (Exp.) 2021-05-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.13	2.13	± 0.40	6.80	6.80	± 1.00	17.50	17.50	± 2.20	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.46	2.46	± 0.16	4.73	4.73	± 0.20	5.24	5.24	± 0.25	
HB HGB	g/dl	6.9	6.9	± 0.4	14.2	14.2	± 0.5	16.7	16.7	± 0.6	
	g/l	69	69	± 4	142	142	± 5	167	167	± 6	
	mmol/l	4.28	4.28	± 0.25	8.82	8.82	± 0.31	10.37	10.37	± 0.37	
HT HCT	%	20.3	20.3	± 1.5	39.7	39.7	± 2.0	47.2	47.2	± 2.5	
	l/l	0.203	0.203	± 0.015	0.397	0.397	± 0.020	0.472	0.472	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	82.5	82.5	± 5.0	84.0	84.0	± 5.0	90.0	90.0	± 5.0	
TGMH MCH	pg	28.0	28.0	± 2.0	30.0	30.0	± 2.0	31.9	31.9	± 2.5	
	fmol	1.74	1.74	± 0.12	1.86	1.86	± 0.12	1.98	1.98	± 0.16	
CCMH MCHC	g/dl	34.0	34.0	± 3.0	35.7	35.7	± 3.0	35.4	35.4	± 3.0	
	g/l	340	340	± 30	357	357	± 30	354	354	± 30	
	mmol/l	21.11	21.11	± 1.86	22.19	22.19	± 1.86	21.99	21.99	± 1.86	
IDR-CV RDW-CV	%	12.5	12.5	± 4.0	12.0	12.0	± 4.0	11.5	11.5	± 4.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	71	71	± 20	250	250	± 30	490	490	± 50	
VPM MPV	µm <sup>3</sup> ; fl	10.0	10.0	± 2.0	9.8	9.8	± 2.0	10.3	10.3	± 2.0	
NEUT	#	1.15	1.15	± 0.35	3.57	3.57	± 0.90	11.06	11.06	± 1.90	
	%	54.0	54.0	± 10.0	52.5	52.5	± 10.0	63.2	63.2	± 10.0	
LYMPHO	#	0.57	0.57	± 0.33	1.75	1.75	± 0.70	2.10	2.10	± 1.50	
	%	26.7	26.7	± 12.0	25.7	25.7	± 8.0	12.0	12.0	± 8.0	
MONO	#	0.15	0.15	± 0.15	0.85	0.85	± 0.77	2.77	2.77	± 2.49	
	%	7.2	7.2	± 7.2	12.5	12.5	± 11.3	15.8	15.8	± 14.2	
EOS	#	0.16	0.16	± 0.16	0.33	0.33	± 0.33	0.75	0.75	± 0.75	
	%	7.4	7.4	± 7.4	4.8	4.8	± 4.8	4.3	4.3	± 4.3	
BASO	#	0.10	0.10	± 0.10	0.31	0.31	± 0.31	0.82	0.82	± 0.82	
	%	4.7	4.7	± 4.7	4.5	4.5	± 4.5	4.7	4.7	± 4.7	
IDR-SD RDW-SD	fl	37.0	37.0	± 4.0	37.0	37.0	± 4.0	39.0	39.0	± 4.0	
ERB NRBC	#	0.33	0.33	± 0.13	0.97	0.97	± 0.30	1.03	1.03	± 0.25	
	%	15.4	15.4	± 4.0	14.3	14.3	± 5.0	5.9	5.9	± 1.5	
PLTS OPT	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	78	± 30	N/A	253	± 40	N/A	500	± 50	

1300074625  
BACK / VERSO  
Ref. TEMP-0821 Rev.44