


LOT PX 427
Rev 2

CONTROL

 (Exp.) 2021-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.25	2.25	± 0.40		6.82	6.82		± 1.00	17.15		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.39	± 0.16	4.55	4.55	± 0.20	5.16	5.16	± 0.25		
HB HGB	g/dl	6.6	6.6	± 0.4	13.6	13.6	± 0.5	16.7	16.7	± 0.6		
	g/l	66	66	± 4	136	136	± 5	167	167	± 6		
	mmol/l	4.10	4.10	± 0.25	8.45	8.45	± 0.31	10.37	10.37	± 0.37		
HT HCT	%	19.4	19.4	± 1.5	37.8	37.8	± 2.0	47.0	47.0	± 2.5		
	l/l	0.194	0.194	± 0.015	0.378	0.378	± 0.020	0.470	0.470	± 0.025		
VGM MCV	µm ³ ; fl	81.0	81.0	± 5.0	83.0	83.0	± 5.0	91.0	91.0	± 5.0		
TGMH MCH	pg	27.6	27.6	± 2.0	29.9	29.9	± 2.0	32.4	32.4	± 2.5		
	fmol	1.71	1.71	± 0.12	1.86	1.86	± 0.12	2.01	2.01	± 0.16		
CCMH MCHC	g/dl	34.1	34.1	± 3.0	36.0	36.0	± 3.0	35.6	35.6	± 3.0		
	g/l	341	341	± 30	360	360	± 30	356	356	± 30		
	mmol/l	21.17	21.17	± 1.86	22.36	22.36	± 1.86	22.09	22.09	± 1.86		
IDR-CV RDW-CV	%	12.4	12.4	± 4.0	13.5	13.5	± 4.0	12.8	12.8	± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	71	71	± 20	252	252	± 30	490	490	± 50		
VPM MPV	µm ³ ; fl	9.5	9.5	± 2.0	9.5	9.5	± 2.0	10.0	10.0	± 2.0		
NEUT	#	1.26	1.26	± 0.35	3.46	3.46	± 0.90	10.67	10.67	± 1.90		
	%	56.0	56.0	± 10.0	50.7	50.7	± 10.0	62.2	62.2	± 10.0		
LYMPHO	#	0.44	0.44	± 0.33	1.48	1.48	± 0.70	1.66	1.66	± 1.50		
	%	19.6	19.6	± 12.0	21.7	21.7	± 8.0	9.7	9.7	± 8.0		
MONO	#	0.29	0.29	± 0.29	1.30	1.30	± 1.17	3.16	3.16	± 2.85		
	%	13.0	13.0	± 13.0	19.0	19.0	± 17.1	18.4	18.4	± 16.6		
EOS	#	0.15	0.15	± 0.15	0.27	0.27	± 0.27	0.86	0.86	± 0.86		
	%	6.6	6.6	± 6.6	3.9	3.9	± 3.9	5.0	5.0	± 5.0		
BASO	#	0.11	0.11	± 0.11	0.32	0.32	± 0.32	0.81	0.81	± 0.81		
	%	4.8	4.8	± 4.8	4.7	4.7	± 4.7	4.7	4.7	± 4.7		
IDR-SD RDW-SD	fl	38.0	38.0	± 4.0	40.0	40.0	± 4.0	42.5	42.5	± 4.0		
ERB NRBC	#	0.25	0.25	± 0.13	0.80	0.80	± 0.30	0.86	0.86	± 0.25		
	%	11.0	11.0	± 4.0	11.7	11.7	± 5.0	5.0	5.0	± 1.5		
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	80	± 30	N/A	264	± 40	N/A	520	± 50		

1300074625
BACK / VERSO
Ref. TEMP-0821 Rev.44