


LOT PX 416
Rev 1

CONTROL

 (Exp.) 2019-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.10	2.10	± 0.40		6.80	6.80		± 1.00	17.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.49	2.49	± 0.16	4.65	4.65	± 0.20	5.31	5.31	± 0.25		
HB HGB	g/dl	6.9	6.9	± 0.4	13.8	13.8	± 0.5	16.7	16.7	± 0.6		
	g/l	69	69	± 4	138	138	± 5	167	167	± 6		
	mmol/l	4.28	4.28	± 0.25	8.57	8.57	± 0.31	10.37	10.37	± 0.37		
HT HCT	%	20.4	20.4	± 1.5	39.1	39.1	± 2.0	47.5	47.5	± 2.5		
	l/l	0.204	0.204	± 0.015	0.391	0.391	± 0.020	0.475	0.475	± 0.025		
VGM MCV	µm ³ ; fl	82.0	82.0	± 5.0	84.0	84.0	± 5.0	89.5	89.5	± 5.0		
TGMH MCH	pg	27.7	27.7	± 2.0	29.7	29.7	± 2.0	31.5	31.5	± 2.5		
	fmol	1.72	1.72	± 0.12	1.84	1.84	± 0.12	1.95	1.95	± 0.16		
CCMH MCHC	g/dl	33.8	33.8	± 3.0	35.3	35.3	± 3.0	35.1	35.1	± 3.0		
	g/l	338	338	± 30	353	353	± 30	351	351	± 30		
	mmol/l	20.99	20.99	± 1.86	21.94	21.94	± 1.86	21.82	21.82	± 1.86		
IDR-CV RDW-CV	%	12.0	12.0	± 4.0	12.0	12.0	± 4.0	12.5	12.5	± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	75	± 20	256	256	± 30	526	526	± 50		
VPM MPV	µm ³ ; fl	10.5	10.5	± 2.0	9.8	9.8	± 2.0	10.1	10.1	± 2.0		
NEUT	#	1.27	1.27	± 0.35	3.81	3.81	± 0.90	11.71	11.71	± 1.90		
	%	60.5	60.5	± 10.0	56.0	56.0	± 10.0	68.9	68.9	± 10.0		
LYMPHO	#	0.43	0.43	± 0.33	1.62	1.62	± 0.70	1.92	1.92	± 1.50		
	%	20.5	20.5	± 12.0	23.8	23.8	± 8.0	11.3	11.3	± 8.0		
MONO	#	0.14	0.14	± 0.11	0.74	0.74	± 0.56	1.60	1.60	± 1.20		
	%	6.7	6.7	± 5.0	10.9	10.9	± 8.2	9.4	9.4	± 7.1		
EOS	#	0.16	0.16	± 0.16	0.31	0.31	± 0.31	0.94	0.94	± 0.94		
	%	7.5	7.5	± 7.5	4.6	4.6	± 4.6	5.5	5.5	± 5.5		
BASO	#	0.10	0.10	± 0.10	0.32	0.32	± 0.32	0.83	0.83	± 0.83		
	%	4.8	4.8	± 4.8	4.7	4.7	± 4.7	4.9	4.9	± 4.9		
IDR-SD RDW-SD	fl	37.5	37.5	± 4.0	39.5	39.5	± 4.0	43.0	43.0	± 4.0		
ERB NRBC	#	0.24	0.24	± 0.13	0.89	0.89	± 0.30	0.87	0.87	± 0.25		
	%	11.4	11.4	± 4.0	13.1	13.1	± 5.0	5.1	5.1	± 1.5		
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	75	± 30	N/A	247	± 40	N/A	495	± 50		

1300060883
BACK / VERSO
Ref: TEMP-0821 Rev.43