


**LOT** PX 425  
Rev 1

**CONTROL**

 (Exp.) 2020-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.05	2.05	± 0.40	6.65	6.65	± 1.00	17.00	17.00	± 2.20	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.34	± 0.16	4.57	4.57	± 0.20	5.10	5.10	± 0.25	
HB	HGB	g/dl	6.5	6.5	± 0.4	13.8	13.8	± 0.5	16.0	16.0	± 0.6	
		g/l	65	65	± 4	138	138	± 5	160	160	± 6	
		mmol/l	4.04	4.04	± 0.25	8.57	8.57	± 0.31	9.94	9.94	± 0.37	
HT	HCT	%	19.1	19.1	± 1.5	38.6	38.6	± 2.0	45.1	45.1	± 2.5	
		l/l	0.191	0.191	± 0.015	0.386	0.386	± 0.020	0.451	0.451	± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	81.5	81.5	± 5.0	84.5	84.5	± 5.0	88.5	88.5	± 5.0	
TGMH	MCH	pg	27.8	27.8	± 2.0	30.2	30.2	± 2.0	31.4	31.4	± 2.5	
		fmol	1.73	1.73	± 0.12	1.88	1.88	± 0.12	1.95	1.95	± 0.16	
CCMH	MCHC	g/dl	34.1	34.1	± 3.0	35.7	35.7	± 3.0	35.4	35.4	± 3.0	
		g/l	341	341	± 30	357	357	± 30	354	354	± 30	
		mmol/l	21.17	21.17	± 1.86	22.19	22.19	± 1.86	22.01	22.01	± 1.86	
IDR-CV	RDW-CV	%	11.8	11.8	± 4.0	11.3	11.3	± 4.0	11.2	11.2	± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	65	65	± 20	260	260	± 30	500	500	± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.7	9.7	± 2.0	10.3	10.3	± 2.0	10.1	10.1	± 2.0	
NEUT		#	1.17	1.17	± 0.35	3.64	3.64	± 0.90	11.29	11.29	± 1.90	
		%	57.0	57.0	± 10.0	54.8	54.8	± 10.0	66.4	66.4	± 10.0	
LYMPHO		#	0.50	0.50	± 0.33	1.66	1.66	± 0.70	2.13	2.13	± 1.50	
		%	24.3	24.3	± 12.0	25.0	25.0	± 8.0	12.5	12.5	± 8.0	
MONO		#	0.14	0.14	± 0.14	0.73	0.73	± 0.66	1.79	1.79	± 1.62	
		%	6.6	6.6	± 6.6	11.0	11.0	± 9.9	10.5	10.5	± 9.5	
EOS		#	0.15	0.15	± 0.15	0.30	0.30	± 0.30	0.99	0.99	± 0.99	
		%	7.4	7.4	± 7.4	4.5	4.5	± 4.5	5.8	5.8	± 5.8	
BASO		#	0.10	0.10	± 0.10	0.31	0.31	± 0.31	0.82	0.82	± 0.82	
		%	4.7	4.7	± 4.7	4.7	4.7	± 4.7	4.8	4.8	± 4.8	
IDR-SD	RDW-SD	fl	37.0	37.0	± 4.0	37.5	37.5	± 4.0	40.5	40.5	± 4.0	
ERB	NRBC	#	0.26	0.26	± 0.13	0.88	0.88	± 0.30	0.90	0.90	± 0.25	
		%	12.7	12.7	± 4.0	13.3	13.3	± 5.0	5.3	5.3	± 1.5	
PLTS	OPT	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	75	± 30	N/A	265	± 40	N/A	510	± 50	

1300074625  
BACK / VERSO  
Ref: TEMP-0821 Rev.44