

ABX Difftrol



LOT PX 415
Rev 3

CONTROL

(Exp.) 2019-03-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1	± 0.4	6.6	6.6	± 1.0	17.4	17.4	± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.49	2.49	± 0.16	4.67	4.67	± 0.20	5.20	5.20	± 0.25	
HB HGB	g/dl	7.0	7.0	± 0.4	13.8	13.8	± 0.5	16.0	16.0	± 0.6	
	g/l	70	70	± 4	138	138	± 5	160	160	± 6	
	mmol/l	4.35	4.35	± 0.25	8.57	8.57	± 0.31	9.94	9.94	± 0.37	
HT HCT	%	20.4	20.4	± 1.5	38.8	38.8	± 2.0	45.8	45.8	± 2.5	
	l/l	0.204	0.204	± 0.015	0.388	0.388	± 0.020	0.458	0.458	± 0.025	
VGM MCV	µm ³ ; fl	82	82	± 5	83	83	± 5	88	88	± 5	
TGMH MCH	pg	28.1	28.1	± 2.0	29.6	29.6	± 2.0	30.8	30.8	± 2.5	
	fmol	1.75	1.75	± 0.12	1.84	1.84	± 0.12	1.91	1.91	± 0.16	
CCMH MCHC	g/dl	34.3	34.3	± 3.0	35.6	35.6	± 3.0	35.0	35.0	± 3.0	
	g/l	343	343	± 30	356	356	± 30	350	350	± 30	
	mmol/l	21.29	21.29	± 1.86	22.11	22.11	± 1.86	21.71	21.71	± 1.86	
IDR-CV RDW-CV	%	11.5	11.5	± 4.0	11.7	11.7	± 4.0	12.3	12.3	± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68	± 20	261	261	± 30	525	525	± 50	
VPM MPV	µm ³ ; fl	9.5	9.5	± 2.0	9.9	9.9	± 2.0	10.3	10.3	± 2.0	
NEUT	#	1.25	1.25	± 0.35	3.95	3.95	± 0.90	12.46	12.46	± 1.90	
	%	59.6	59.6	± 10.0	59.8	59.8	± 10.0	71.6	71.6	± 10.0	
LYMPHO	#	0.43	0.43	± 0.33	1.54	1.54	± 0.70	1.79	1.79	± 1.50	
	%	20.7	20.7	± 12.0	23.4	23.4	± 8.0	10.3	10.3	± 8.0	
MONO	#	0.13	0.13	± 0.13	0.50	0.50	± 0.33	1.37	1.37	± 0.91	
	%	6.1	6.1	± 6.1	7.6	7.6	± 5.1	7.9	7.9	± 5.3	
EOS	#	0.19	0.19	± 0.19	0.30	0.30	± 0.30	0.94	0.94	± 0.94	
	%	8.9	8.9	± 8.9	4.5	4.5	± 4.5	5.4	5.4	± 5.4	
BASO	#	0.10	0.10	± 0.10	0.31	0.31	± 0.21	0.84	0.84	± 0.42	
	%	4.7	4.7	± 4.7	4.7	4.7	± 4.7	4.8	4.8	± 4.8	
IDR-SD RDW-SD	fl	37.5	37.5	± 8.0	38.0	38.0	± 8.0	41.0	41.0	± 8.0	
ERB NRBC	#	0.23	0.23	± 0.13	0.88	0.88	± 0.30	0.84	0.84	± 0.25	
	%	10.8	10.8	± 4.0	13.4	13.4	± 5.0	4.8	4.8	± 1.5	
PLTS OPT	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	71	± 30	N/A	250	± 40	N/A	503	± 50	

1300045049 BACK / VERSO Ref: TEMP-0821 Rev.42