


LOT PX 424
Rev 1

CONTROL

 (Exp.) 2020-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.10	2.10		± 0.40	6.60		6.60	± 1.00		
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.45	2.45	± 0.16	4.63	4.63	± 0.20	5.07	5.07	± 0.25	
HB	HGB	g/dl	6.6	6.6	± 0.4	13.8	13.8	± 0.5	16.0	16.0	± 0.6	
		g/l	66	66	± 4	138	138	± 5	160	160	± 6	
		mmol/l	4.10	4.10	± 0.25	8.57	8.57	± 0.31	9.94	9.94	± 0.37	
HT	HCT	%	19.5	19.5	± 1.5	38.4	38.4	± 2.0	44.6	44.6	± 2.5	
		l/l	0.195	0.195	± 0.015	0.384	0.384	± 0.020	0.446	0.446	± 0.025	
VGM	MCV	µm ³ ; fl	79.5	79.5	± 5.0	83.0	83.0	± 5.0	88.0	88.0	± 5.0	
TGMH	MCH	pg	26.9	26.9	± 2.0	29.8	29.8	± 2.0	31.6	31.6	± 2.5	
		fmol	1.67	1.67	± 0.12	1.85	1.85	± 0.12	1.96	1.96	± 0.16	
CCMH	MCHC	g/dl	33.9	33.9	± 3.0	35.9	35.9	± 3.0	35.9	35.9	± 3.0	
		g/l	339	339	± 30	359	359	± 30	359	359	± 30	
		mmol/l	21.04	21.04	± 1.86	22.30	22.30	± 1.86	22.27	22.27	± 1.86	
IDR-CV	RDW-CV	%	12.0	12.0	± 4.0	11.5	11.5	± 4.0	11.5	11.5	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68	± 20	255	255	± 30	518	518	± 50	
VPM	MPV	µm ³ ; fl	10.4	10.4	± 2.0	10.4	10.4	± 2.0	10.5	10.5	± 2.0	
NEUT		#	1.24	1.24	± 0.35	3.70	3.70	± 0.90	11.90	11.90	± 1.90	
		%	58.9	58.9	± 10.0	56.0	56.0	± 10.0	70.0	70.0	± 10.0	
LYMPHO		#	0.49	0.49	± 0.33	1.79	1.79	± 0.70	2.07	2.07	± 1.50	
		%	23.2	23.2	± 12.0	27.1	27.1	± 8.0	12.2	12.2	± 8.0	
MONO		#	0.13	0.13	± 0.13	0.52	0.52	± 0.47	1.34	1.34	± 1.21	
		%	6.1	6.1	± 6.1	7.9	7.9	± 7.1	7.9	7.9	± 7.1	
EOS		#	0.15	0.15	± 0.15	0.28	0.28	± 0.28	0.87	0.87	± 0.87	
		%	7.0	7.0	± 7.0	4.2	4.2	± 4.2	5.1	5.1	± 5.1	
BASO		#	0.10	0.10	± 0.10	0.32	0.32	± 0.32	0.82	0.82	± 0.82	
		%	4.8	4.8	± 4.8	4.8	4.8	± 4.8	4.8	4.8	± 4.8	
IDR-SD	RDW-SD	fl	37.0	37.0	± 4.0	37.5	37.5	± 4.0	39.5	39.5	± 4.0	
ERB	NRBC	#	0.29	0.29	± 0.13	1.06	1.06	± 0.30	0.99	0.99	± 0.25	
		%	14.0	14.0	± 4.0	16.0	16.0	± 5.0	5.8	5.8	± 1.5	
PLTS	OPT	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	72	± 30	N/A	258	± 40	N/A	515	± 50	

1300074625
BACK / VERSO
Ref. TEMP-0821 Rev.44