


**LOT** PX 408  
Rev 1

**CONTROL**

 (Exp.) 2018-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.2	2.2		± 0.4	7.0		7.0	± 1.0		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.34	± 0.16	4.46	4.46	± 0.20	5.06	5.06	± 0.25	
HB HGB	g/dl	6.5	6.5	± 0.4	12.7	12.7	± 0.5	15.8	15.8	± 0.6	
	g/l	65	65	± 4	127	127	± 5	158	158	± 6	
	mmol/l	4.04	4.04	± 0.25	7.89	7.89	± 0.31	9.81	9.81	± 0.37	
HT HCT	%	19.3	19.3	± 1.5	36.4	36.4	± 2.0	44.7	44.7	± 2.5	
	l/l	0.193	0.193	± 0.015	0.364	0.364	± 0.020	0.447	0.447	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	83	83	± 5	82	82	± 5	88	88	± 5	
TGMH MCH	pg	27.8	27.8	± 2.0	28.5	28.5	± 2.0	31.2	31.2	± 2.5	
	fmol	1.73	1.73	± 0.12	1.77	1.77	± 0.12	1.94	1.94	± 0.16	
CCMH MCHC	g/dl	33.5	33.5	± 3.0	34.7	34.7	± 3.0	35.5	35.5	± 3.0	
	g/l	335	335	± 30	347	347	± 30	355	355	± 30	
	mmol/l	20.78	20.78	± 1.86	21.56	21.56	± 1.86	22.04	22.04	± 1.86	
IDR-CV RDW-CV	%	15.1	15.1	± 4.0	14.9	14.9	± 4.0	15.1	15.1	± 4.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	73	± 20	254	254	± 30	497	497	± 50	
VPM MPV	µm <sup>3</sup> ; fl	10.4	10.4	± 2.0	10.2	10.2	± 2.0	9.4	9.4	± 2.0	
NEUT	#	1.11	1.11	± 0.35	3.56	3.56	± 0.90	10.40	10.40	± 1.90	
	%	50.3	50.3	± 10.0	50.6	50.6	± 10.0	60.2	60.2	± 10.0	
LYMPHO	#	0.52	0.52	± 0.33	1.62	1.62	± 0.70	2.01	2.01	± 1.50	
	%	23.7	23.7	± 12.0	23.1	23.1	± 8.0	11.6	11.6	± 8.0	
MONO	#	0.37	0.37	± 0.25	1.26	1.26	± 0.84	3.39	3.39	± 2.26	
	%	16.8	16.8	± 11.2	18.0	18.0	± 12.0	19.6	19.6	± 13.1	
EOS	#	0.10	0.10	± 0.10	0.25	0.25	± 0.25	0.66	0.66	± 0.66	
	%	4.6	4.6	± 4.6	3.6	3.6	± 3.6	3.8	3.8	± 3.8	
BASO	#	0.10	0.10	± 0.10	0.33	0.33	± 0.22	0.82	0.82	± 0.41	
	%	4.6	4.6	± 2.5	4.7	4.7	± 3.0	4.8	4.8	± 2.5	
IDR-SD RDW-SD	fl	39.5	39.5	± 4.0	40.0	40.0	± 4.0	44.6	44.6	± 4.0	
ERB NRBC	#	0.30	0.30	± 0.13	0.88	0.88	± 0.30	0.87	0.87	± 0.25	
	%	13.7	13.7	± 4.0	12.5	12.5	± 5.0	5.0	5.0	± 1.5	
PLTS OPT	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	74	± 30	N/A	248	± 40	N/A	475	± 50	

1300040597  
BACK / VERSO  
Ref. TEMP-0821 Rev.41