


**LOT** PX 420  
Rev 2

**CONTROL**

 (Exp.) 2020-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.30	2.30	± 0.40	6.85	6.85	± 1.00	17.90	17.90	± 2.20	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.54	2.54	± 0.16	4.51	4.51	± 0.20	5.17	5.17	± 0.25	
HB HGB	g/dl	7.3	7.3	± 0.4	13.5	13.5	± 0.5	16.1	16.1	± 0.6	
	g/l	73	73	± 4	135	135	± 5	161	161	± 6	
	mmol/l	4.53	4.53	± 0.25	8.38	8.38	± 0.31	10.00	10.00	± 0.37	
HT HCT	%	21.3	21.3	± 1.5	37.9	37.9	± 2.0	45.5	45.5	± 2.5	
	l/l	0.213	0.213	± 0.015	0.379	0.379	± 0.020	0.455	0.455	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	84.0	84.0	± 5.0	84.0	84.0	± 5.0	88.0	88.0	± 5.0	
TGMH MCH	pg	28.7	28.7	± 2.0	29.9	29.9	± 2.0	31.1	31.1	± 2.5	
	fmol	1.78	1.78	± 0.12	1.86	1.86	± 0.12	1.93	1.93	± 0.16	
CCMH MCHC	g/dl	34.2	34.2	± 3.0	35.6	35.6	± 3.0	35.4	35.4	± 3.0	
	g/l	342	342	± 30	356	356	± 30	354	354	± 30	
	mmol/l	21.25	21.25	± 1.86	22.13	22.13	± 1.86	21.98	21.98	± 1.86	
IDR-CV RDW-CV	%	11.5	11.5	± 4.0	12.0	12.0	± 4.0	11.5	11.5	± 4.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68	68	± 20	250	250	± 30	538	538	± 50	
VPM MPV	µm <sup>3</sup> ; fl	10.0	10.0	± 2.0	9.8	9.8	± 2.0	11.3	11.3	± 2.0	
NEUT	#	1.38	1.38	± 0.35	3.93	3.93	± 0.90	11.99	11.99	± 1.90	
	%	59.8	59.8	± 10.0	57.4	57.4	± 10.0	67.0	67.0	± 10.0	
LYMPHO	#	0.43	0.43	± 0.33	1.47	1.47	± 0.70	1.79	1.79	± 1.50	
	%	18.9	18.9	± 12.0	21.5	21.5	± 8.0	10.0	10.0	± 8.0	
MONO	#	0.17	0.17	± 0.17	0.73	0.73	± 0.64	2.08	2.08	± 1.87	
	%	7.5	7.5	± 7.5	10.7	10.7	± 9.3	11.6	11.6	± 10.4	
EOS	#	0.21	0.21	± 0.21	0.40	0.40	± 0.40	1.18	1.18	± 1.18	
	%	9.0	9.0	± 9.0	5.9	5.9	± 5.9	6.6	6.6	± 6.6	
BASO	#	0.11	0.11	± 0.11	0.31	0.31	± 0.31	0.86	0.86	± 0.86	
	%	4.8	4.8	± 4.8	4.5	4.5	± 4.5	4.8	4.8	± 4.8	
IDR-SD RDW-SD	fl	37.0	37.0	± 4.0	38.0	38.0	± 4.0	39.5	39.5	± 4.0	
ERB NRBC	#	0.21	0.21	± 0.13	0.79	0.79	± 0.30	0.84	0.84	± 0.25	
	%	9.3	9.3	± 4.0	11.5	11.5	± 5.0	4.7	4.7	± 1.5	
PLTS OPT	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	79	± 30	N/A	258	± 40	N/A	540	± 50	

1300060883  
BACK / VERSO  
Ref. TEMP-0821 Rev.43