


LOT PX 411
Rev 1

CONTROL

 (Exp.) 2018-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1	± 0.4	6.6	6.6	± 1.0	16.8	16.8	± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	± 0.16	4.70	4.70	± 0.20	5.24	5.24	± 0.25	
HB HGB	g/dl	6.9	6.9	± 0.4	13.9	13.9	± 0.5	16.2	16.2	± 0.6	
	g/l	69	69	± 4	139	139	± 5	162	162	± 6	
	mmol/l	4.28	4.28	± 0.25	8.63	8.63	± 0.31	10.06	10.06	± 0.37	
HT HCT	%	19.7	19.7	± 1.5	39.0	39.0	± 2.0	45.6	45.6	± 2.5	
	l/l	0.197	0.197	± 0.015	0.390	0.390	± 0.020	0.456	0.456	± 0.025	
VGM MCV	µm ³ ; fl	82	82	± 5	83	83	± 5	87	87	± 5	
TGMH MCH	pg	28.8	28.8	± 2.0	29.6	29.6	± 2.0	30.9	30.9	± 2.5	
	fmol	1.79	1.79	± 0.12	1.84	1.84	± 0.12	1.92	1.92	± 0.16	
CCMH MCHC	g/dl	35.1	35.1	± 3.0	35.6	35.6	± 3.0	35.5	35.5	± 3.0	
	g/l	351	351	± 30	356	356	± 30	355	355	± 30	
	mmol/l	21.77	21.77	± 1.86	22.13	22.13	± 1.86	22.07	22.07	± 1.86	
IDR-CV RDW-CV	%	14.0	14.0	± 4.0	14.5	14.5	± 4.0	13.5	13.5	± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	± 20	265	265	± 30	530	530	± 50	
VPM MPV	µm ³ ; fl	10.5	10.5	± 2.0	10.0	10.0	± 2.0	10.0	10.0	± 2.0	
NEUT	#	1.25	1.25	± 0.35	3.96	3.96	± 0.90	11.83	11.83	± 1.90	
	%	59.4	59.4	± 10.0	60.0	60.0	± 10.0	70.4	70.4	± 10.0	
LYMPHO	#	0.49	0.49	± 0.33	1.58	1.58	± 0.70	1.81	1.81	± 1.50	
	%	23.3	23.3	± 12.0	23.9	23.9	± 8.0	10.8	10.8	± 8.0	
MONO	#	0.14	0.14	± 0.09	0.53	0.53	± 0.35	1.53	1.53	± 1.02	
	%	6.5	6.5	± 4.3	8.1	8.1	± 5.4	9.1	9.1	± 6.1	
EOS	#	0.13	0.13	± 0.13	0.24	0.24	± 0.24	0.86	0.86	± 0.86	
	%	6.2	6.2	± 6.2	3.6	3.6	± 3.6	5.1	5.1	± 5.1	
BASO	#	0.10	0.10	± 0.10	0.29	0.29	± 0.29	0.77	0.77	± 0.77	
	%	4.6	4.6	± 4.6	4.4	4.4	± 4.4	4.6	4.6	± 4.6	
IDR-SD RDW-SD	fl	37.0	37.0	± 4.0	38.5	38.5	± 4.0	40.0	40.0	± 4.0	
ERB NRBC	#	0.27	0.27	± 0.13	0.86	0.86	± 0.30	0.84	0.84	± 0.25	
	%	13.0	13.0	± 4.0	13.0	13.0	± 5.0	5.0	5.0	± 1.5	
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	75	± 30	N/A	250	± 40	N/A	475	± 50	

1300045049
BACK / VERSO
Ref. TEMP-0821 Rev.42