


LOT PX 433
Rev 1

CONTROL

 (Exp.) 2022-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.10	2.10	± 0.40	6.55	6.55	± 1.00	16.50	16.50	± 2.20	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.43	± 0.16	4.59	4.59	± 0.20	5.18	5.18	± 0.25	
HB	HGB	g/dl	6.8	6.8	± 0.4	14.0	14.0	± 0.5	16.8	16.8	± 0.6	
		g/l	68	68	± 4	140	140	± 5	168	168	± 6	
		mmol/l	4.22	4.22	± 0.25	8.69	8.69	± 0.31	10.43	10.43	± 0.37	
HT	HCT	%	19.9	19.9	± 1.5	38.6	38.6	± 2.0	47.1	47.1	± 2.5	
		l/l	0.199	0.199	± 0.015	0.386	0.386	± 0.020	0.471	0.471	± 0.025	
VGM	MCV	µm ³ ; fl	82.0	82.0	± 5.0	84.0	84.0	± 5.0	91.0	91.0	± 5.0	
TGMH	MCH	pg	28.0	28.0	± 2.0	30.5	30.5	± 2.0	32.4	32.4	± 2.5	
		fmol	1.74	1.74	± 0.12	1.89	1.89	± 0.12	2.01	2.01	± 0.16	
CCMH	MCHC	g/dl	34.1	34.1	± 3.0	36.3	36.3	± 3.0	35.6	35.6	± 3.0	
		g/l	341	341	± 30	363	363	± 30	356	356	± 30	
		mmol/l	21.19	21.19	± 1.86	22.55	22.55	± 1.86	22.13	22.13	± 1.86	
IDR-CV	RDW-CV	%	13.0	13.0	± 4.0	12.0	12.0	± 4.0	12.0	12.0	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	77	77	± 20	266	266	± 30	510	510	± 50	
VPM	MPV	µm ³ ; fl	10.5	10.5	± 2.0	10.2	10.2	± 2.0	10.0	10.0	± 2.0	
NEUT		#	1.25	1.25	± 0.35	4.00	4.00	± 0.90	12.06	12.06	± 1.90	
		%	59.6	59.6	± 10.0	61.1	61.1	± 10.0	73.1	73.1	± 10.0	
LYMPHO		#	0.48	0.48	± 0.33	1.66	1.66	± 0.70	1.91	1.91	± 1.50	
		%	22.9	22.9	± 12.0	25.3	25.3	± 8.0	11.6	11.6	± 8.0	
MONO		#	0.10	0.10	± 0.10	0.30	0.30	± 0.27	0.94	0.94	± 0.85	
		%	4.6	4.6	± 4.6	4.6	4.6	± 4.1	5.7	5.7	± 5.1	
EOS		#	0.17	0.17	± 0.17	0.28	0.28	± 0.28	0.81	0.81	± 0.81	
		%	8.0	8.0	± 8.0	4.2	4.2	± 4.2	4.9	4.9	± 4.9	
BASO		#	0.10	0.10	± 0.10	0.31	0.31	± 0.21	0.78	0.78	± 0.39	
		%	4.9	4.9	± 4.9	4.8	4.8	± 4.8	4.7	4.7	± 4.7	
IDR-SD	RDW-SD	fl	38.0	38.0	± 4.0	36.0	36.0	± 4.0	40.0	40.0	± 4.0	
ERB	NRBC	#	0.27	0.27	± 0.13	0.95	0.95	± 0.30	0.86	0.86	± 0.25	
		%	13.0	13.0	± 4.0	14.5	14.5	± 5.0	5.2	5.2	± 1.5	
PLTS	OPT	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	83	± 30	N/A	274	± 40	N/A	530	± 50	

1300074625
BACK / VERSO
Ref. TEMP-0821 Rev.45