


LOT PX 412
Rev 1

CONTROL

 (Exp.) 2018-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1		± 0.4	6.6		6.6	± 1.0		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.36	± 0.16	4.65	4.65	± 0.20	5.08	5.08	± 0.25	
HB HGB	g/dl	6.8	6.8	± 0.4	14.0	14.0	± 0.5	15.9	15.9	± 0.6	
	g/l	68	68	± 4	140	140	± 5	159	159	± 6	
	mmol/l	4.22	4.22	± 0.25	8.69	8.69	± 0.31	9.87	9.87	± 0.37	
HT HCT	%	19.6	19.6	± 1.5	38.6	38.6	± 2.0	44.2	44.2	± 2.5	
	l/l	0.196	0.196	± 0.015	0.386	0.386	± 0.020	0.442	0.442	± 0.025	
VGM MCV	µm ³ ; fl	83	83	± 5	83	83	± 5	87	87	± 5	
TGMH MCH	pg	28.8	28.8	± 2.0	30.1	30.1	± 2.0	31.3	31.3	± 2.5	
	fmol	1.79	1.79	± 0.12	1.87	1.87	± 0.12	1.94	1.94	± 0.16	
CCMH MCHC	g/dl	34.7	34.7	± 3.0	36.3	36.3	± 3.0	36.0	36.0	± 3.0	
	g/l	347	347	± 30	363	363	± 30	360	360	± 30	
	mmol/l	21.56	21.56	± 1.86	22.53	22.53	± 1.86	22.34	22.34	± 1.86	
IDR-CV RDW-CV	%	14.0	14.0	± 4.0	14.5	14.5	± 4.0	14.5	14.5	± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	± 20	263	263	± 30	517	517	± 50	
VPM MPV	µm ³ ; fl	9.9	9.9	± 2.0	9.7	9.7	± 2.0	9.6	9.6	± 2.0	
NEUT	#	1.24	1.24	± 0.35	3.91	3.91	± 0.90	10.60	10.60	± 1.90	
	%	59.2	59.2	± 10.0	59.3	59.3	± 10.0	69.3	69.3	± 10.0	
LYMPHO	#	0.50	0.50	± 0.33	1.72	1.72	± 0.70	2.03	2.03	± 1.50	
	%	23.7	23.7	± 12.0	26.0	26.0	± 8.0	13.3	13.3	± 8.0	
MONO	#	0.11	0.11	± 0.07	0.44	0.44	± 0.29	1.07	1.07	± 0.71	
	%	5.4	5.4	± 3.60	6.7	6.7	± 4.50	7.0	7.0	± 4.70	
EOS	#	0.15	0.15	± 0.15	0.23	0.23	± 0.23	0.89	0.89	± 0.89	
	%	7.1	7.1	± 7.1	3.5	3.5	± 3.50	5.8	5.8	± 5.8	
BASO	#	0.10	0.10	± 0.10	0.30	0.30	± 0.30	0.70	0.70	± 0.70	
	%	4.6	4.6	± 4.6	4.5	4.5	± 4.5	4.6	4.6	± 4.6	
IDR-SD RDW-SD	fl	38.5	38.5	± 8.0	41.5	41.5	± 8.0	43.0	43.0	± 8.0	
ERB NRBC	#	0.29	0.29	± 0.13	0.96	0.96	± 0.30	0.93	0.93	± 0.25	
	%	14.0	14.0	± 4.0	14.6	14.6	± 5.0	6.1	6.1	± 1.5	
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	74	± 30	N/A	235	± 40	N/A	476	± 50	

1300045049
BACK / VERSO
Ref. TEMP-0821 Rev.42