



## ABX Difftrol



**LOT** PX 432

**CONTROL**



(Exp.) 2022-01-05  
 (YYYY - MM - DD)

Rev 3

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio														
			CONTROL		L	TOLERANCES TOLERANCE	CONTROL		N	TOLERANCES TOLERANCE	CONTROL		H	TOLERANCES TOLERANCE			
			YUMIZEN		H1500		H2500	YUMIZEN			H1500	H2500	YUMIZEN		H1500	H2500	
			H1500	H2500				H1500	H2500				H1500				H2500
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.00	2.00	± 0.40	6.60	6.60	± 1.00	16.85	16.85	± 2.20						
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.47	2.47	± 0.16	4.57	4.57	± 0.20	5.10	5.10	± 0.25						
HB	HGB	g/dl	7.1	7.1	± 0.4	13.3	13.3	± 0.5	16.1	16.1	± 0.6						
		g/l	71	71	± 4	133	133	± 5	161	161	± 6						
		mmol/l	4.41	4.41	± 0.25	8.26	8.26	± 0.31	10.00	10.00	± 0.37						
HT	HCT	%	20.6	20.6	± 1.5	37.2	37.2	± 2.0	45.1	45.1	± 2.5						
		l/l	0.206	0.206	± 0.015	0.372	0.372	± 0.020	0.451	0.451	± 0.025						
VGM	MCV	µm <sup>3</sup> ·fl	83.5	83.5	± 5.0	81.5	81.5	± 5.0	88.5	88.5	± 5.0						
TGMH	MCH	pg	28.7	28.7	± 2.0	29.1	29.1	± 2.0	31.6	31.6	± 2.5						
		fmol	1.79	1.79	± 0.12	1.81	1.81	± 0.12	1.96	1.96	± 0.16						
CCMH	MCHC	g/dl	34.4	34.4	± 3.0	35.7	35.7	± 3.0	35.7	35.7	± 3.0						
		g/l	344	344	± 30	357	357	± 30	357	357	± 30						
		mmol/l	21.38	21.38	± 1.86	22.18	22.18	± 1.86	22.15	22.15	± 1.86						
IDR-CV	RDW-CV	%	14.0	14.0	± 4.0	14.0	14.0	± 4.0	12.0	12.0	± 4.0						
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	± 20	260	260	± 30	525	525	± 50						
VPM	MPV	µm <sup>3</sup> ·fl	10.5	10.5	± 2.0	9.4	9.4	± 2.0	10.2	10.2	± 2.0						
NEUT	#	1.21	1.21	± 0.35	3.70	3.70	± 0.90	11.09	11.09	± 1.90							
	%	60.5	60.5	± 10.0	56.0	56.0	± 10.0	65.8	65.8	± 10.0							
LYMPHO	#	0.40	0.40	± 0.33	1.48	1.48	± 0.70	1.75	1.75	± 1.50							
	%	19.9	19.9	± 12.0	22.4	22.4	± 8.0	10.4	10.4	± 8.0							
MONO	#	0.16	0.16	± 0.16	0.82	0.82	± 0.74	2.27	2.27	± 2.04							
	%	7.9	7.9	± 7.9	12.4	12.4	± 11.2	13.5	13.5	± 12.2							
EOS	#	0.14	0.14	± 0.14	0.31	0.31	± 0.31	0.94	0.94	± 0.94							
	%	7.1	7.1	± 7.1	4.7	4.7	± 4.7	5.6	5.6	± 5.6							
BASO	#	0.09	0.09	± 0.09	0.30	0.30	± 0.20	0.79	0.79	± 0.40							
	%	4.6	4.6	± 2.5	4.5	4.5	± 3.0	4.7	4.7	± 2.5							
IDR-SD	RDW-SD	fl	39.0	39.0	± 4.0	38.0	38.0	± 4.0	40.0	40.0	± 4.0						
ERB	NRBC	#	0.22	0.22	± 0.13	0.82	0.82	± 0.30	0.79	0.79	± 0.25						
		%	10.9	10.9	± 4.0	12.4	12.4	± 5.0	4.7	4.7	± 1.5						
PLTS	OPT	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	79	± 30	N/A	265	± 40	N/A	530	± 50						

1300074625  
 BACK / VERSO  
 Ref: TEMP-0821 Rev.45