


LOT PX 413
Rev 2

CONTROL

 (Exp.) 2018-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1	± 0.4	6.6	6.6	± 1.0	17.0	17.0	± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.45	2.45	± 0.16	4.67	4.67	± 0.20	5.21	5.21	± 0.25	
HB HGB	g/dl	6.9	6.9	± 0.4	14.1	14.1	± 0.5	16.6	16.6	± 0.6	
	g/l	69	69	± 4	141	141	± 5	166	166	± 6	
	mmol/l	4.28	4.28	± 0.25	8.76	8.76	± 0.31	10.31	10.31	± 0.37	
HT HCT	%	20.6	20.6	± 1.5	39.7	39.7	± 2.0	47.4	47.4	± 2.5	
	l/l	0.206	0.206	± 0.015	0.397	0.397	± 0.020	0.474	0.474	± 0.025	
VGM MCV	µm ³ ; fl	84	84	± 5	85	85	± 5	91	91	± 5	
TGMH MCH	pg	28.2	28.2	± 2.0	30.2	30.2	± 2.0	31.9	31.9	± 2.5	
	fmol	1.75	1.75	± 0.12	1.87	1.87	± 0.12	1.98	1.98	± 0.16	
CCMH MCHC	g/dl	33.5	33.5	± 3.0	35.5	35.5	± 3.0	35.0	35.0	± 3.0	
	g/l	335	335	± 30	355	355	± 30	350	350	± 30	
	mmol/l	20.82	20.82	± 1.86	22.06	22.06	± 1.86	21.74	21.74	± 1.86	
IDR-CV RDW-CV	%	14.2	14.2	± 4.0	13.7	13.7	± 4.0	13.8	13.8	± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74	± 20	254	254	± 30	510	510	± 50	
VPM MPV	µm ³ ; fl	9.8	9.8	± 2.0	9.7	9.7	± 2.0	10.0	10.0	± 2.0	
NEUT	#	1.32	1.32	± 0.35	3.92	3.92	± 0.90	12.19	12.19	± 1.90	
	%	62.7	62.7	± 10.0	59.4	59.4	± 10.0	71.7	71.7	± 10.0	
LYMPHO	#	0.40	0.40	± 0.33	1.57	1.57	± 0.70	1.75	1.75	± 1.50	
	%	19.0	19.0	± 12.0	23.8	23.8	± 8.0	10.3	10.3	± 8.0	
MONO	#	0.12	0.12	± 0.08	0.44	0.44	± 0.29	1.19	1.19	± 0.79	
	%	5.5	5.5	± 3.7	6.7	6.7	± 4.5	7.0	7.0	± 4.7	
EOS	#	0.17	0.17	± 0.17	0.36	0.36	± 0.36	1.09	1.09	± 1.09	
	%	8.0	8.0	± 8.0	5.4	5.4	± 5.4	6.4	6.4	± 6.4	
BASO	#	0.10	0.10	± 0.10	0.31	0.31	± 0.31	0.78	0.78	± 0.78	
	%	4.8	4.8	± 4.8	4.7	4.7	± 4.7	4.6	4.6	± 4.6	
IDR-SD RDW-SD	fl	40.0	40.0	± 8.0	39.8	39.8	± 8.0	44.2	44.2	± 8.0	
ERB NRBC	#	0.21	0.21	± 0.13	0.88	0.88	± 0.30	0.80	0.80	± 0.25	
	%	10.1	10.1	± 4.0	13.3	13.3	± 5.0	4.7	4.7	± 1.5	
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	84	± 30	N/A	255	± 40	N/A	502	± 50	

1300045049
BACK / VERSO
Ref: TEMP-0821 Rev.42