


LOT PX 423
Rev 2

CONTROL

 (Exp.) 2020-07-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | Lysebio | | | | | | | | | |
|--------------------------|--------|--|---------|-------|-------------------------|---------|-------|-------------------------|---------|-------|-------------------------|-------------------------|
| | | | CONTROL | | L | CONTROL | | N | CONTROL | | H | TOLERANCES TOLERANCE |
| | | | YUMIZEN | | TOLERANCES TOLERANCE | YUMIZEN | | TOLERANCES TOLERANCE | YUMIZEN | | TOLERANCES TOLERANCE | |
| | | | H1500 | H2500 | | H1500 | H2500 | | H1500 | H2500 | | |
| GB | WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.17 | 2.17 | ± 0.40 | 6.75 | 6.75 | ± 1.00 | 16.71 | 16.71 | ± 2.20 | |
| GR | RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.39 | 2.39 | ± 0.16 | 4.66 | 4.66 | ± 0.20 | 5.17 | 5.17 | ± 0.25 | |
| HB | HGB | g/dl | 6.6 | 6.6 | ± 0.4 | 13.8 | 13.8 | ± 0.5 | 16.3 | 16.3 | ± 0.6 | |
| | | g/l | 66 | 66 | ± 4 | 138 | 138 | ± 5 | 163 | 163 | ± 6 | |
| | | mmol/l | 4.10 | 4.10 | ± 0.25 | 8.57 | 8.57 | ± 0.31 | 10.12 | 10.12 | ± 0.37 | |
| HT | HCT | % | 19.5 | 19.5 | ± 1.5 | 38.9 | 38.9 | ± 2.0 | 45.8 | 45.8 | ± 2.5 | |
| | | l/l | 0.195 | 0.195 | ± 0.015 | 0.389 | 0.389 | ± 0.020 | 0.458 | 0.458 | ± 0.025 | |
| VGM | MCV | µm ³ ; fl | 81.5 | 81.5 | ± 5.0 | 83.5 | 83.5 | ± 5.0 | 88.5 | 88.5 | ± 5.0 | |
| TGMH | MCH | pg | 27.6 | 27.6 | ± 2.0 | 29.6 | 29.6 | ± 2.0 | 31.5 | 31.5 | ± 2.5 | |
| | | fmol | 1.71 | 1.71 | ± 0.12 | 1.84 | 1.84 | ± 0.12 | 1.96 | 1.96 | ± 0.16 | |
| CCMH | MCHC | g/dl | 33.9 | 33.9 | ± 3.0 | 35.5 | 35.5 | ± 3.0 | 35.6 | 35.6 | ± 3.0 | |
| | | g/l | 339 | 339 | ± 30 | 355 | 355 | ± 30 | 356 | 356 | ± 30 | |
| | | mmol/l | 21.04 | 21.04 | ± 1.86 | 22.02 | 22.02 | ± 1.86 | 22.12 | 22.12 | ± 1.86 | |
| IDR-CV | RDW-CV | % | 12.5 | 12.5 | ± 4.0 | 12.5 | 12.5 | ± 4.0 | 11.8 | 11.8 | ± 4.0 | |
| PLAQ. | PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 70 | 70 | ± 20 | 250 | 250 | ± 30 | 485 | 485 | ± 50 | |
| VPM | MPV | µm ³ ; fl | 9.8 | 9.8 | ± 2.0 | 9.4 | 9.4 | ± 2.0 | 9.7 | 9.7 | ± 2.0 | |
| NEUT | | # | 1.31 | 1.31 | ± 0.35 | 3.81 | 3.81 | ± 0.90 | 11.60 | 11.60 | ± 1.90 | |
| | | % | 60.3 | 60.3 | ± 10.0 | 56.5 | 56.5 | ± 10.0 | 69.4 | 69.4 | ± 10.0 | |
| LYMPHO | | # | 0.45 | 0.45 | ± 0.33 | 1.68 | 1.68 | ± 0.70 | 1.82 | 1.82 | ± 1.50 | |
| | | % | 20.7 | 20.7 | ± 12.0 | 24.9 | 24.9 | ± 8.0 | 10.9 | 10.9 | ± 8.0 | |
| MONO | | # | 0.13 | 0.13 | ± 0.13 | 0.65 | 0.65 | ± 0.59 | 1.52 | 1.52 | ± 1.37 | |
| | | % | 6.1 | 6.1 | ± 6.1 | 9.6 | 9.6 | ± 8.6 | 9.1 | 9.1 | ± 8.2 | |
| EOS | | # | 0.18 | 0.18 | ± 0.18 | 0.29 | 0.29 | ± 0.29 | 0.99 | 0.99 | ± 0.99 | |
| | | % | 8.1 | 8.1 | ± 8.1 | 4.3 | 4.3 | ± 4.3 | 5.9 | 5.9 | ± 5.9 | |
| BASO | | # | 0.10 | 0.10 | ± 0.10 | 0.32 | 0.32 | ± 0.31 | 0.79 | 0.79 | ± 0.77 | |
| | | % | 4.8 | 4.8 | ± 4.8 | 4.7 | 4.7 | ± 4.7 | 4.7 | 4.7 | ± 4.7 | |
| IDR-SD | RDW-SD | fl | 37.0 | 37.0 | ± 4.0 | 39.5 | 39.5 | ± 4.0 | 39.5 | 39.5 | ± 4.0 | |
| ERB | NRBC | # | 0.29 | 0.29 | ± 0.13 | 1.03 | 1.03 | ± 0.30 | 0.91 | 0.91 | ± 0.25 | |
| | | % | 13.2 | 13.2 | ± 4.0 | 15.3 | 15.3 | ± 5.0 | 5.4 | 5.4 | ± 1.5 | |
| PLTS | OPT | 10 ⁹ /mm ³ ; 10 ⁹ /l | N/A | 75 | ± 30 | N/A | 254 | ± 40 | N/A | 485 | ± 50 | |

1300074625
BACK / VERSO
Ref. TEMP-0821 Rev.44