


**LOT** PX 432  
Rev 3

**CONTROL**

 (Exp.) 2022-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Whitediff									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H550	H500 OT H500 CT		H550	H500 OT H500 CT		H550	H500 OT H500 CT		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.30	2.30		± 0.40	7.15		7.15	± 1.00		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.38	± 0.16	4.56	4.56	± 0.20	5.13	5.13	± 0.25	
HB HGB	g/dl	6.9	6.9	± 0.4	13.0	13.0	± 0.5	15.8	15.8	± 0.6	
	g/l	69	69	± 4	130	130	± 5	158	158	± 6	
	mmol/l	4.28	4.28	± 0.25	8.07	8.07	± 0.31	9.81	9.81	± 0.37	
HT HCT	%	20.2	20.2	± 1.5	37.8	37.8	± 2.0	45.7	45.7	± 2.5	
	l/l	0.202	0.202	± 0.015	0.378	0.378	± 0.020	0.457	0.457	± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	85.0	85.0	± 5.0	83.0	83.0	± 5.0	89.0	89.0	± 5.0	
TGMH MCH	pg	29.0	29.0	± 2.0	28.5	28.5	± 2.0	30.8	30.8	± 2.5	
	fmol	1.80	1.80	± 0.12	1.77	1.77	± 0.12	1.91	1.91	± 0.16	
CCMH MCHC	g/dl	34.1	34.1	± 3.0	34.3	34.3	± 3.0	34.6	34.6	± 3.0	
	g/l	341	341	± 30	343	343	± 30	346	346	± 30	
	mmol/l	21.18	21.18	± 1.86	21.30	21.30	± 1.86	21.49	21.49	± 1.86	
IDR-CV RDW-CV	%	17.0	17.0	± 4.0	15.0	15.0	± 4.0	13.5	13.5	± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	75	± 20	265	265	± 30	534	534	± 50	
VPM MPV	µm <sup>3</sup> ; fl	11.0	11.0	± 2.0	9.7	9.7	± 2.0	10.9	10.9	± 2.0	
NEUT	#	1.16	1.16	± 0.35	3.65	3.65	± 0.90	11.00	11.00	± 1.90	
	%	50.3	50.3	± 10.0	51.1	51.1	± 10.0	64.7	64.7	± 10.0	
LYMPHO	#	0.65	0.65	± 0.33	2.25	2.25	± 0.70	2.67	2.67	± 1.50	
	%	28.4	28.4	± 12.0	31.4	31.4	± 8.0	15.7	15.7	± 8.0	
MONO	#	0.22	0.22	± 0.22	0.87	0.87	± 0.87	2.18	2.18	± 2.18	
	%	9.4	9.4	± 9.4	12.2	12.2	± 12.2	12.8	12.8	± 12.8	
EOS	#	0.15	0.15	± 0.15	0.26	0.26	± 0.26	0.83	0.83	± 0.83	
	%	6.7	6.7	± 6.7	3.7	3.7	± 3.7	4.9	4.9	± 4.9	
BASO	#	0.12	0.12	± 0.12	0.11	0.11	± 0.11	0.32	0.32	± 0.32	
	%	5.2	5.2	± 5.2	1.6	1.6	± 1.6	1.9	1.9	± 1.9	
IDR-SD RDW-SD	fl	43.0	43.0	± 8.0	41.0	41.0	± 8.0	42.5	42.5	± 8.0	

Ref: TEMP-0821 Rev.45 FRONT / RECTO 1300074625