


**LOT** PX 055  
Rev 2

**CONTROL**

 (Exp.) 2015-07-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Whitediff									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H500			H500			H500			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.3</b>		± 0.4	<b>7.1</b>		± 1.0	<b>16.8</b>		± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.43</b>		± 0.16	<b>4.68</b>		± 0.20	<b>5.21</b>		± 0.25	
HB HGB	g/dl	<b>7.0</b>		± 0.4	<b>13.7</b>		± 0.5	<b>16.2</b>		± 0.6	
	g/l	<b>70</b>		± 4	<b>137</b>		± 5	<b>162</b>		± 6	
	mmol/l	<b>4.35</b>		± 0.25	<b>8.51</b>		± 0.31	<b>10.06</b>		± 0.37	
HT HCT	%	<b>19.7</b>		± 1.5	<b>38.9</b>		± 2.0	<b>45.4</b>		± 2.5	
	l/l	<b>0.197</b>		± 0.015	<b>0.389</b>		± 0.020	<b>0.454</b>		± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>81</b>		± 5	<b>83</b>		± 5	<b>87</b>		± 5	
TGMH MCH	pg	<b>28.8</b>		± 2.0	<b>29.3</b>		± 2.0	<b>31.1</b>		± 2.5	
	fmol	<b>1.79</b>		± 0.12	<b>1.82</b>		± 0.12	<b>1.93</b>		± 0.16	
CCMH MCHC	g/dl	<b>35.6</b>		± 3.0	<b>35.3</b>		± 3.0	<b>35.7</b>		± 3.0	
	g/l	<b>356</b>		± 30	<b>353</b>		± 30	<b>357</b>		± 30	
	mmol/l	<b>22.11</b>		± 1.86	<b>21.92</b>		± 1.86	<b>22.17</b>		± 1.86	
IDR RDW	%	<b>11.1</b>		± 4.0	<b>11.4</b>		± 4.0	<b>10.9</b>		± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>71</b>		± 20	<b>258</b>		± 30	<b>486</b>		± 50	
VPM MPV	µm <sup>3</sup> ; fl	<b>9.0</b>		± 2.0	<b>9.1</b>		± 2.0	<b>9.0</b>		± 2.0	
NEUT	#	<b>1.40</b>		± 0.35	<b>4.11</b>		± 0.90	<b>12.05</b>		± 1.90	
	%	<b>60.6</b>		± 10.0	<b>57.9</b>		± 10.0	<b>71.7</b>		± 10.0	
LYMPHO	#	<b>0.57</b>		± 0.33	<b>2.32</b>		± 0.70	<b>2.74</b>		± 1.50	
	%	<b>24.9</b>		± 12.0	<b>32.7</b>		± 8.0	<b>16.3</b>		± 8.0	
MONO	#	<b>0.08</b>		± 0.08	<b>0.34</b>		± 0.34	<b>0.86</b>		± 0.86	
	%	<b>3.3</b>		± 3.3	<b>4.8</b>		± 4.8	<b>5.1</b>		± 5.1	
EOS	#	<b>0.18</b>		± 0.18	<b>0.25</b>		± 0.25	<b>0.89</b>		± 0.89	
	%	<b>8.0</b>		± 8.0	<b>3.5</b>		± 3.5	<b>5.3</b>		± 5.3	
BASO	#	<b>0.07</b>		± 0.07	<b>0.08</b>		± 0.08	<b>0.27</b>		± 0.27	
	%	<b>3.2</b>		± 3.2	<b>1.1</b>		± 1.1	<b>1.6</b>		± 1.6	

9930096-A  
Ref: TEMP-0821 Rev.36 BACK / VERSO