


**LOT** PX 075  
Rev 1

**CONTROL**

 (Exp.) 2015-09-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Whitediff									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H500			H500			H500			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.6</b>		± 0.4	<b>7.4</b>		± 1.0	<b>17.7</b>		± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.35</b>		± 0.16	<b>4.65</b>		± 0.20	<b>5.19</b>		± 0.25	
HB HGB	g/dl	<b>6.7</b>		± 0.4	<b>13.6</b>		± 0.5	<b>16.3</b>		± 0.6	
	g/l	<b>67</b>		± 4	<b>136</b>		± 5	<b>163</b>		± 6	
	mmol/l	<b>4.16</b>		± 0.25	<b>8.45</b>		± 0.31	<b>10.12</b>		± 0.37	
HT HCT	%	<b>19.0</b>		± 1.5	<b>39.1</b>		± 2.0	<b>46.2</b>		± 2.5	
	l/l	<b>0.190</b>		± 0.015	<b>0.391</b>		± 0.020	<b>0.462</b>		± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>81</b>		± 5	<b>84</b>		± 5	<b>89</b>		± 5	
TGMH MCH	pg	<b>28.5</b>		± 2.0	<b>29.2</b>		± 2.0	<b>31.4</b>		± 2.5	
	fmol	<b>1.77</b>		± 0.12	<b>1.81</b>		± 0.12	<b>1.95</b>		± 0.16	
CCMH MCHC	g/dl	<b>35.2</b>		± 3.0	<b>34.8</b>		± 3.0	<b>35.3</b>		± 3.0	
	g/l	<b>352</b>		± 30	<b>348</b>		± 30	<b>353</b>		± 30	
	mmol/l	<b>21.86</b>		± 1.86	<b>21.61</b>		± 1.86	<b>21.92</b>		± 1.86	
IDR RDW	%	<b>12.2</b>		± 4.0	<b>11.8</b>		± 4.0	<b>11.1</b>		± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>73</b>		± 20	<b>254</b>		± 30	<b>492</b>		± 50	
VPM MPV	µm <sup>3</sup> ; fl	<b>9.7</b>		± 2.0	<b>9.8</b>		± 2.0	<b>9.8</b>		± 2.0	
NEUT	#	<b>1.35</b>		± 0.35	<b>4.08</b>		± 0.90	<b>11.75</b>		± 1.90	
	%	<b>52.0</b>		± 10.0	<b>55.2</b>		± 10.0	<b>66.4</b>		± 10.0	
LYMPHO	#	<b>0.74</b>		± 0.33	<b>2.30</b>		± 0.70	<b>2.76</b>		± 1.50	
	%	<b>28.5</b>		± 12.0	<b>31.1</b>		± 8.0	<b>15.6</b>		± 8.0	
MONO	#	<b>0.24</b>		± 0.24	<b>0.64</b>		± 0.64	<b>1.82</b>		± 1.82	
	%	<b>9.1</b>		± 9.1	<b>8.6</b>		± 8.6	<b>10.3</b>		± 10.3	
EOS	#	<b>0.14</b>		± 0.14	<b>0.27</b>		± 0.27	<b>0.94</b>		± 0.94	
	%	<b>5.4</b>		± 5.4	<b>3.6</b>		± 3.6	<b>5.3</b>		± 5.3	
BASO	#	<b>0.13</b>		± 0.13	<b>0.11</b>		± 0.11	<b>0.42</b>		± 0.42	
	%	<b>5.0</b>		± 5.0	<b>1.5</b>		± 1.5	<b>2.4</b>		± 2.4	

9930096-A  
Ref: TEMP-0821 Rev.36 BACK / VERSO