


**LOT** PX 095  
Rev 1

**CONTROL**

 (Exp.) 2015-11-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Whitediff									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H500			H500			H500			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4		± 0.4	7.1		± 1.0	17.4		± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.40		± 0.16	4.56		± 0.20	5.35		± 0.25	
HB HGB	g/dl	6.9		± 0.4	13.3		± 0.5	16.9		± 0.6	
	g/l	69		± 4	133		± 5	169		± 6	
	mmol/l	4.28		± 0.25	8.26		± 0.31	10.49		± 0.37	
HT HCT	%	20.2		± 1.5	38.4		± 2.0	48.1		± 2.5	
	l/l	0.202		± 0.015	0.384		± 0.020	0.481		± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	84		± 5	84		± 5	90		± 5	
TGMH MCH	pg	28.8		± 2.0	29.2		± 2.0	31.6		± 2.5	
	fmol	1.79		± 0.12	1.81		± 0.12	1.96		± 0.16	
CCMH MCHC	g/dl	34.2		± 3.0	34.7		± 3.0	35.1		± 3.0	
	g/l	342		± 30	347		± 30	351		± 30	
	mmol/l	21.24		± 1.86	21.55		± 1.86	21.80		± 1.86	
IDR RDW	%	16.0		± 4.0	14.0		± 4.0	14.0		± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73		± 20	250		± 30	497		± 50	
VPM MPV	µm <sup>3</sup> ; fl	9.9		± 2.0	10.2		± 2.0	9.9		± 2.0	
NEUT	#	1.41		± 0.35	4.20		± 0.90	12.99		± 1.90	
	%	58.1		± 10.0	59.0		± 10.0	74.7		± 10.0	
LYMPHO	#	0.73		± 0.33	2.20		± 0.70	2.54		± 1.50	
	%	30.3		± 12.0	31.0		± 8.0	14.6		± 8.0	
MONO	#	0.11		± 0.11	0.28		± 0.28	0.61		± 0.61	
	%	4.5		± 4.5	4.0		± 4.0	3.5		± 3.5	
EOS	#	0.13		± 0.13	0.29		± 0.29	0.87		± 0.87	
	%	5.5		± 5.5	4.0		± 4.0	5.0		± 5.0	
BASO	#	0.04		± 0.04	0.14		± 0.14	0.39		± 0.39	
	%	1.6		± 1.6	2.0		± 2.0	2.2		± 2.2	

9930096-A  
Ref: TEMP-0821 Rev.36 BACK / VERSO