


LOT PX 115
Rev 1

CONTROL

 (Exp.) 2016-01-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H500			H500			H500			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4		± 0.4	7.4		± 1.0	16.5		± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.40		± 0.16	4.55		± 0.20	5.13		± 0.25	
HB	HGB	g/dl	6.8		± 0.4	13.3		± 0.5	15.8		± 0.6	
		g/l	68		± 4	133		± 5	158		± 6	
		mmol/l	4.22		± 0.25	8.26		± 0.31	9.81		± 0.37	
HT	HCT	%	19.9		± 1.5	38.3		± 2.0	45.4		± 2.5	
		l/l	0.199		± 0.015	0.383		± 0.020	0.454		± 0.025	
VGM	MCV	µm ³ ; fl	83		± 5	84		± 5	89		± 5	
TGMH	MCH	pg	28.3		± 2.0	29.2		± 2.0	30.8		± 2.5	
		fmol	1.76		± 0.12	1.81		± 0.12	1.91		± 0.16	
CCMH	MCHC	g/dl	34.1		± 3.0	34.8		± 3.0	34.6		± 3.0	
		g/l	341		± 30	348		± 30	346		± 30	
		mmol/l	21.18		± 1.86	21.61		± 1.86	21.49		± 1.86	
IDR	RDW	%	15.0		± 4.0	13.6		± 4.0	13.4		± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	82		± 20	276		± 30	529		± 50	
VPM	MPV	µm ³ ; fl	9.6		± 2.0	9.9		± 2.0	9.2		± 2.0	
NEUT		#	1.15		± 0.35	3.84		± 0.90	11.10		± 1.90	
		%	47.4		± 10.0	51.6		± 10.0	67.4		± 10.0	
LYMPHO		#	0.73		± 0.33	2.21		± 0.70	2.35		± 1.50	
		%	30.1		± 12.0	29.7		± 8.0	14.3		± 8.0	
MONO		#	0.27		± 0.27	0.98		± 0.98	1.89		± 1.89	
		%	11.0		± 11.0	13.2		± 13.2	11.5		± 11.5	
EOS		#	0.15		± 0.15	0.29		± 0.29	0.81		± 0.81	
		%	6.4		± 6.4	3.9		± 3.9	4.9		± 4.9	
BASO		#	0.12		± 0.12	0.12		± 0.12	0.32		± 0.32	
		%	5.1		± 5.1	1.6		± 1.6	1.9		± 1.9	

9930096-A
Ref: TEMP-0821 Rev.36 BACK / VERSO