


**LOT** PX 016  
Rev 1

**CONTROL**

 (Exp.) 2016-03-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN			YUMIZEN			YUMIZEN			
			H500		TOLERANCES TOLERANCE	H500		TOLERANCES TOLERANCE	H500		TOLERANCES TOLERANCE	
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.3</b>		± 0.4	<b>7.2</b>		± 1.0	<b>17.2</b>		± 2.2	
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.32</b>		± 0.16	<b>4.61</b>		± 0.20	<b>5.22</b>		± 0.25	
HB	HGB	g/dl	<b>6.6</b>		± 0.4	<b>13.5</b>		± 0.5	<b>16.3</b>		± 0.6	
		g/l	<b>66</b>		± 4	<b>135</b>		± 5	<b>163</b>		± 6	
		mmol/l	<b>4.10</b>		± 0.25	<b>8.38</b>		± 0.31	<b>10.12</b>		± 0.37	
HT	HCT	%	<b>19.3</b>		± 1.5	<b>38.7</b>		± 2.0	<b>46.5</b>		± 2.5	
		l/l	<b>0.193</b>		± 0.015	<b>0.387</b>		± 0.020	<b>0.465</b>		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	<b>83</b>		± 5	<b>84</b>		± 5	<b>89</b>		± 5	
TGMH	MCH	pg	<b>28.4</b>		± 2.0	<b>29.3</b>		± 2.0	<b>31.2</b>		± 2.5	
		fmol	<b>1.76</b>		± 0.12	<b>1.82</b>		± 0.12	<b>1.94</b>		± 0.16	
CCMH	MCHC	g/dl	<b>34.3</b>		± 3.0	<b>34.9</b>		± 3.0	<b>35.1</b>		± 3.0	
		g/l	<b>343</b>		± 30	<b>349</b>		± 30	<b>351</b>		± 30	
		mmol/l	<b>21.30</b>		± 1.86	<b>21.67</b>		± 1.86	<b>21.80</b>		± 1.86	
IDR	RDW	%	<b>15.0</b>		± 4.0	<b>13.5</b>		± 4.0	<b>13.2</b>		± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>75</b>		± 20	<b>276</b>		± 30	<b>530</b>		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	<b>9.2</b>		± 2.0	<b>9.3</b>		± 2.0	<b>9.0</b>		± 2.0	
NEUT		#	<b>1.19</b>		± 0.35	<b>3.92</b>		± 0.90	<b>11.95</b>		± 1.90	
		%	<b>51.7</b>		± 10.0	<b>54.5</b>		± 10.0	<b>69.5</b>		± 10.0	
LYMPHO		#	<b>0.74</b>		± 0.33	<b>2.16</b>		± 0.70	<b>2.68</b>		± 1.50	
		%	<b>32.0</b>		± 12.0	<b>30.0</b>		± 8.0	<b>15.6</b>		± 8.0	
MONO		#	<b>0.16</b>		± 0.16	<b>0.71</b>		± 0.71	<b>1.46</b>		± 1.46	
		%	<b>6.9</b>		± 6.9	<b>9.8</b>		± 9.8	<b>8.5</b>		± 8.5	
EOS		#	<b>0.15</b>		± 0.15	<b>0.30</b>		± 0.30	<b>0.86</b>		± 0.86	
		%	<b>6.5</b>		± 6.5	<b>4.2</b>		± 4.2	<b>5.0</b>		± 5.0	
BASO		#	<b>0.07</b>		± 0.07	<b>0.11</b>		± 0.11	<b>0.24</b>		± 0.24	
		%	<b>2.9</b>		± 2.9	<b>1.5</b>		± 1.5	<b>1.4</b>		± 1.4	

9930096-A  
Ref: TEMP-0821 Rev.36 BACK / VERSO