


LOT PX 056
Rev 4

CONTROL

 (Exp.) 2016-07-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.3	± 0.4		7.2	7.1		± 1.0	16.9		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.37	2.33	± 0.16	4.64	4.58	± 0.20	5.18	5.07	± 0.25		
HB HGB	g/dl	6.7	6.7	± 0.4	13.4	13.4	± 0.5	15.6	15.5	± 0.6		
	g/l	67	67	± 4	134	134	± 5	156	155	± 6		
	mmol/l	4.16	4.16	± 0.25	8.32	8.32	± 0.31	9.69	9.63	± 0.37		
HT HCT	%	19.4	19.6	± 1.5	38.5	38.9	± 2.0	44.5	44.1	± 2.5		
	l/l	0.194	0.196	± 0.015	0.385	0.389	± 0.020	0.445	0.441	± 0.025		
VGM MCV	µm ³ ; fl	82	84	± 5	83	85	± 5	86	87	± 5		
TGMH MCH	pg	28.3	28.8	± 2.0	28.9	29.3	± 2.0	30.1	30.6	± 2.5		
	fmol	1.76	1.79	± 0.12	1.79	1.82	± 0.12	1.87	1.90	± 0.16		
CCMH MCHC	g/dl	34.5	34.2	± 3.0	34.8	34.4	± 3.0	35.0	35.1	± 3.0		
	g/l	345	342	± 30	348	344	± 30	350	351	± 30		
	mmol/l	21.42	21.24	± 1.86	21.61	21.36	± 1.86	21.74	21.80	± 1.86		
IDR RDW	%	14.8	15.2	± 4.0	13.8	14.2	± 4.0	13.2	13.6	± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	80	72	± 20	280	256	± 30	535	505	± 50		
VPM MPV	µm ³ ; fl	9.5	9.9	± 2.0	9.9	10.2	± 2.0	9.5	9.8	± 2.0		
NEUT	#	1.50	1.45	± 0.35	3.87	3.77	± 0.90	11.83	11.60	± 1.90		
	%	62.6	62.6	± 10.0	53.7	53.0	± 10.0	70.0	70.3	± 10.0		
LYMPHO	#	0.50	0.45	± 0.33	2.13	2.03	± 0.70	2.21	1.88	± 1.50		
	%	21.0	19.4	± 12.0	29.6	28.6	± 8.0	13.1	11.4	± 8.0		
MONO	#	0.18	0.20	± 0.18	0.80	0.91	± 0.80	1.62	1.83	± 1.62		
	%	7.3	8.8	± 7.3	11.1	12.7	± 11.1	9.6	11.1	± 9.6		
EOS	#	0.11	0.11	± 0.11	0.30	0.31	± 0.30	1.00	0.99	± 0.99		
	%	4.7	4.6	± 4.6	4.2	4.4	± 4.2	5.9	6.0	± 5.9		
BASO	#	0.11	0.11	± 0.11	0.10	0.09	± 0.09	0.24	0.20	± 0.20		
	%	4.4	4.6	± 4.4	1.4	1.3	± 1.3	1.4	1.2	± 1.2		

Ref: TEMP-0821 Rev.37 BACK / VERSO 1300019740-A