


LOT PX 400
Rev 2

CONTROL

 (Exp.) 2016-09-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.3	± 0.4	7.3	7.3	± 1.0	17.3	17.2	± 2.2		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38	2.36	± 0.16	4.65	4.62	± 0.20	5.14	5.09	± 0.25		
HB HGB	g/dl	6.9	6.8	± 0.4	13.2	13.1	± 0.5	16.0	15.8	± 0.6		
	g/l	69	68	± 4	132	131	± 5	160	158	± 6		
	mmol/l	4.28	4.22	± 0.25	8.20	8.14	± 0.31	9.94	9.81	± 0.37		
HT HCT	%	19.5	19.8	± 1.5	37.2	38.3	± 2.0	44.7	45.8	± 2.5		
	l/l	0.195	0.198	± 0.015	0.372	0.383	± 0.020	0.447	0.458	± 0.025		
VGM MCV	µm ³ ; fl	82	84	± 5	80	83	± 5	87	90	± 5		
TGMH MCH	pg	29.0	28.8	± 2.0	28.4	28.4	± 2.0	31.1	31.0	± 2.5		
	fmol	1.80	1.79	± 0.12	1.76	1.76	± 0.12	1.93	1.93	± 0.16		
CCMH MCHC	g/dl	35.4	34.3	± 3.0	35.5	34.2	± 3.0	35.8	34.5	± 3.0		
	g/l	354	343	± 30	355	342	± 30	358	345	± 30		
	mmol/l	21.98	21.30	± 1.86	22.05	21.24	± 1.86	22.23	21.42	± 1.86		
IDR RDW	%	13.8	13.8	± 4.0	13.0	13.5	± 4.0	12.5	13.6	± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	80	69	± 20	275	252	± 30	528	491	± 50		
VPM MPV	µm ³ ; fl	9.8	10.6	± 2.0	9.9	10.3	± 2.0	9.3	9.5	± 2.0		
NEUT	#	1.33	1.29	± 0.35	3.80	3.88	± 0.90	11.91	12.07	± 1.90		
	%	55.6	56.0	± 10.0	52.0	53.1	± 10.0	68.9	70.2	± 10.0		
LYMPHO	#	0.66	0.60	± 0.33	2.39	2.31	± 0.70	2.70	2.44	± 1.50		
	%	27.4	26.3	± 12.0	32.7	31.7	± 8.0	15.6	14.2	± 8.0		
MONO	#	0.16	0.16	± 0.16	0.66	0.66	± 0.66	1.52	1.50	± 1.50		
	%	6.6	7.0	± 6.6	9.1	9.1	± 9.1	8.8	8.7	± 8.7		
EOS	#	0.13	0.13	± 0.13	0.31	0.30	± 0.30	0.73	0.76	± 0.73		
	%	5.4	5.7	± 5.4	4.2	4.1	± 4.1	4.2	4.4	± 4.2		
BASO	#	0.12	0.12	± 0.12	0.15	0.15	± 0.15	0.43	0.43	± 0.43		
	%	5.0	5.0	± 5.0	2.0	2.0	± 2.0	2.5	2.5	± 2.5		

Ref: TEMP-0821 Rev.37 BACK / VERSO 1300019740-A