


LOT PX 401
Rev 1

CONTROL

 (Exp.) **2016-11-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3	± 0.4		7.2	7.2		± 1.0	17.4		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.25	2.23	± 0.16	4.50	4.47	± 0.20	5.12	5.07	± 0.25		
HB HGB	g/dl	6.6	6.6	± 0.4	13.1	13.3	± 0.5	15.5	15.6	± 0.6		
	g/l	66	66	± 4	131	133	± 5	155	156	± 6		
	mmol/l	4.10	4.10	± 0.25	8.14	8.26	± 0.31	9.63	9.69	± 0.37		
HT HCT	%	18.5	18.5	± 1.5	37.4	37.5	± 2.0	44.0	44.1	± 2.5		
	l/l	0.185	0.185	± 0.015	0.374	0.375	± 0.020	0.440	0.441	± 0.025		
VGM MCV	µm ³ ; fl	82	83	± 5	83	84	± 5	86	87	± 5		
TGMH MCH	pg	29.3	29.6	± 2.0	29.1	29.8	± 2.0	30.3	30.8	± 2.5		
	fmol	1.82	1.84	± 0.12	1.81	1.85	± 0.12	1.88	1.91	± 0.16		
CCMH MCHC	g/dl	35.8	35.7	± 3.0	35.1	35.4	± 3.0	35.2	35.4	± 3.0		
	g/l	358	357	± 30	351	354	± 30	352	354	± 30		
	mmol/l	22.23	22.17	± 1.86	21.80	21.98	± 1.86	21.86	21.98	± 1.86		
IDR RDW	%	13.8	15.0	± 4.0	13.5	15.1	± 4.0	13.0	14.2	± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	67	± 20	255	238	± 30	496	500	± 50		
VPM MPV	µm ³ ; fl	9.7	10.1	± 2.0	9.3	10.0	± 2.0	9.8	10.3	± 2.0		
NEUT	#	1.24	1.25	± 0.35	3.98	3.82	± 0.90	12.23	12.06	± 1.90		
	%	53.7	53.7	± 10.0	55.3	53.4	± 10.0	70.3	69.9	± 10.0		
LYMPHO	#	0.59	0.57	± 0.33	2.41	2.39	± 0.70	2.68	2.54	± 1.50		
	%	25.6	24.6	± 12.0	33.5	33.5	± 8.0	15.4	14.7	± 8.0		
MONO	#	0.22	0.25	± 0.22	0.45	0.60	± 0.45	1.31	1.49	± 1.31		
	%	9.5	10.8	± 9.5	6.3	8.3	± 6.3	7.5	8.6	± 7.5		
EOS	#	0.14	0.14	± 0.14	0.21	0.20	± 0.20	0.75	0.75	± 0.75		
	%	6.2	5.9	± 5.9	2.9	2.8	± 2.8	4.3	4.3	± 4.3		
BASO	#	0.12	0.12	± 0.12	0.14	0.14	± 0.14	0.44	0.44	± 0.44		
	%	5.0	5.0	± 5.0	2.0	2.0	± 2.0	2.5	2.5	± 2.5		

Ref: TEMP-0821 Rev.37 BACK / VERSO 1300019740-A