


**LOT** PX 403  
Rev 2

**CONTROL**

 (Exp.) 2017-03-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	Whitediff									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		H500 OT v1 only	H500 OT H500 CT		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	± 0.4	7.3	7.2	± 1.0	18.0	17.6	± 2.2	
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.33	2.32	± 0.16	4.63	4.59	± 0.20	5.16	5.13	± 0.25	
HB	HGB	g/dl	6.8	6.7	± 0.4	13.4	13.5	± 0.5	15.9	16.0	± 0.6	
		g/l	68	67	± 4	134	135	± 5	159	160	± 6	
		mmol/l	4.22	4.16	± 0.25	8.32	8.38	± 0.31	9.87	9.94	± 0.37	
HT	HCT	%	19.3	19.3	± 1.5	38.4	38.6	± 2.0	45.4	45.7	± 2.5	
		l/l	0.193	0.193	± 0.015	0.384	0.386	± 0.020	0.454	0.457	± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	83	83	± 5	83	84	± 5	88	89	± 5	
TGMH	MCH	pg	29.2	28.9	± 2.0	28.9	29.4	± 2.0	30.8	31.2	± 2.5	
		fmol	1.81	1.79	± 0.12	1.79	1.83	± 0.12	1.91	1.94	± 0.16	
CCMH	MCHC	g/dl	35.2	34.8	± 3.0	34.9	35.0	± 3.0	35.0	35.0	± 3.0	
		g/l	352	348	± 30	349	350	± 30	350	350	± 30	
		mmol/l	21.86	21.61	± 1.86	21.67	21.74	± 1.86	21.74	21.74	± 1.86	
IDR	RDW	%	12.2	14.8	± 4.0	12.0	14.2	± 4.0	11.5	14.7	± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	73	± 20	255	255	± 30	500	506	± 50	
VPM	MPV	µm <sup>3</sup> ; fl	10.5	10.4	± 2.0	10.6	10.4	± 2.0	10.6	10.3	± 2.0	
NEUT		#	1.30	1.27	± 0.35	3.74	3.57	± 0.90	12.10	11.55	± 1.90	
		%	54.2	52.9	± 10.0	51.2	49.5	± 10.0	67.2	65.7	± 10.0	
LYMPHO		#	0.68	0.64	± 0.33	2.28	2.26	± 0.70	2.52	2.49	± 1.50	
		%	28.5	26.8	± 12.0	31.3	31.3	± 8.0	14.0	14.2	± 8.0	
MONO		#	0.17	0.23	± 0.17	0.89	1.01	± 0.89	2.09	2.25	± 2.09	
		%	7.0	9.7	± 7.0	12.2	14.0	± 12.2	11.6	12.8	± 11.6	
EOS		#	0.13	0.13	± 0.13	0.24	0.23	± 0.23	0.85	0.85	± 0.85	
		%	5.3	5.6	± 5.3	3.3	3.2	± 3.2	4.7	4.8	± 4.7	
BASO		#	0.12	0.12	± 0.12	0.15	0.14	± 0.14	0.45	0.44	± 0.44	
		%	5.0	5.0	± 5.0	2.0	2.0	± 2.0	2.5	2.5	± 2.5	

Ref: TEMP-0821 Rev.39 BACK / VERSO 1300027264-A