



# Yumizen G PT Reco 10 for Yumizen G Line



D 4 9 1 0 9 1 1 3 6 3 7 6 0 9 2 3 D



D 5 1 0 0 1 0 1 5 0 0 1 3 2 3 3 0 D



D 6 1 6 3 2 5 0 1 9 2 1 0 1 1 0 3 D



D 7 1 3 7 D

<b>Reference No.:</b>		<b>1300036376</b>	
<b>Lot No.:</b>		<b>910911 (rev. 2)</b>	
<b>Expiry date:</b>		<b>2023-09-30</b>	
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>
<b>Optical</b>	<b>Yumizen G Line</b>		<b>%</b>
		<b>100</b>	<b>10,1</b>
		<b>50</b>	<b>13,2</b>
		<b>33</b>	<b>16,3</b>
		<b>25</b>	<b>19,2</b>
	<b>1,03</b>	<b>MNPT</b>	<b>10,1</b>
<b>Important notice</b>			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the <b>Min. Time</b> must be set to 7 seconds and the <b>Lag time</b> to 5 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>			