



# Yumizen G PT 5 for Yumizen G Line



D 4 9 2 0 1 0 1 3 6 3 3 8 0 1 2 4 D



D 5 1 0 0 1 2 6 5 0 0 1 5 7 3 3 0 D



D 6 1 9 0 2 5 0 2 2 0 1 2 6 1 2 5 D



D 7 1 3 7 D

<b>Reference No.:</b>		<b>1300036338</b>		
<b>Lot No.:</b>		<b>920101 (rev. 1)</b>		
<b>Expiry date:</b>		<b>2024-01-31</b>		
<b>Coagulometers</b>		<b>ISI</b>	<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G Line</b>		<b>1,25</b>	<b>%</b>
		<b>100</b>		<b>12,6</b>
		<b>50</b>		<b>15,7</b>
		<b>33</b>		<b>19,0</b>
		<b>25</b>		<b>22,0</b>
			<b>MNPT</b>	<b>12,6</b>
<b>Important notice</b>				
<ul style="list-style-type: none"> <li>- The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</li> <li>- The efficiency of the reconstitution of the reagent may be improved by performing it at 37°C.</li> </ul>				